

AMERICAN NATIONAL STANDARD



*for Ophthalmics –  
Laser Systems for  
Corneal Reshaping*



**ANSI<sup>®</sup>**  
**Z80.11-2007**

American National Standard  
for Ophthalmics –

**Laser Systems for  
Corneal Reshaping**

Secretariat

**Information Technology Industry Council**

Approved July 31, 2007

**American National Standards Institute, Inc.**

# American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

**CAUTION NOTICE:** This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

## *Developed by*

The Accredited Committee Z80 for Ophthalmic Standards -

Optical Laboratories Association  
Z80 Secretariat  
11096 Lee Highway  
A101  
Fairfax, VA 22030-5039

## *Published by*

Optical Laboratories Association  
11096 Lee Highway  
A101  
Fairfax, VA 22030-5039

Copyright © 2007 by Optical Laboratories Association  
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

## Contents

	Page
Foreword .....	iv
<b>1</b> Scope and purpose .....	1
<b>2</b> Normative references .....	1
<b>3</b> Definitions .....	3
<b>4</b> Mechanical, thermal, and environmental requirements .....	5
4.1 Combination of different devices .....	5
4.2 Materials .....	5
4.3 Resistance to transport and storage conditions .....	5
<b>5</b> Safety requirements .....	6
5.1 Protection against contaminants .....	6
5.2 Protection against toxins and allergens .....	6
5.3 Photobiological hazards .....	6
5.4 Thermal hazards.....	6
5.5 Mechanical hazards.....	6
5.6 Electrical safety .....	7
5.7 Radiation safety.....	7
5.7.1 Light hazards .....	7
5.8 Gas safety (for gas lasers) .....	7
5.9 Safety in use .....	7
5.10 System hazard analysis.....	8
<b>6</b> Optical requirements .....	8
6.1 Alignment system .....	8
6.2 Fail safe monitoring .....	8
6.3 System calibration .....	8
<b>7</b> System control and performance.....	9
7.1 Software .....	9
<b>8</b> Clinical evaluation.....	9
8.1 Clinical investigation plan .....	9
8.2 Surgical procedure .....	9
8.3 Reporting periods and evaluations .....	9
8.4 Adverse events.....	10
<b>9</b> Test methods.....	10
9.1 Verification of alignment system.....	10
9.1.1 Materials .....	10
9.1.2 Procedure .....	10

	Page
9.2 Verification of the cylinder axis alignment .....	10
9.2.1 Materials .....	10
9.2.2 Procedure .....	11
<b>10</b> Accompanying documents .....	11
<b>11</b> Marking.....	11
<b>Annexes</b>	
<b>A</b> Spectral weighting function for ultraviolet radiation hazard analysis .....	12
<b>B</b> Methods for system calibration.....	14
B.1 Plastic plate ablation and measurement .....	14
B.1.1 Materials .....	14
B.1.2 Procedure .....	14
B.2 Laminated calibration plate method .....	14
B.2.1 Materials .....	14
B.2.2 Procedure .....	15
<b>C</b> Characterization of laser ablation beams and treatment patterns .....	16
C.1 Ablation characteristics of the beam .....	16
C.2 Mathematical models and simulations .....	17
C.3 Validation of ablation algorithm software.....	17
<b>D</b> Guidance on clinical study design of refractive procedures that use laser systems for corneal reshaping.....	18
D.1 General.....	18
D.2 Study objectives .....	18
D.3 Design of the clinical study .....	19
D.4 Study duration .....	19
D.5 Enrollement of subjects .....	19
D.6 Inclusion and exclusion criteria for subject selection.....	20
D.6.1 Inclusion criteria .....	20
D.6.2 Exclusion criteria .....	20
D.7 Examination schedule .....	21
D.8 Evaluations and methodology .....	23
D.8.1 Visual acuity and manifest refraction.....	23
D.8.2 Measurement of intraocular pressure.....	24
D.8.3 Subject questionnaire.....	25
D.8.4 Mesopic pupil size .....	25
D.8.5 Contrast sensitivity .....	25
D.8.5.1 Grating contrast sensitivity testing.....	25

	Page
D.8.6 Low contrast letter acuity testing .....	26
D.8.7 Specular microscopy .....	27
D.9 Adverse device effects/adverse events .....	29
<b>E</b> Statistical sample size considerations .....	30
E.1 Statistical symbols and definitions .....	30
E.2 Calculation of necessary sample sizes .....	31
E.2.1 Sample size based on safety estimates .....	31
E.2.2 Sample size based on effectiveness estimates using noninferiority hypothesis testing .....	32
E.3 Clinical substudies .....	34
E.3.1 Sample size for a contract sensitivity study .....	34
E.3.2 Sample size for endothelial cell density study .....	34
<b>F</b> Presentation of results of clinical studies .....	36
F.1 General .....	36
F.2 Accountability of subjects .....	36
F.3 Refractive stability .....	38
F.4 Safety .....	39
F.5 Effectiveness .....	39
F.6 Retreatment .....	40
<b>G</b> Bibliography .....	41
<b>Tables</b>	
D.1 Recommended postoperative examination schedule .....	22
E.1 Statistical symbols and definitions .....	30
E.2 Normal quantiles to use in equations .....	31
F.1 Accountability by postoperative visit .....	37

**Foreword** (This foreword is not part of American National Standard ANSI Z80.11-2007.)

This American National Standard was developed to address the expressed needs of those members of the ophthalmic community who correct the refractive errors of the human eye using laser refractive correction procedures, those who manufacture the lasers systems for corneal reshaping used to perform these procedures, and those who insure the public interest by ensuring that these systems are made in such a way so that they may perform their function in a safe and effective way when used correctly by those skilled in their use.

It must be realized that correcting the refractive error of the human eye with laser corrective surgery is a medical procedure involving not only a laser system for corneal reshaping but also (1) other devices used during surgery, (2) the assessment of the refractive state of the eye prior to surgery, (3) decisions on the best approach to take for treatment that involve not only the judgment of the physician but the desires of the patient, (4) the postsurgical care, and (5) the effects of healing, known and unknown. This standard only addresses the laser system for corneal reshaping and makes no attempt to standardize the procedure itself.

However, in response to a perceived need, informative annexes have been included in the standard to give guidance on types of clinical testing deemed to be adequate to ensure that the entire procedure is safe and effective. It was felt that a service would be performed for those in the field if this information were to be placed in a public document where it would be readily available to all.

While it is true that the outcome of a laser refractive procedure will not prove acceptable if the laser system for corneal reshaping used to perform it is not adequate for the task, it cannot be assumed that a laser system is inadequate if outcomes are not acceptable, as this may be the result of deficiencies in other important parts of the medical procedure. Thus, no claim is made that, if a laser system for corneal reshaping complies with this standard for the tasks it is designed to perform, that surgical procedures performed with the laser will have acceptable outcomes.

This standard was created by a special working group created by the Z80 Subcommittee on Medical Ophthalmic Devices and included experts in the field of laser refractive correction from the clinical, manufacturing and academic areas of the ophthalmic community and by experts from the regulatory agency given oversight in this field in the United States of America.

This standard contains seven annexes. Annexes A and B are normative and are considered part of the standard. Annexes C through G are informative and are not considered part of this standard.

Suggestions for improvement of this standard will be welcome. They should be sent to the Optical Laboratories Association, P.O. Box 2000, Merrifield, VA 22116-2000.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas C. White, M.D., Chairman  
Quido Cappelli, Vice-Chairman  
Robert Rosenberg, O.D., Secretary  
Daniel Torgersen, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Advance Medical Technologies Association .....	Douglas J. Fortunato Glenn Davies (Alt.) Stanley J. Rogaski (Alt.) Bernie Liebler (Alt.)
American Academy of Ophthalmology .....	Thomas C. White Gerhard Cibis (Alt.) Norman Lanphear (Alt.) Paul F. Vinger (Alt.)
American Academy of Optometry .....	David S. Loshin
American Ceramic Society .....	Lyle Rubin
American Glaucoma Society .....	Herbert Hoover (Alt.) Steven J. Gedde
American Optometric Association .....	Douglas J. Rhee (Alt.) Donald Pitts William J. Benjamin (Alt.) William L. Brown (Alt.) Robert Rosenberg (Alt.) Jeffrey Weaver (Alt.)
American Society of Cataract and Refractive Surgery .....	Stephen Klyce Jack T. Holladay (Alt.) Stephen H. Johnson (Alt.)
Contact Lens Institute .....	Ed Schilling
Contact Lens Manufacturers Association .....	Quido Cappelli Jan Svochak (Alt.)
Department of Veterans Affairs .....	John Townsend Sharon R. Atkin (Alt.)
Federated Cornea Societies .....	Michael W. Belin David Glasser (Alt.)
Food & Drug Administration .....	Donald Calogero Robert Landry (Alt.)
National Association of Optometrists & Opticians .....	Franklin D. Rozak
National Academy of Opticianry .....	James Elkin
Optical Laboratories Association .....	Daniel Torgersen Gregory S. Jacobs (Alt.) Virginia "Susie" Leshner (Alt.) Jonathan Schwartz (Alt.)
Optical Society of America .....	Richard A. Phillips
Opticians Association of America .....	Catherine Langley Tina M. Schott (Alt.)
Prevent Blindness .....	Christine Bradley Jeff Todd (Alt.)
Sunglass Association of America .....	Kenneth L. Frederick Scott Macguffie (Alt.) Rick Van Arnam (Alt.)
US Leader to ISO TC 172/SC7 .....	Charles E. Campbell
Vision Council of America .....	Jeff Endres Steve Drake (Alt.) Neil Roche (Alt.) Dick Whitney (Alt.) Kenneth O. Wood (Alt.)

The Working Group for Laser Systems for Corneal Reshaping, which falls under the Medical Ophthalmic Devices Subcommittee, had the following members who worked on the writing of this standard:

Charles E. Campbell, WG Chair  
Gerhard Cibis, Subcommittee Chair

John J. Alpar MD	Guy M. Kezirian MD	Alan Russell PhD
Michael W. Belin MD	Michael Kliewer PhD	Alex Sacharoff PhD
Don Calagero	Stephen Klyce PhD	Stephen C. Schallhorn MD
Ian Cox PhD	Kitty Lombardo	David Sliney PhD
Bruce Drum PhD	Scott MacCrea MD	Doyle Stulting MD
Daniel Durrie MD	Shirley K. McGarvey	Yoshitaka Suzuki
Malvina Eydelman MD	Mohinder Merchea OD PhD	Larry Thibos PhD
Robert Faaland	Audrey Munnerlyn	Mark Von Tress PhD
Barbara Fant PharmD	Charles Munnerlyn PhD	Thomas C. White MD
Emily Graham	Terrence O'Brien MD	Bonita Wong OD
Gene Hilmantel OD	Mark Odrich MD	
Jack T. Holladay MD	Jan S. Peterson	

---

American National Standard  
for Ophthalmics –

Laser Systems for  
Corneal Reshaping

## 1 Scope and purpose

This standard applies to any laser system whose primary intended use is to alter the shape of the cornea through the removal of corneal tissue, resulting in the improvement of visual performance.

This standard addresses the vocabulary, performance requirements, labeling, and clinical investigations necessary for this type of device.

## 2 Normative references

The following standards contain provisions that, through reference in this text, constitute provisions of this American National Standard. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of the IEC and ISO maintain registers of currently valid International Standards.

ANSI Z136.1-2000, *Safe Use of Lasers*

ANSI Z136.3-2005, *Safe Use of Lasers in Health Care Facilities*

IEC 60601-1:2001, *Medical Electrical Equipment – Part 1: General Requirements for Safety*

IEC 60601-1-1:2000, *Medical Electrical Equipment – Part 1-1: General Requirements for Safety – Collateral Standard: Safety Requirements for Medical Electrical Systems*

IEC 60601-1-2:1993, *Medical Electrical Equipment – Part 1: General Requirements for Safety – 2. Collateral Standard: Electromagnetic Compatibility – Requirements and Tests*

IEC 60601-1-4:2000, *Medical Electrical Equipment – Part 1-4: General Requirements for Safety – Collateral Standard: Programmable Electrical Medical Systems*

IEC 60601-2-22:1995, *Medical Electrical Equipment – Part 2: Particular Requirements for the Safety of Diagnostic and Therapeutic Laser Equipment*

IEC 60825-1:2001, *Safety of Laser Products – Part 1: Equipment Classification, Requirements and User's Guide*

ISO 8598:1996, *Optics and optical instruments – Focimeters*