

American National Standard

*for Ophthalmics –
Corneal Topography Systems –
Standard Terminology, Requirements*



ANSI[®]
Z80.23-2008 (R2013)
(Revision of
ANSI Z80.23-1999)

American National Standard
for Ophthalmics –
Corneal Topography Systems –
Standard Terminology, Requirements

Secretariat
The Vision Council

Approved August 4, 2008
Reaffirmed November 21, 2013

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards -

The Vision Council
Z80 Secretariat
225 Reinekers Lane
Alexandria, VA 22314

Published by

The Vision Council
225 Reinekers Lane
Alexandria, VA 22314

Copyright © 2013 by The Vision Council
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	ii
1 Scope and purpose	1
2 Normative references	1
3 Terminology	2
4 Requirements	8
5 Test methods and test devices	8
6 Accompanying documents	14
7 Marking	14
Tables	
1 Conic section descriptors	7
2 Test surfaces for type testing	10
3 Tolerance levels for test surface measurements	11
4 Analysis zones for accuracy and repeatability testing	13
Figures	
1 Illustration of axial curvature, K_a , axial radius of curvature, r_a , meridional curvature, K_m , and meridional radius of curvature, r_m	6
2 Illustration of the corneal vertex and the apex	7
Annexes	
A Test surfaces for corneal topographers	15
B Standardized displays for corneal topographers	17
C Calculation of area weighting values	20
D Preprocessing of axial elevation data prior to analysis of pared data sets	22

Foreword (This foreword is not part of American National Standard ANSI Z80.23-2008 (R2013).

This American National Standard continues to address the expressed needs of those members of the ophthalmic community who use corneal topography in clinical settings, those who manufacture corneal topographers and those who teach others regarding the use of the information collected by corneal topographers. In particular there continues to be a need for standardization of the terms and definitions used in the field, for standardization of the methods used for characterizing the performance of these instruments and for standardization of displays of corneal topographical information. The experts who worked together to create this standard felt that at this time there is not sufficient consensus within the ophthalmic community to set performance requirements for these instruments beyond those for minimum area measured and measurement sample density. The standard continues to address standardization of the methods for testing these instruments, for assessing their performance, and for reporting the results thus obtained.

The number and type of test surfaces to be used has been changed to include only test surfaces for which the results can be verified. When these surfaces are tilted or rotated the expected surface measurements are easy to predict. These surfaces were considered to be adequate as minimum verification surfaces for corneal topographers; if a corneal topography system can measure these surfaces well, it will be a clinically useful instrument. The method for standardization of color maps has been changed in an effort to improve the user's ability to discern just-noticeable-differences in corneal topography. The user always has the option of using a scale with less resolution but with greater range, as long as the scale recommended in this document is available to be used.

This standard was created by a special working group created by the Z80 Subcommittee on Ophthalmic Instruments and included experts in the field of corneal topography from the clinical, manufacturing and academic areas of the ophthalmic community.

This standard contains four annexes. Annex A is informative and is not considered to be part of this standard. Annexes B, C, and D are normative and are considered to be part of this standard.

Suggestions for improvement of this standard will be welcome. They should be sent to the Vision Council, 225 Reinekers Lane, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas White, M.D., Chairman
Quido Cappelli, Vice-Chairman
Robert Rosenberg, O.D., Secretary
Daniel Torgersen, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Advance Medical Technologies Association	Douglas J. Fortunato Glenn Davies (Alt.) Bernie Liebler (Alt.) Richard Courtney (Alt.)

<i>Organization Represented</i>	<i>Name of Representative</i>
American Academy of Ophthalmology	Thomas C. White Gerhard Cibis (Alt.) Norman Lanphear (Alt.) Paul F. Vinger (Alt.)
American Academy of Optometry	David S. Loshin
American Ceramic Society	Lyle Rubin
American Glaucoma Society	Herbert Hoover (Alt.) Steven J. Gedde
American Optometric Association	Douglas J. Rhee (Alt.) William L. Brown William J. Benjamin (Alt.) Robert Rosenberg (Alt.) Jeffrey Weaver (Alt.)
American Society of Cataract and Refractive Surgery	Stephen Klyce Jack T. Holladay (Alt.) Stephen H. Johnson (Alt.)
Contact Lens Institute	Tom Henteleff
Contact Lens Manufacturers Association	Peter Mathers (Alt.) Quido Cappelli Jan Suochak (Alt.)
Department of Veterans Affairs	John Townsend Sharon R. Atkin (Alt.)
Federated Cornea Societies	Michael W. Belin David Glasser (Alt.)
Food & Drug Administration	Donald Calogero Robert Landry (Alt.) Bruce Drum (Alt.) Robert H. James (Alt.) Dexiu Shi (Alt.)
National Association of Optometrists & Opticians	Franklin D. Rozak Joe Dezenzo (Alt.)
Optical Laboratories Association	Daniel Torgersen Gregory S. Jacobs (Alt.) Susie Leshner (Alt.) Jonathan Schwartz (Alt.)
Optical Society of America	(Representation Vacant)
Opticians Association of America	Tom Hicks Catherine Langley (Alt.)
Prevent Blindness	Christine Bradley Jeff Todd (Alt.)
Sunglass Association of America	Kenneth L. Frederick Scott Macguffie (Alt.) Rick Van Arnam (Alt.)
US Leader to ISO TC 172/SC7	Charles E. Campbell
Vision Council of America	Jeff Endres Ken Wood (Alt.) Steve Drake (Alt.) Neil Roche (Alt.) Dick Whitney (Alt.)

The subcommittee on Ophthalmic Instruments, which developed this standard, had the following members:

William L. Brown, O. D., Ph.D, Chairman	Charles E. Campbell
	Robert Landry
	David Loshin
	Robert Rosenberg
	Thomas White

American National Standard for Ophthalmics – Corneal Topography Systems – Standard Terminology, Requirements

1 Scope and purpose

1.1 Scope

This American National Standard applies to instruments, systems and methods that are intended to measure the shape of the cornea of the human eye over a majority of its central anterior surface. The measurements may be of the curvature of the surface in local areas, three dimensional topographical measurements of the surface or other more global parameters used to characterize the surface. Instruments classified as ophthalmometers or keratometers are not covered by this standard.

1.2 Purpose

This standard defines certain terms that are peculiar to the characterization of the corneal shape so that they may be standardized throughout the field of vision care and have common meaning for all those who have occasion to participate in this area.

This standard sets forth minimum requirements for instruments and systems that fall into the class of corneal topographers.

This standard sets forth tests and verification procedures that will verify that a system or instrument complies with the standard and so qualifies as a corneal topographer in the meaning of this standard.

This standard sets forth certain tests and verification procedures that will allow the verification of capabilities of systems that are beyond the minimum required for corneal topographers.

2 Normative references

The following standard contains provisions that, through reference in this text, constitute provisions of this American National Standard. At the time of publication, the edition indicated was valid. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below.

ANSI Z80.20-1998, *Ophthalmics – Contact Lenses – Standard terminology, tolerances, measurements and physicochemical properties*¹⁾

ANSI Z80.28-2004, *Ophthalmic Instruments – Methods for reporting optical aberrations of eyes*¹⁾

ISO 8429:1986, *Optics and optical instruments – Ophthalmology – Gradual dial scale*¹⁾

¹⁾ For electronic copies of some standards, visit ANSI's Electronic Standards Store (ESS) at www.ansi.org. For printed versions of all these standards, contact Global Engineering Documents, 15 Inverness Way East, Englewood, CO 80112-5704, (800) 854-7179.