

ASME PTC 19.2-2010
[Revision of ASME/ANSI PTC 19.2-1987 (R2004)]

Pressure Measurement

Instruments and Apparatus Supplement

Performance Test Codes

AN AMERICAN NATIONAL STANDARD



INTENTIONALLY LEFT BLANK

ASME PTC 19.2-2010
[Revision of ASME/ANSI PTC 19.2-1987 (R2004)]

Pressure Measurement

Instruments and Apparatus Supplement

Performance Test Codes

AN AMERICAN NATIONAL STANDARD



Three Park Avenue • New York, NY • 10016 USA

Date of Issuance: October 29, 2010

This Code will be revised when the Society approves the issuance of a new edition. There will be no addenda issued to PTC 19.2-2010.

ASME issues written replies to inquiries concerning interpretations of technical aspects of this Code. Periodically certain actions of the ASME PTC Committee may be published as Code Cases. Code Cases and interpretations are published on the ASME Web site under the Committee Pages at <http://cstools.asme.org> as they are issued.

ASME is the registered trademark of The American Society of Mechanical Engineers.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The Standards Committee that approved the code or standard was balanced to assure that individuals from competent and concerned interests have had an opportunity to participate. The proposed code or standard was made available for public review and comment that provides an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not "approve," "rate," or "endorse" any item, construction, proprietary device, or activity.

ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable letters patent, nor assumes any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representative(s) or person(s) affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations of this document issued in accordance with the established ASME procedures and policies, which precludes the issuance of interpretations by individuals.

No part of this document may be reproduced in any form,
in an electronic retrieval system or otherwise,
without the prior written permission of the publisher.

The American Society of Mechanical Engineers
Three Park Avenue, New York, NY 10016-5990

Copyright © 2010 by
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS
All rights reserved
Printed in U.S.A.

CONTENTS

Notice	vi
Foreword	vii
Acknowledgments	vii
Committee Roster	viii
Correspondence With the PTC 19.2 Committee	ix
Section 1 Object and Scope	1
1-1 Object	1
1-2 Scope	1
1-3 Uncertainty	1
Section 2 Definitions and Terms	2
2-1 Introduction	2
2-2 Definitions	2
2-3 Units	3
2-4 Dynamic Measurements	3
2-5 Use of Control and Operating Instrumentation	4
2-6 Two-Phase Fluid Systems	5
Section 3 Measurement Devices	6
3-1 Types of Devices	6
3-2 Pressure Transmitters and Their Applications	12
3-3 Elastic Gauges and Their Applications	31
Section 4 Calibration and Standards	36
Section 5 Measurement Installations	37
5-1 Pressure Taps	37
5-2 Pressure Probes	37
5-3 Connecting Piping	47
5-4 Diaphragm Seals	48
5-5 Installation Effects	55
Section 6 Uncertainties in Pressure Measurement	57
6-1 Introduction	57
6-2 Combined Standard Uncertainty and Expanded Uncertainty	57
6-3 Random Standard Uncertainty	57
6-4 Using Elemental Random Error Sources	57
6-5 Systematic Standard Uncertainty	58
6-6 Propagation of Measurement Uncertainties Into a Result	58
6-7 Uncertainty of Measurements (Example)	58
Figures	
2-2-1 Pressure Terminology	3
3-1.1-1 Typical Pressure Devices	7
3-1.2.1-1 Wheatstone Bridge	7
3-1.2.1-2 Bonded Strain Gauge	8
3-1.2.2-1 Typical Deposited-Thin-Film Strain Gauge	8
3-1.2.3-1 Typical Piezoresistive Pressure Sensor	9
3-1.2.4-1 Typical Variable Capacitive Pressure Sensor	10
3-1.2.5-1 Photoelectric Pressure Sensor	10
3-1.2.6-1 Inductive Pressure Sensor	11
3-1.2.7-1 Preferred Schematic Representation of the Linear Variable Differential Transformer (LVDT)	11

3-1.2.8-1	Piezoelectric Pressure Sensor	11
3-1.2.9-1	Pressure Transducer With Vibrating Element	12
3-2.1.1-1	Typical Flow Installations	15
3-2.1.1-2	Exploded View of Differential-Pressure Transmitter	16
3-2.1.2.1-1	Flange-Mounted Transmitters	17
3-2.1.2.1-2	Flange Transmitter Mounted Directly to Tank Nozzle	18
3-2.1.2.1-3	Open-Tank Installation With Remote-Seal Type of Transmitter	18
3-2.1.2.2-1	Schematic Diagram of Closed-Tank Transmitter Primary	19
3-2.1.2.2-2	Closed-Tank Installation, Dry Leg	20
3-2.1.2.2-3	Closed-Tank Installation, Dry Leg Transmitter Above Datum Line	21
3-2.1.2.2-4	Closed-Tank Installation, Dry Leg Transmitter Below Datum Line	22
3-2.1.2.2-5	Closed-Tank Installation, Wet Leg	23
3-2.1.2.2-6	Closed-Tank Installation, Wet Leg Transmitter Above Datum Line	24
3-2.1.2.2-7	Closed-Tank Installation, Wet Leg Transmitter Below Datum Line	25
3-2.1.2.2-8	Closed-Tank Installation, Dry Leg Transmitter Above Upper Process Tap	26
3-2.1.2.3-1	Repeater-Type Level-Measurement Device	26
3-2.1.3.1-1	Hydrostatic Head Provides One Method of Density Measurement	27
3-2.1.3.1-2	Differential Hydrostatic Head Increases Sensitivity of Density Measurement	27
3-2.1.3.2-1	Common Method of Measuring Density of a Process Liquid	28
3-2.2.1-1	Typical Drum Water Level	28
3-3.1.1-1	Bourdon Gauge	29
3-3.1.1-2	Bourdon Tubes	30
3-3.1.2-1	Bellows Gauge	30
3-3.1.3-1	Slack Diaphragm Gauge	31
4-1	Pressure-Measurement Calibration Hierarchy	36
5-1-1	Tap Geometry	38
5-1-2	Pressure-Tap Flow Field	38
5-1.1-1	Errors for Different Size Taps in Fully Developed Pipe Flow	38
5-1.2-1	Relative Tap Errors as Percent of Dynamic Pressure	39
5-2.1.1-1	Impact Tube	40
5-2.1.1-2	Variation of Total Pressure Indication With Angle of Attack and Geometry for Pitot Tubes	41
5-2.1.2-1	Keil Probe	42
5-2.1.2-2	Total Pressure Location on a Cylinder in a Flow Field	42
5-2.2.1-1	Static Tube	42
5-2.2.1-2	Pitot-Static Tube	43
5-2.2.2-1	Cylindrical Probe, Principle of Operation	43
5-2.2.2-2	Wedge-Type Probe	44
5-2.2.2-3	Spherical- and Cone-Type Probes	44
5-2.2.3-1	Basket Probe	45
5-2.2.3-2	Alternate Basket Probe	45
5-2.2.4-1	Magnitude of Probe-Blockage Effects, Pressure Error	46
5-2.2.4-2	Magnitude of Probe-Blockage Effects, Mach Number	47
5-4.3-1	Seal Assembled With Capillary	49
5-4.3-2	Diaphragm Seals and In-Line Diaphragm Seals Installed	49
5-4.3-3	Diaphragm Seals and Probe Seal Installed	50
5-4.3-4	Threaded Diaphragm Seal	50
5-4.3-5	Flanged Diaphragm Seals	50
5-4.3-6	Integral Flange and Diaphragm Seal	51
5-4.3-7	Separate Flange and Diaphragm Seal	51
5-4.3-8	Pancake Diaphragm Seal With Capillary Connection	51
5-4.3-9	Pancake Diaphragm Seal Installed	52
5-4.3-10	Flange Extension	52
5-4.3-11	Pancake Extension	53
5-4.7-1	Vapor Pressure Curve of a Fill-Fluid	54
5-4.9-1	Response Time	55
5-4.10-1	Installation Consideration of Pressure Instrument and Diaphragm Seals	56

Tables

2-3-1	Pressure Conversion Factors	4
3-1.3-1	Pressure Instrument Summary	13
3-2.1.2.2-1	Seal-Fluid Selection Chart	19
3-2.1.2.2-2	Type of Calibration Required for Various Transmitter Applications	24
3-2.1.3.2-1	Variations in Density for Different Fluids	25
6-7-1	Uncertainty Due to Random Error	59
6-7-2	Throttle-Pressure Uncertainties	59
6-7-3	Exhaust-Pressure Uncertainties	59
6-7-4	Condensate-Flow Uncertainties	59

Form

3-3.6-1	Recording of Gauge-Test Data Sample	35
---------	---	----

Nonmandatory Appendices

A	Piston Gauges	61
B	Manometers	67
C	Low-Absolute-Pressure (Vacuum) Instruments	72
D	References and Bibliography	82

NOTICE

All Performance Test Codes must adhere to the requirements of ASME PTC 1, General Instructions. The following information is based on that document and is included here for emphasis and for the convenience of the user of the Supplement. It is expected that the Code user is fully cognizant of Sections 1 and 3 of ASME PTC 1 and has read them prior to applying this Supplement.

ASME Performance Test Codes provide test procedures that yield results of the highest level of accuracy consistent with the best engineering knowledge and practice currently available. They were developed by balanced committees representing all concerned interests and specify procedures, instrumentation, equipment-operating requirements, calculation methods, and uncertainty analysis.

When tests are run in accordance with a Code, the test results themselves, without adjustment for uncertainty, yield the best available indication of the actual performance of the tested equipment. ASME Performance Test Codes do not specify means to compare those results to contractual guarantees. Therefore, it is recommended that the parties to a commercial test agree before starting the test and preferably before signing the contract on the method to be used for comparing the test results to the contractual guarantees. It is beyond the scope of any Code to determine or interpret how such comparisons shall be made.

FOREWORD

This Instruments and Apparatus Supplement to The American Society of Mechanical Engineers (ASME) Performance Test Codes (PTC) 19 Series provides information on instrumentation and associated procedures for tests involving measurement of pressure. It is intended to promote results consistent with the best engineering knowledge and practice in industry.

The object and scope of any test should be agreed upon in writing by all parties to the test prior to the test.

ASME PTC 2, *Definitions and Values Code* and ASME PTC 19.1, *Test Uncertainty* may be especially useful references when using this Supplement.

The previous Supplement replaced an older version published in 1964. The previous edition was approved by the Board on Performance Test Codes on September 23, 1986, and adopted by the American National Standard Institute (ANSI) as an American National Standard on August 25, 1987.

Subsequent to the 1987 revision, the PTC 19.2 Committee was reactivated to work on the current revision. This revision uses updated pressure-measurement technologies. Obsolete or rarely used pressure-measuring devices were deleted, resulting in a substantially reduced number of pressure-measurement devices. Some of the less frequently used field devices, such as manometers and piston gauges, were moved to the two appendices.

This edition of PTC 19.2, *Pressure Measurement* was approved by the PTC Standards Committee on December 18, 2009, and approved and adopted as a Standard practice of the Society by action of the Board on Standardization and Testing on January 19, 2010. It was also approved as an American National Standard, by the ANSI Board of Standards Review, on April 22, 2010.

ACKNOWLEDGMENTS

The Committee gratefully acknowledges the contribution and leadership role of Charles Doran.

ASME PTC COMMITTEE

Performance Test Codes

(The following is the roster of the Committee at the time of approval of this Code.)

STANDARDS COMMITTEE OFFICERS

M. P. McHale, *Chair*
J. R. Friedman, *Vice Chair*
J. H. Karian, *Secretary*

STANDARDS COMMITTEE PERSONNEL

P. G. Albert, General Electric Co.	S. J. Korellis, Dynegy Generation
R. P. Allen, Consultant	M. P. McHale, McHale & Associates, Inc.
J. M. Burns, Burns Engineering Services	P. M. McHale, McHale & Associates, Inc.
W. C. Campbell, Southern Company Services	T. K. Kirkpatrick, <i>Alternate</i> , McHale & Associates, Inc.
M. J. Dooley, Alstom Power	J. W. Milton, Reliant Energy
J. R. Friedman, Siemens Energy, Inc.	S. P. Nuspl, The Babcock & Wilcox Co.
G. J. Gerber, Consultant	R. R. Priestley, General Electric Co.
P. M. Gerhart, University of Evansville	J. A. Silvaggio, Siemens Demag Delaval Turbomachinery, Inc.
T. C. Heil, <i>Retired</i> , The Babcock & Wilcox Co.	W. G. Steele, Mississippi State University
S. A. Scavuzzo, <i>Alternate</i> , The Babcock & Wilcox Co.	T. L. Toburen, T2E3
R. E. Henry, Sargent & Lundy	G. E. Weber, Midwest Generation, EME
J. H. Karian, The American Society of Mechanical Engineers	J. C. Westcott, Mustan Corp.
D. R. Keyser, Survice Engineering	W. C. Wood, Duke Power Co.

PTC 19.2 COMMITTEE – PRESSURE MEASUREMENT

R. Weissner, <i>Chair</i> , Wika Instrument	J. Conti, Consultant
J. H. Karian, <i>Secretary</i> , The American Society of Mechanical Engineers	J. A. Silvaggio, Siemens Demag Delaval Turbomachinery, Inc.
Z. Akhtar, Bechtel Power	

CORRESPONDENCE WITH THE PTC 19.2 COMMITTEE

General. ASME Codes are developed and maintained with the intent to represent the consensus of concerned interests. As such, users of this Code may interact with the Committee by requesting interpretations, proposing revisions, and attending Committee meetings. Correspondence should be addressed to

Secretary, PTC 19.2 Standards Committee
The American Society of Mechanical Engineers
Three Park Avenue
New York, NY 10016-5990
<http://go.asme.org/Inquiry>

Proposing Revisions. Revisions are made periodically to the Supplement to incorporate changes that appear necessary or desirable, as demonstrated by the experience gained from the application of the Supplement. Approved revisions will be published periodically.

The Committee welcomes proposals for revisions to this Supplement. Such proposals should be as specific as possible, citing the paragraph number(s), the proposed wording, and a detailed description of the reasons for the proposal including any pertinent documentation.

Proposing a Case. Cases may be issued for the purpose of providing alternative rules when justified, to permit early implementation of an approved revision when the need is urgent, or to provide rules not covered by existing provisions. Cases are effective immediately upon ASME approval and shall be posted on the ASME PTC Committee Web page.

Requests for Cases shall provide a Statement of Need and Background Information. The request should identify the Code and the paragraph, figure, or table number(s), and be written as a Question and a Reply in the same format as existing Cases. Requests for Cases should also indicate the applicable Code edition(s) to which the proposed Case applies.

Interpretations. Upon request, the PTC 19.2 Standards Committee will render an interpretation of any requirement of the Code. Interpretations can be rendered only in response to a written request sent to the Secretary of the PTC 19.2 Standards Committee.

The request for interpretation should be clear and unambiguous. It is further recommended that the inquirer submit his request in the following format:

- Subject:** Cite the applicable paragraph number(s) and the topic of the inquiry.
- Edition:** Cite the applicable edition of the Supplement for which the interpretation is being requested.
- Question:** Phrase the question as a request for an interpretation of a specific requirement suitable for general understanding and use, not as a request for an approval of a proprietary design or situation. The inquirer may also include any plans or drawings that are necessary to explain the question; however, they should not contain proprietary names or information.

Requests that are not in this format will be rewritten in this format by the Committee prior to being answered, which may inadvertently change the intent of the original request.

ASME procedures provide for reconsideration of any interpretation when or if additional information that might affect an interpretation is available. Further, persons aggrieved by an interpretation may appeal to the cognizant ASME Committee. ASME does not “approve,” “certify,” “rate,” or “endorse” any item, construction, proprietary device, or activity.

Attending Committee Meetings. The PTC 19.2 Standards Committee holds meetings, which are open to the public. Persons wishing to attend any meeting or telephone conference should contact the Secretary of the PTC 19.2 Standards Committee.

INTENTIONALLY LEFT BLANK

Section 1

Object and Scope

1-1 OBJECT

The object of this Supplement is to give instructions and guidance for the accurate determination of pressure values in support of the ASME Performance Test Codes. The choice of method, instruments, required calculations, and corrections to be applied depends on the purpose of the measurement, the allowable uncertainty, and the characteristics of the equipment being tested.

1-2 SCOPE

The methods for pressure measurement and the protocols used for data transmission are provided in this Supplement. Guidance is given for setting up the instrumentation and determining the uncertainty of the measurement. Information regarding the instrument type,

design, applicable pressure range, accuracy, output, and relative cost is provided.

Information on pressure-measuring devices that are not normally used in field environments is given in the Nonmandatory Appendices A (Piston Gauges), B (Manometers), and C [Low-Absolute-Pressure (Vacuum) Instruments].

1-3 UNCERTAINTY

The methods provided in this Supplement are designed to assist in the evaluation of measurement uncertainty based on current technology and engineering knowledge, taking into account published instrumentation specifications and measurement and application techniques. This Supplement provides guidance in the use of methods to establish the pressure-measurement uncertainty.