



**ASA/ANSI S12.11-2013/Part 1 / ISO
10302-1:2011**

Reaffirmed by ANSI July 13, 2018

AMERICAN NATIONAL STANDARD

**Acoustics – Measurement of airborne noise
emitted and structure-borne vibration induced
by small air-moving devices –
Part 1: Airborne noise measurement
(a nationally adopted international standard)**

Secretariat:

Acoustical Society of America

Approved on April 2, 2013:

American National Standards Institute, Inc.

Abstract

ANSI/ASA S12.11 - 2013/Part 1 / ISO 10302 - 1:2011 specifies methods for measuring the airborne noise emitted by small air - moving devices (AMDs), such as those used for cooling electronic, electrical, and mechanical equipment where the sound power level of the AMD is of interest.

Examples of these AMDs include propeller fans, tube - axial fans, vane - axial fans, centrifugal fans, motorized impellers, and their variations.

This part of this American National Standard describes the test apparatus and methods for determining the airborne noise emitted by small AMDs as a function of the volume flow rate and the fan static pressure developed by the AMD on the test apparatus. It is intended for use by AMD manufacturers, by manufacturers who use AMDs for cooling electronic equipment and similar applications, and by testing laboratories. It provides a method for AMD manufacturers, equipment manufacturers and testing laboratories to obtain comparable results. Results of measurements made in accordance with this part of this American National Standard are expected to be used for engineering information and performance verification, and the methods can be cited in purchase specifications and contracts between buyers and sellers. The ultimate purpose of the measurements is to provide data to assist the designers of electronic, electrical or mechanical equipment which contains one or more AMDs.

Based on experimental data, a method is given for calculating the maximum volume flow rate of the scaled plenum up to which this part of this American National Standard is applicable.

AMERICAN NATIONAL STANDARD

Acoustics – Measurement of airborne noise emitted and structure-borne vibration induced by small air-moving devices – Part 1: Airborne noise measurement (a nationally adopted international standard)

ANSI/ASA S12.11-2013/Part 1 /
ISO 10302-1 :2011

Accredited Standards Committee S12, Noise

Standards Secretariat
Acoustical Society of America
35 Pinelawn Road, Suite 114 E
Melville, NY 11747-3177

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



ANSI/ASA S12.11-2013/Part 1 / ISO 10302-1:2011
(Revision of ANSI/ASA S12.11-2003/Part 1 / ISO 10302:1996 (MOD) (R 2008))

AMERICAN NATIONAL STANDARD

**Acoustics – Measurement of airborne noise emitted
and structure-borne vibration induced by small air-
moving devices – Part 1: Airborne noise measurement
(a nationally adopted international standard)**

Secretariat:

Acoustical Society of America

Approved on April 2, 2013 by:

American National Standards Institute, Inc.

Abstract

ANSI/ASA S12.11-2013/Part 1 / ISO 10302-1:2011 specifies methods for measuring the airborne noise emitted by small air-moving devices (AMDs), such as those used for cooling electronic, electrical, and mechanical equipment where the sound power level of the AMD is of interest.

Examples of these AMDs include propeller fans, tube-axial fans, vane-axial fans, centrifugal fans, motorized impellers, and their variations.

This part of this American National Standard describes the test apparatus and methods for determining the airborne noise emitted by small AMDs as a function of the volume flow rate and the fan static pressure developed by the AMD on the test apparatus. It is intended for use by AMD manufacturers, by manufacturers who use AMDs for cooling electronic equipment and similar applications, and by testing laboratories. It provides a method for AMD manufacturers, equipment manufacturers and testing laboratories to obtain comparable results. Results of measurements made in accordance with this part of this American National Standard are expected to be used for engineering information and performance verification, and the methods can be cited in purchase specifications and contracts between buyers and sellers. The ultimate purpose of the measurements is to provide data to assist the designers of electronic, electrical or mechanical equipment which contains one or more AMDs.

Based on experimental data, a method is given for calculating the maximum volume flow rate of the scaled plenum up to which this part of this American National Standard is applicable.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
Standards Secretariat
35 Pinelawn Road, Suite 114E
Melville, New York 11747-3177
Telephone: 1 (631) 390-0215
Fax: 1 (631) 390-0217
E-mail: asastds@aip.org

© 2013 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

These materials are subject to copyright claims of ISO, ANSI, and ASA. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of the Acoustical Society of America (ASA). All requests pertaining to this standard should be submitted to the ASA at the address above.

Contents

1	Scope	1
2	Normative references	1
3	Terms and definitions.....	2
3.1	General definitions	2
3.2	Acoustical definitions.....	3
3.3	Aerodynamic definitions.....	3
4	Limitations of measurement.....	5
5	Design and performance requirements for test plenum.....	6
5.1	General.....	6
5.2	Test plenum: main assembly.....	6
5.3	Mounting panel assembly.....	7
5.4	Adjustable exit port assembly	7
5.5	Insertion loss of test plenum.....	8
5.6	Instrumentation for static pressure measurement.....	8
6	Installation.....	8
6.1	Installation of test plenum in test room.....	8
6.2	Direction of airflow	9
6.3	Mounting of air-moving device	9
7	Operation of air-moving device.....	9
7.1	Input power	9
7.2	Points of operation (AC and DC air-moving devices).....	10
8	Measurement procedures	11
8.1	General.....	11
8.2	Microphone positions for measurements in an essentially free field over a reflecting plane.....	11
8.3	Preparations for measurements.....	12
8.4	Operational test of air-moving device.....	12
9	Measurement uncertainty.....	13
10	Information to be recorded.....	14
11	Information to be reported.....	14
	Annex A (normative) Micro-fan p - q curve measurement method	26
	A.1 Scope.....	26
	Annex B (informative) Effects of air density	28
	Annex C (informative) Data formats for presentation.....	29
	C.1 General.....	29
	C.2 Example	29
	Annex D (informative) Air-moving device acoustical noise specification	33
	D.1 General.....	33

D.2	Specification formats for a single air-moving device	33
D.3	Specification formats for a lot of air-moving devices.....	33
Annex E (informative) Guidance on the development of information on measurement uncertainty.....		
		34
E.1	General.....	34
E.2	Model function.....	34
E.3	Input quantities and their contributions to measurement uncertainty	36
E.4	Uncertainty budget.....	39
E.5	Combined standard uncertainty and expanded uncertainty.....	40

Tables

Table 1 — Estimated values of the standard deviation of reproducibility of sound power levels of air-moving devices determined in accordance with this part of this American National Standard.....	13
Table 2 — Co-ordinates of hemispherical measurement surface for sources emitting discrete tones (10 measurement heights).....	24
Table D.1 — Example of an uncertainty budget for determination of sound power	40

Figures

Figure 1 — Test plenum (full size)	16
Figure 2 — Test plenum — Film attachment detail.....	17
Figure 3 — Test plenum — Gusset and vibration isolation.....	17
Figure 4 — Test plenum — Pressure ring.....	18
Figure 5 — Mounting panel assembly.....	19
Figure 6 — Adjustable exit port assembly	20
Figure 7 — Adjustable exit port assembly — Aperture plate (stainless steel).....	21
Figure 8 — Adjustable exit port assembly — Slider (stainless steel).....	22
Figure 9 — Schematic correlation between p - q curve versus system impedance curves (not to scale).....	23
Figure 10 — Hemispherical surface — 10 measurement points.....	24
Figure 11 — Coaxial circular five or more paths in parallel planes for microphone traverses in a free field over a reflecting plane	25
Figure 12 — Schematic correlation of applicability of this annex and corresponding noise emissions (not to scale)	27

Foreword

[*This Foreword is for information only, and is not a part of the American National Standard ANSI/ASA S12.11-2013 Part 1/ISO 10302-1:2011 American National Standard Acoustics – Measurement of airborne noise emitted and structure-borne vibration induced by small air-moving devices – Part 1: Airborne noise measurement. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the standard.*]

This standard comprises a part of a group of definitions, standards, and specifications for use in noise. It was developed and approved by Accredited Standards Committee S12 Noise, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S12 is as follows:

Standards, specifications, and terminology in the field of acoustical noise pertaining to methods of measurement, evaluation, and control, including biological safety, tolerance, and comfort, and physical acoustics as related to environmental and occupational noise.

This standard is an identical national adoption of ISO 10302-1:2011. It revises and replaces ANSI/ASA S12.11-2003/Part 1 / ISO 10302:1996 (MOD) (R 2008) *American National Standard Acoustics – Measurement of noise and vibration of small air-moving devices – Part 1: Airborne noise emission*, which was a Modified Nationally Adopted International Standard. ISO 10302-1:2011 was prepared by Technical Committee ISO/TC 43 Subcommittee SC 1, Noise.

In conformance with ANSI and ISO rules, the words "this part of this American National Standard" replace the words "this part of ISO 10302" where they appear in the ISO document, decimal points were substituted in place of the decimal commas used in ISO documents, and American English spelling is used in place of British English spelling.

The ANSI/ASA equivalents for some of the ISO standards referenced in this standard are given below:

- ANSI/ASA S12.51/ISO 3741 is an identical national adoption of ISO 3741;
- ANSI/ASA S12.54/ISO 3744 is an identical national adoption of ISO 3744; and
- ANSI/ASA S12.55/ISO 3745 is an identical national adoption of ISO 3745.

At time of publication, BSR/ASA S12.11-201x/Part 2 / ISO 10302-2:2011 is expected to be approved as the identical national adoption of ISO 10302-2:2011.

ANSI/ASA S12.10-2010/Part 1 is essentially identical to ISO 7779:2010 as they are both adoptions of 10th Edition of ECMA-74 (2008).

At the time this Standard was submitted to Accredited Standards Committee S12, Noise for approval, the membership was as follows:

W.J. Murphy, *Chair*
S.J. Lind, *Vice-Chair*

S.B. Blaeser, *Secretary*

3M Occupational Health & Environmental Safety Division.....E.H. Berger

Acoustical Society of America.....R.D. Hellweg
.....D. Lubman (Alt.)

Air-Conditioning, Heating and Refrigeration Institute.....S.J. Lind
.....D. Abbate (Alt.)

Air Movement & Control Association, Inc	J.A. Brooks
.....	M. Stevens (Alt.)
American Academy of Audiology	T. Ricketts
.....	C. Schweitzer (Alt.)
American Academy of Otolaryngology - Head and Neck Surgery	R.A. Dobie
.....	L.A. Michael (Alt.)
American Industrial Hygiene Association	D. Driscoll
.....	S.N. Hacker (Alt.)
American Speech-Language-Hearing Association	L.A. Wilber
.....	N. DiSarno (Alt.)
Caterpillar, Inc.	K.G. Meitl
Compressed Air and Gas Institute	R.C. Johnson
.....	D.R. Bookshar (Alt.)
Council for Accreditation in Occupational Hearing Conservation	VACANT
.....	L.D. Hager (Alt.)
Emerson Electric – Copeland Corporation	A.T. Herfat
.....	G. Williamson (Alt.)
ETS – Lindgren Acoustic Systems	D. Winker
.....	M. Black (Alt.)
ExxonMobil	B. Moulton
G.R.A.S. Sound & Vibration	J. Soendergaard
.....	B. Schustrich (Alt.)
GE Aviation	R. Majjigi
.....	R. Babbitt (Alt.)
General Motors	D.B. Moore
InfoComm International	R. Derbyshire
.....	J. Bocchiaro (Alt.)
Information Technology Industry Council	W.M. Beltman
.....	J. Rosenberg (Alt.)
Institute of Noise Control Engineering	R.J. Peppin
.....	B. Tinianov (Alt.)
International Safety Equipment Association	J. Birkner
.....	C.Z. Fargo (Alt.)
John Deere	K.B. Washburn
.....	K. Cone (Alt.)

National Council of Acoustical Consultants	J. Erdreich
.....	G.E. Winzer (Alt.)
National Hearing Conservation Association	P.A. Brogan
National Institute for Occupational Safety and Health	W.J. Murphy
.....	E. Zechmann (Alt.)
National Institute of Standards and Technology – National Voluntary Laboratory Accreditation Program	B.A. Sandoval
National Park Service	M. McKenna
.....	K. Fristrup (Alt.)
Noise Control Engineering, Inc.	M.A. Bahtiarian
.....	R. Fischer (Alt.)
Noise Pollution Clearinghouse	L. Blomberg
PCB Group	K. Cox
.....	L. Harbaugh (Alt.)
Power Tool Institute, Inc.	W.D. Spencer
.....	M. Hickok (Alt.)
Schomer and Associates, Inc.	P.D. Schomer
Sperian Hearing Protection, LLC	R. Ghent
.....	B. Witt (Alt.)
U.S. Air Force	R.L. McKinley
.....	H. Gallagher (Alt.)
U.S. Army Aeromedical Research Lab	W.A. Ahroon
U.S. Army Construction Engineering Research Laboratory	M.J. White
.....	M. Swearingen (Alt.)
U.S. Army Public Health Command	C. Stewart
.....	M. Grantham (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	J. Gaston
.....	P. Fedele (Alt.)
U.S. Department of Transportation	K. Grasty
U.S. Naval Surface Warfare Center - Carderock	M. Craun
Université du Québec ETS	J. Voix
.....	F. Laville (Alt.)

Individual Experts of Accredited Standards Committee S12, Noise, were:

P.K. Baade	W.W. Lang	R.J. Peppin
E.H. Berger	D. Lubman	J. Schmitt
B.M. Brooks	D. Michaud	P.D. Schomer
A.J. Campanella	N.P. Miller	L.C. Sutherland
L.S. Finegold	W.J. Murphy	W.R. Thornton
R.D. Godfrey	M.A. Nobile	L.A. Wilber
R.D. Hellweg		G.E. Winzer

Working Group S12/WG 3, Measurement of Noise from Information Technology and Telecommunications Equipment, which assisted Accredited Standards Committee S12, Noise, in the development of this standard, had the following membership.

S. Bard, Chair

D. Ali	R.D. Hellweg	C. Saunder
B.A. Bard	K.X.C. Man	J. Schmitt
E. Baugh	A. Nava	R. Underwood
W.M. Beltman	M.A. Nobile	L. Wittig
J. DeMoss	V. Ojeda	J. Wong
E. Dunens	R.J. Peppin	Y. Xu
	D. Quinlan	

Suggestions for improvements of this standard will be welcomed. They should be sent to Accredited Standards Committee S12, Noise, in care of the Standards Secretariat of the Acoustical Society of America, 35 Pinelawn Road, Suite 114E, Melville, New York 11747-3177. Telephone: 631-390-0215; FAX: 631-390-0217; E-mail: asastds@aip.org.

Introduction

This part of this American National Standard specifies in detail methods for determining and reporting the airborne noise emissions of small air-moving devices (AMDs) used primarily for cooling electronic equipment, such as that for information technology and telecommunications.

To provide compatibility with measurements of acoustical noise emitted by such equipment, this part of this American National Standard uses the noise emission descriptors and sound power measurement methods of ISO 7779. The descriptor of overall airborne noise emission of the AMD under test is the A-weighted sound power level. The one-third-octave-band sound power level is the detailed descriptor of the noise emission. Octave-band sound power levels may be provided in addition to the one-third-octave-band sound power levels.

American National Standard

Acoustics – Measurement of airborne noise emitted and structure-borne vibration induced by small air-moving devices – Part 1: Airborne noise measurement (a nationally adopted international standard)

1 Scope

This part of this American National Standard specifies methods for measuring the airborne noise emitted by small air-moving devices (AMDs), such as those used for cooling electronic, electrical, and mechanical equipment where the sound power level of the AMD is of interest.

Examples of these AMDs include propeller fans, tube-axial fans, vane-axial fans, centrifugal fans, motorized impellers, and their variations.

This part of this American National Standard describes the test apparatus and methods for determining the airborne noise emitted by small AMDs as a function of the volume flow rate and the fan static pressure developed by the AMD on the test apparatus. It is intended for use by AMD manufacturers, by manufacturers who use AMDs for cooling electronic equipment and similar applications, and by testing laboratories. It provides a method for AMD manufacturers, equipment manufacturers and testing laboratories to obtain comparable results. Results of measurements made in accordance with this part of this American National Standard are expected to be used for engineering information and performance verification, and the methods can be cited in purchase specifications and contracts between buyers and sellers. The ultimate purpose of the measurements is to provide data to assist the designers of electronic, electrical or mechanical equipment which contains one or more AMDs.

Based on experimental data, a method is given for calculating the maximum volume flow rate of the scaled plenum up to which this part of this American National Standard is applicable.

2 Normative references

The following referenced documents are indispensable for the application of this standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3741, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Precision methods for reverberation test rooms*

ISO 3744, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Engineering methods for an essentially free field over a reflecting plane*

ISO 3745, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Precision methods for anechoic test rooms and hemi-anechoic test rooms*

ISO 5801:2007, *Industrial fans — Performance testing using standardized airways*