

ASME MFC-16–2014
(Revision of ASME MFC-16–2007)

Measurement of Liquid Flow in Closed Conduits With Electromagnetic Flowmeters

AN AMERICAN NATIONAL STANDARD



**The American Society of
Mechanical Engineers**

ASME MFC-16–2014
(Revision of ASME MFC-16–2007)

Measurement of Liquid Flow in Closed Conduits With Electromagnetic Flowmeters

AN AMERICAN NATIONAL STANDARD



**The American Society of
Mechanical Engineers**

Two Park Avenue • New York, NY • 10016 USA

Date of Issuance: March 14, 2014

This Standard will be revised when the Society approves the issuance of a new edition.

ASME issues written replies to inquiries concerning interpretations of technical aspects of this Standard. Interpretations are published on the Committee Web page and under go.asme.org/InterpsDatabase. Periodically certain actions of the ASME MFC Committee may be published as Cases. Cases are published on the ASME Web site under the MFC Committee Page at go.asme.org/MFCcommittee as they are issued.

Errata to codes and standards may be posted on the ASME Web site under the Committee Pages to provide corrections to incorrectly published items, or to correct typographical or grammatical errors in codes and standards. Such errata shall be used on the date posted.

The Committee Page can be found at go.asme.org/MFCcommittee. There is an option available to automatically receive an e-mail notification when errata are posted to a particular code or standard. This option can be found on the appropriate Committee Page after selecting “Errata” in the “Publication Information” section.

ASME is the registered trademark of The American Society of Mechanical Engineers.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The Standards Committee that approved the code or standard was balanced to assure that individuals from competent and concerned interests have had an opportunity to participate. The proposed code or standard was made available for public review and comment that provides an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not “approve,” “rate,” or “endorse” any item, construction, proprietary device, or activity.

ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable letters patent, nor assumes any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representative(s) or person(s) affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations of this document issued in accordance with the established ASME procedures and policies, which precludes the issuance of interpretations by individuals.

No part of this document may be reproduced in any form,
in an electronic retrieval system or otherwise,
without the prior written permission of the publisher.

The American Society of Mechanical Engineers
Two Park Avenue, New York, NY 10016-5990

Copyright © 2014 by
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS
All rights reserved
Printed in U.S.A.

CONTENTS

Foreword	iv
Committee Roster	v
Correspondence With the MFC Committee	vi
1 Scope	1
2 References	1
3 Definitions and Symbols	1
4 Theory and Measurement Technique	1
5 Flowmeter Descriptions	4
6 Application Considerations	4
7 Equipment Markings	7
8 Calibration	7
Figures	
4-1 Industrial Electromagnetic Flowmeters	2
4.2-1 Examples of Electromagnetic Field (B_0) Variation With Time	3
4.3-1 Examples of Electrodes for an Electromagnetic Flowmeter	3
6.4-1 Electromagnetic Flowmeter System	5
Table	
3.2-1 Symbols	2
Nonmandatory Appendices	
A Added Details Regarding Theory and Measurement Technique	9
B Liner Material Guidelines	11
C Manufacturer-Specified Accuracy	13
D Calculation Examples	15
E Bibliography	16

FOREWORD

This Standard was prepared by Subcommittee 16 of the ASME Committee on the Measurement of Liquid Flow in Closed Conduits. The chair of the subcommittee is indebted to the many individuals who contributed to this document.

Electromagnetic flowmeters were introduced to the process industries in the mid 1950s. They quickly became accepted flowmeters for difficult applications. Subsequent improvements in technology and reductions in cost have transformed these flowmeters into one of the leading contenders for general use in water-based and other electrically conducting liquid applications.

Due to differences in design of the various electromagnetic flowmeters in the marketplace, this Standard cannot address detailed performance limitations in specific applications. It covers issues that are common to all meters, including application considerations.

The flow industry has been changing from the use of the names "primary" and "secondary" to "sensor" and "transmitter." Previous editions of ASME MFC-16 did use primary and secondary in their figures and text. This new edition uses the sensor and transmitter terminology.

Suggestions for improvement of this Standard will be welcomed. They should be sent to The American Society of Mechanical Engineers; Attn: Secretary, MFC Standards Committee; Two Park Avenue; New York, NY 10016-5990.

This revision was approved as an American National Standard on January 28, 2014.

ASME MFC COMMITTEE

Measurement of Fluid Flow in Closed Conduits

(The following is the roster of the Committee at the time of approval of this Standard.)

STANDARDS COMMITTEE OFFICERS

R. J. DeBoom, *Chair*
Z. D. Husain, *Vice Chair*
D. C. Wyatt, *Vice Chair*
C. J. Gomez, *Secretary*

STANDARDS COMMITTEE PERSONNEL

C. J. Blechinger , <i>Honorary Member</i> , Consultant	G. E. Mattingly , The Catholic University of America
R. M. Bough , Rolls-Royce Corp.	R. W. Miller , <i>Honorary Member</i> , R. W. Miller & Associates, Inc.
M. S. Carter , Flow Systems, Inc.	A. M. Quraishi , American Gas Association
R. J. DeBoom , Consultant	W. F. Seidl , <i>Honorary Member</i> , Consultant
D. Faber , <i>Contributing Member</i> , Badger Meter, Inc.	D. W. Spitzer , <i>Contributing Member</i> , Spitzer and Boyes, LLC
C. J. Gomez , The American Society of Mechanical Engineers	R. N. Steven , Colorado Engineering Experiment Station, Inc.
F. D. Goodson , Emerson Process Management — Daniel	J. H. Vignos , <i>Honorary Member</i> , Consultant
Z. D. Husain , Chevron Corp.	D. E. Wiklund , Emerson–Rosemount Measurement — Rosemount
C. G. Langford , <i>Honorary Member</i> , Consultant	J. D. Wright , <i>Contributing Member</i> , National Institute of Standards and Technology
W. M. Mattar , Invensys/Foxboro Co.	D. C. Wyatt , Wyatt Engineering

SUBCOMMITTEE 16 — ELECTROMAGNETIC FLOWMETERS

R. J. DeBoom , <i>Chair</i> , Consultant	S. B. Rogers , Emerson Process Management — Rosemount
C. A. Diederichs , Emerson Process Management — Rosemount	C. J. Rongione , ABB Instrumentation
M. J. Keilty , Endress + Hauser Flowtec AG	D. W. Spitzer , Spitzer and Boyes, LLC
M. M. Lloyd , The Dow Chemical Co.	S. Y. Tung , City of Houston, Public Works and Engineering
W. M. Mattar , Invensys/Foxboro Co.	P. A. Warburton , Yokogawa Corp. of America
R. W. Miller , <i>Contributing Member</i> , R. W. Miller & Associates, Inc.	D. C. Wyatt , Wyatt Engineering
B. K. Rao , Consultant	

CORRESPONDENCE WITH THE MFC COMMITTEE

General. ASME Standards are developed and maintained with the intent to represent the consensus of concerned interests. As such, users of this Standard may interact with the Committee by requesting interpretations, proposing revisions, and attending Committee meetings. Correspondence should be addressed to

Secretary, MFC Standards Committee
The American Society of Mechanical Engineers
Two Park Avenue
New York, NY 10016-5990
<http://go.asme.org/Inquiry>

Proposing Revisions. Revisions are made periodically to the Standard to incorporate changes that appear necessary or desirable, as demonstrated by the experience gained from the application of the Standard. Approved revisions will be published periodically.

The Committee welcomes proposals for revisions to this Standard. Such proposals should be as specific as possible, citing the paragraph number(s), the proposed wording, and a detailed description of the reasons for the proposal, including any pertinent documentation.

Proposing a Case. Cases may be issued for the purpose of providing alternative rules when justified, to permit early implementation of an approved revision when the need is urgent, or to provide rules not covered by existing provisions. Cases are effective immediately upon ASME approval and shall be posted on the ASME Committee Web page.

Requests for Cases shall provide a Statement of Need and Background Information. The request should identify the Standard and the paragraph, figure, or table number(s), and be written as a Question and Reply in the same format as existing Cases. Requests for Cases should also indicate the applicable edition(s) of the Standard to which the proposed Case applies.

Interpretations. Upon request, the MFC Committee will render an interpretation of any requirement of the Standard. Interpretations can only be rendered in response to a written request sent to the Secretary of the MFC Standards Committee at go.asme.org/Inquiry.

The request for interpretation should be clear and unambiguous. It is further recommended that the inquirer submit his/her request in the following format:

Subject: Cite the applicable paragraph number(s) and the topic of the inquiry.
Edition: Cite the applicable edition of the Standard for which the interpretation is being requested.
Question: Phrase the question as a request for an interpretation of a specific requirement suitable for general understanding and use, not as a request for an approval of a proprietary design or situation. The inquirer may also include any plans or drawings that are necessary to explain the question; however, they should not contain proprietary names or information.

Requests that are not in this format may be rewritten in the appropriate format by the Committee prior to being answered, which may inadvertently change the intent of the original request.

ASME procedures provide for reconsideration of any interpretation when or if additional information that might affect an interpretation is available. Further, persons aggrieved by an interpretation may appeal to the cognizant ASME Committee or Subcommittee. ASME does not "approve," "certify," "rate," or "endorse" any item, construction, proprietary device, or activity.

Attending Committee Meetings. The MFC Standards Committee regularly holds meetings that are open to the public. Persons wishing to attend any meeting should contact the Secretary of the MFC Standards Committee. Future Committee meeting dates and locations can be found on the Committee Page at go.asme.org/MFCcommittee.

MEASUREMENT OF LIQUID FLOW IN CLOSED CONDUITS WITH ELECTROMAGNETIC FLOWMETERS

1 SCOPE

This Standard is applicable to industrial electromagnetic flowmeters and their application in the measurement of liquid flow. The electromagnetic flowmeters covered by this Standard utilize an alternating electrical current (AC) or pulsed direct-current (pulsed-DC) to generate a magnetic field in electrically conductive and electrically homogeneous liquids or slurries flowing in a completely filled, closed conduit.

This Standard does not cover the following:

- insertion-type electromagnetic flowmeters
- electromagnetic flowmeters used in surgical, therapeutic, or other health and medical applications
- applications of industrial flowmeters involving nonconductive liquids
- highly conductive liquids (e.g., liquid metals)

2 REFERENCES

The following document forms a part of this Standard to the extent specified herein. The latest edition shall apply.

ISO 13359, Measurement of conductive liquid flow in closed conduits — Flanged electromagnetic flowmeters — Overall length

Publisher: International Organization for Standardization (ISO) Central Secretariat, 1, ch. de la Voie-Creuse, Case postale 56, CH-1211, Genève 20, Switzerland/Suisse

3 DEFINITIONS AND SYMBOLS

3.1 Definitions

accuracy of measurement: closeness of the agreement between the result of a measurement and a true value of the measurand.

NOTE: Accuracy is a qualitative concept; for the quantitative concept, see *uncertainty*.

calibration: the experimental determination of the relationship between the quantity being measured and the device that measures it, usually by comparison with a standard, then (typically) correcting the output of that device to bring it to the desired value, within a specified tolerance, for a particular value of the input.

flowmeter sensor: includes the flow tube, process connections, electromagnetic coils, and electrodes. Flowmeter sensor is also known by other names, e.g., flowmeter sensor device, sensor device, and sensor.

flowmeter transmitter: includes the electronic transmitter, measurement of the emf_v , and, in most cases, the power for the electromagnet coils of the flowmeter sensor.

meter factor: the number determined by liquid calibration that enables the output flow signal to be related to the volumetric flow rate under defined reference conditions; often expressed as the reciprocal of mean *K*-factor.

uncertainty (of measurement): parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand.

verification: provision of objective evidence that a given item fulfills requirements.

EXAMPLE: Use of independent flow calibration to confirm that performance properties and/or legal requirements of a measuring system are met.

3.2 Symbols

See Table 3.2-1.

4 THEORY AND MEASUREMENT TECHNIQUE

Industrial electromagnetic flowmeters are composed of the following basic components (see Fig. 4-1):

(a) a nonmagnetic tube with a nonconductive inner surface

(b) a magnetic field passing through the tube and perpendicular to the axis of the tube at the center of the flow tube

(c) a minimum of two electrodes on opposite sides of the tube in a cross-sectional plane passing through the center of the flow tube, the straight line between these two electrodes being perpendicular to the magnetic field at the center of the flow tube

4.1 Flow-Related Electromotive Force

Faraday's law of induction applied to this physical configuration predicts the generation of an electromotive force (a voltage) between the electrodes when a