



BSI Standards Publication

**Ambient air — Method
for the measurement
of benz[a]anthracene,
benzo[b]fluoranthene,
benzo[j]fluoranthene,
benzo[k]fluoranthene,
dibenz[a,h]anthracene,
indeno[1,2,3-cd]pyrene and
benzo[ghi]perylene**

National foreword

This Published Document is the UK implementation of CEN/TS 16645:2014.

The UK participation in its preparation was entrusted to Technical Committee EH/2/3, Ambient atmospheres.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 81790 8

ICS 13.040.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 April 2014.

Amendments issued since publication

Date	Text affected
------	---------------

TECHNICAL SPECIFICATION
 SPÉCIFICATION TECHNIQUE
 TECHNISCHE SPEZIFIKATION

CEN/TS 16645

April 2014

ICS 13.040.20

English Version

**Ambient air - Method for the measurement of benz[a]anthracene,
 benzo[b]fluoranthene, benzo[j]fluoranthene,
 benzo[k]fluoranthene, dibenz[a,h]anthracene, indeno[1,2,3-
 cd]pyrene and benzo[ghi]perylene**

Air ambiant - Méthode pour la mesure de
 benz[a]anthracène, benzo[b]fluoranthène,
 benzo[j]fluoranthène, benzo[k]fluoranthène,
 dibenz[a,h]anthracène, indeno[1,2,3-cd]pyrène et
 benzo[ghi]perylène

Außenluft - Verfahren zur Messung von Benz[a]anthracen,
 Benzo[b]fluoranthen, Benzo[j]fluoranthen,
 Benzo[k]fluoranthen, Dibenz[a,h]anthracen, Indeno[1,2,3-
 cd]pyren und Benzo[ghi]perylen

This Technical Specification (CEN/TS) was approved by CEN on 16 December 2013 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
 COMITÉ EUROPÉEN DE NORMALISATION
 EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	4
Introduction	5
1 Scope	6
2 Normative references	7
3 Terms and definitions	7
4 Symbols and abbreviations	9
4.1 Symbols	9
4.2 Abbreviations	10
5 Principle of the method	10
6 Requirements	10
6.1 Siting requirements	10
6.2 Sampling requirements	11
6.3 Analysis	11
7 Reagents and gases	13
7.1 Solvents	13
7.2 Gases	13
7.3 External standard	13
7.4 Internal standard	13
7.5 Surrogate standard	13
7.6 Stock standard solution	13
7.7 Certified reference material	13
8 Apparatus	14
8.1 Sampling equipment	14
8.2 Sample preparation/extraction	14
8.3 Laboratory apparatus	14
9 Sampling	15
9.1 Preparation of the equipment before sampling	15
9.2 Handling of filters	15
9.3 Preparation of filters	15
9.4 Sample collection, transport and storage	16
10 Sample preparation	16
10.1 Cleaning of the laboratory apparatus	16
10.2 Extraction	16
11 Analysis	17
11.1 General	17
11.2 HPLC/FLD analysis	17
11.3 GC-MS analysis	18
12 Quantification	19
12.1 HPLC/FLD analysis	19
12.2 GC-MS Analysis	21
12.3 Concentration of the PAH compounds in ambient air	22
13 Quality control	23
13.1 Reagent blank check	23

13.2	Calibration drift check.....	23
13.3	Control solutions.....	23
13.4	Recovery efficiency check	23
13.5	Chromatographic interference check.....	23
13.6	Laboratory filter blank check	24
13.7	Field filter blank check.....	24
13.8	External quality assessment	24
14	Determination of measurement uncertainty	24
14.1	Introduction.....	24
14.2	Parameters contributing to measurement uncertainty	25
14.3	Recommendations for use	27
15	Interferences	27
15.1	General	27
15.2	HPLC/FLD method.....	27
15.3	GC-MS method.....	28
16	Reporting of results	28
Annex A	(informative) Sampling systems with oxidant denuder.....	30
Annex B	(informative) Extraction methods (examples of experimental conditions)	33
Annex C	(informative) Example for clean-up procedure.....	35
Annex D	(informative) Parameters for analysis (examples)	36
Annex E	(informative) Assessment of performance indicators and uncertainty contributions	38
Annex F	(informative) Calculation of uncertainty using interlaboratory comparison and field validation data	47
Bibliography	51

Foreword

This document (CEN/TS 16645:2014) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

The measurement methods specified in this document are based on methods that were worked out during the laboratory and field validation tests for the European Standards EN 15549 [5] (determination of BaP in ambient air) and EN 15980 [6] (determination of the deposition of several particle bound PAH compounds). Many data on the performance of the extraction and analysis of benz[*a*]anthracene (BaA), benzo[*b*]fluoranthene (BbF), benzo[*j*]fluoranthene (BjF), benzo[*k*]fluoranthene (BkF), dibenz[*a,h*]anthracene (DBahA), indeno[1,2,3-*cd*]pyrene (INP) and benzo[*ghi*]perylene (BghiP) in deposition or PM10 samples were gathered. These data do not cover the complete measurement procedure including sampling and consequently enable publication of a Technical Specification instead of a European Standard.

It is the long-term goal to combine this document with EN 15549 [5], once adequate progress has been made in the development of reliably working oxidant denuders.

1 Scope

This Technical Specification specifies a measurement method for the determination of the particle bound polycyclic aromatic hydrocarbon (PAH) compounds benz[a]anthracene (BaA), benzo[b]fluoranthene (BbF), benzo[j]fluoranthene (BjF), benzo[k]fluoranthene (BkF), dibenz[a,h]anthracene (DBahA), indeno[1,2,3-cd]pyrene (INP) and benzo[ghi]perylene (BghiP) in ambient air, which can be used in the framework of Council Directive 2008/50/EC [10] and Directive 2004/107/EC [11]. This document specifies performance characteristics and performance criteria for this measurement method. The performance characteristics of the measurement method are based on a sampling period of 24 h.

This Technical Specification describes a measurement method which comprises sampling of the selected PAH compounds as part of the PM10 particles, sample extraction and analysis by high performance liquid chromatography (HPLC) with fluorescence detector (FLD) or by gas chromatography with mass spectrometric detection (GC-MS). The method is applicable for the measurement of the PAH compounds in the concentration range from approx. 0,04 ng/m³ to approximately 20 ng/m³ for BaA, BbF, BjF, BkF, BaP, INP and BghiP and 0,02 ng/m³ to approximately 2 ng/m³ for DBahA. Table 1 shows examples for concentrations of the compounds (annual mean values) for sampling sites with different characteristics.

Table 1 — Examples of annual mean values of PAH compounds in PM10 at sampling sites with different characteristics (in ng/m³)

Compound	Industrial ^a	Urban background ^b	Traffic ^c	Rural background ^d
BaA	0,85	0,24	0,24	0,06
BbF	2,44	0,62	0,48 ^e	0,16 ^e
BjF	0,89	0,27		
BkF	0,89	0,24	0,17	0,15
BaP	1,15	0,29	0,27	0,13
BghiP	1,31 ^f	0,20 ^g	0,34	0,09
DBahA	0,20	0,10	0,05 ^f	0,07 ^h
INP	1,60	0,43	0,23	0,08

^a Bottrop (Germany, 2010), HPLC/FLD.
^b Mülheim-Styrum (Germany, 2010), HPLC/FLD.
^c London Crystal Palace Parade (UK, 2010), GC-MS.
^d Harwell (UK, 2010), GC-MS.
^e (Bbf+BjF).
^f Wijk aan Zee (The Netherlands, 2011), GC-MS.
^g Rotterdam (The Netherlands, 2011), GC-MS.
^h (DBacA+DBahA).

The lower limit of the applicable range depends on the noise level of the detector and the variability of the laboratory filter blank.