

PD IEC/TR 62354:2014



BSI Standards Publication

General testing procedures for medical electrical equipment

bsi.

...making excellence a habit.™

National foreword

This Published Document is the UK implementation of IEC/TR 62354:2014.

The UK participation in its preparation was entrusted by Technical Committee CH/62, Electrical Equipment in Medical Practice, to Subcommittee CH/62/1, Common aspects of Electrical Equipment used in Medical Practice.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014.
Published by BSI Standards Limited 2014

ISBN 978 0 580 87366 9
ICS 11.040.01

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 October 2014.

Amendments issued since publication

Date	Text affected
-------------	----------------------



TECHNICAL REPORT

General testing procedures for medical electrical equipment

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

PRICE CODE **XH**

ICS 11.040

ISBN 978-2-8322-1856-3

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	6
INTRODUCTION.....	9
1 Scope and object.....	10
2 Normative references	10
3 Terms, definitions, abbreviations and acronyms.....	11
3.1 Terms and definitions.....	11
3.2 Abbreviations and acronyms	12
4 Types of tests.....	12
4.1 GENERAL	12
4.2 Visual inspection.....	13
5 State of the ME EQUIPMENT	13
6 Number of samples.....	13
7 Applicable test items to the clauses of IEC 60601-1	13
8 Sequence of tests.....	13
9 General testing condition.....	14
10 Power sources for tests	15
10.1 General.....	15
10.2 Connection to a separate power source	16
10.3 Connection to an external d.c. power source.....	16
10.4 Source of power for ME EQUIPMENT	16
10.5 SUPPLY MAINS for testing ME EQUIPMENT	16
11 Measurement and test equipment.....	17
11.1 General requirements	17
11.2 Accuracy.....	18
11.3 Safety criteria for selection	18
11.4 Calibration	18
12 Treatments of unit symbols and measured values.....	19
13 PROCEDURES for testing, including particular conditions	19
13.1 General.....	19
13.2 Tests to be performed by inspection	20
13.3 Measurements and tests performed on non-energized equipment.....	41
13.4 Measurements and tests for equipment that is operating.....	96
Annex A (informative) Sequence of testing.....	165
A.1 Sequence of testing (IEC 60601-1:1988).....	165
A.2 Sequence of testing (IEC 60601-1:2005).....	166
Annex B (informative) Information typically required for product safety testing (Guide)	168
B.1 Purpose	168
B.2 Description	168
B.3 Intended use environment.....	168
B.4 Construction	168
B.5 List of safety-related components and relevant approvals	168
B.6 Test system	168
B.7 Power	168
B.8 Grounding.....	169

B.9	Modes of operation; configurations	169
B.10	Failure modes	169
B.11	RISK ANALYSIS according with ISO 14971	169
B.12	Software	169
B.13	Auxiliary equipment	169
B.14	Transformers and chokes	169
Annex C (informative)	Testing and measuring equipment	170
Annex D (informative)	Suitable measuring supply circuits	171
Annex E (informative)	Preventive maintenance	174
E.1	General	174
E.2	Cleaning and disinfection	174
E.3	Preventive maintenance checklist	174
E.4	OPERATOR checks	174
Annex F (informative)	Test probes	175
Annex G (informative)	Index of tests (IEC 60601-1:2005 clauses order)	177
Annex H (informative)	Index of tests for an INTERNALLY POWERED EQUIPMENT – battery only – (IEC 60601-1:2005 clauses order)	179
Annex I (informative)	Index of tests (IEC 60601-1:2005 alphabetic order)	181
Annex J (informative)	Index of tests for an INTERNALLY POWERED EQUIPMENT – battery only – (IEC 60601-1:2005 alphabetic order)	183
Annex K (informative)	Production line tests	185
K.1	Production-line dielectric voltage-withstand test	185
K.2	Production-line grounding-continuity test	186
K.3	Production-line EARTH LEAKAGE CURRENT test	186
K.4	Recommended features for specific test equipment	186
Annex L (informative)	Evaluation of the laboratory power source characteristics	189
L.1	Purpose	189
L.2	Application	189
L.3	Definitions	190
L.4	Testing	191
Annex M (informative)	Traceability of calibrations and calibration intervals	194
M.1	Purpose	194
M.2	Traceability of calibrations	194
M.3	Calibration intervals for test equipment requiring calibration	194
Annex N (informative)	Guidance for preparation, attachment, extension, use of thermocouples and acceptance of thermocouple wire	196
N.1	General	196
N.2	Preparation	196
N.3	Placement	196
N.4	Attachment	197
N.5	Extension	198
N.6	Use	199
N.7	Acceptance of thermocouples wire	199
Annex O (informative)	Guideline for safe laboratory work	201
O.1	BASIC SAFETY guidelines for working with test instruments	201
O.2	Basic guidelines for performing safety tests	203
O.3	Basic guidelines regarding test personnel and test areas	205

O.4	Contents of a documented safe environment for working in a testing laboratory	206
	Bibliography.....	207
	Index of defined terms	209
	Figure 1 – Area of the bottom of an ENCLOSURE as specified in 11.3 b) 1)	37
	Figure 2 – Baffle	37
	Figure 3 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 1	51
	Figure 4 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 2	51
	Figure 5 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 3	52
	Figure 6 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 4	52
	Figure 7 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 5	52
	Figure 8 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 6	52
	Figure 9 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 7	53
	Figure 10 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 8	53
	Figure 11 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 9	54
	Figure 12 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 10	54
	Figure 13 – Human body test mass	73
	Figure 14 – Application of test voltage to bridged PATIENT CONNECTIONS (common mode) for DEFIBRILLATION-PROOF APPLIED PARTS	103
	Figure 15 – Application of test voltage to individual PATIENT CONNECTIONS (differential mode) for DEFIBRILLATION-PROOF APPLIED PARTS	104
	Figure 16 – Application of test voltage to test the delivered defibrillation energy (energy reduction test)	108
	Figure 17 – Example of a measuring device and its frequency characteristics	110
	Figure 18 – Measuring circuit for the EARTH LEAKAGE CURRENT of CLASS I equipment, with or without APPLIED PARTS	112
	Figure 19 – Measuring circuit for the TOUCH CURRENT	115
	Figure 20 – Measuring circuit for the PATIENT LEAKAGE CURRENT from the PATIENT CONNECTION to earth	117
	Figure 21 – Measuring circuit for the total PATIENT LEAKAGE CURRENT with all PATIENT CONNECTIONS of all APPLIED PARTS of the same type (TYPE B APPLIED PARTS, TYPE BF APPLIED PARTS or TYPE CF APPLIED PARTS) connected together	118
	Figure 22 – Measuring circuit for the PATIENT LEAKAGE CURRENT via the PATIENT CONNECTION(S) of an F-TYPE APPLIED PART to earth caused by an external voltage on the PATIENT CONNECTION(S)	120
	Figure 23 – Measuring circuit for the PATIENT LEAKAGE CURRENT from PATIENT CONNECTION(S) to earth caused by an external voltage on a metal ACCESSIBLE PART that is not PROTECTIVELY EARTHED	121
	Figure 24 – Measuring circuit for the PATIENT LEAKAGE CURRENT from PATIENT CONNECTION(S) to earth caused by an external voltage on a SIGNAL INPUT/OUTPUT PART	123
	Figure 25 – Measuring circuit for the PATIENT AUXILIARY CURRENT	125
	Figure 26 – Ratio between hydraulic test pressure and maximum permissible working pressure	129
	Figure 27 – Spark ignition test apparatus	136
	Figure 28 – Maximum allowable current I as a function of the maximum allowable voltage U measured in a purely resistive circuit in an OXYGEN RICH ENVIRONMENT	139

Figure 29 – Maximum allowable voltage U as a function of the capacitance C measured in a capacitive circuit used in an OXYGEN RICH ENVIRONMENT	139
Figure 30 – Maximum allowable current I as a function of the inductance L measured in an inductive circuit in an OXYGEN RICH ENVIRONMENT.....	140
Figure D.1 – Measuring supply circuit with one side of the SUPPLY MAINS at approximately earth potential.....	171
Figure D.2 – Measuring supply circuit with SUPPLY MAINS approximately symmetrical to earth potential.....	171
Figure D.3 – Measuring supply circuit for polyphase ME EQUIPMENT specified for connection to a polyphase SUPPLY MAINS	171
Figure D.4 – Measuring supply circuit for single-phase ME EQUIPMENT specified for connection to a polyphase SUPPLY MAINS	172
Figure D.5 – Measuring supply circuit for ME EQUIPMENT having a separate power supply unit or intended to receive its power from another equipment in an ME SYSTEM.....	172
Figure F.1 – Standard test finger	175
Figure F.2 – Test hook.....	176
Figure F.3 – Test pin.....	176
Figure F.4 – Ball-pressure test apparatus	176
Figure N.1 – Thermocouple preparation	196
Figure N.2 – Securing of thermocouples	197
Figure N.3 – Example of confinement of a thermocouple.....	198
Figure N.4 – Example where thermocouple connectors need not be used	199
Table 1 – Units outside the SI units system that may be used	19
Table 2 – Tests to be performed by inspection.....	20
Table 3 – NOMINAL cross-sectional area of conductors of a POWER SUPPLY CORD	31
Table 4 – Acceptable perforation of the bottom of an ENCLOSURE	38
Table 5 – Measurements and tests performed on non-energized equipment	41
Table 6 – Testing of cord anchorages	56
Table 7 – Acceptable gaps ^a	62
Table 8 – Drop height	87
Table 9 – Test torques for rotating controls.....	93
Table 10 – Measurements and tests for equipment that is operating	96
Table 11 – Allowable maximum temperatures for skin contact with ME EQUIPMENT APPLIED PARTS	147
Table 12 – Allowable maximum temperatures for ME EQUIPMENT parts that are likely to be touched.....	147
Table 13 – Allowable maximum temperatures of parts.....	148
Table 14 – Temperature limits of motor windings	148
Table 15 – Maximum motor winding steady-state temperature	149
Table 16 – Maximum allowable temperatures of transformer windings under overload and short-circuit conditions at 25 °C (\pm 5 °C) ambient temperature.....	149
Table 17 – Test current for transformers	162
Table C.1 – IEC 60601-1:1988+AMD 1:1991 and AMD 2:1995.....	170
Table D.1 – Legends of symbols for Figure D.1 to Figure D.5	173
Table L.1 – Method for testing a single phase laboratory power source.....	191

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**GENERAL TESTING PROCEDURES
FOR MEDICAL ELECTRICAL EQUIPMENT**
FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62354, which is a technical report, has been prepared by subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2009. This edition constitutes a technical revision intended to align the guidance in this technical report with Amendment 1 to IEC 60601:2005. Several tests have been updated and additional test procedures added. The following tests have been added or significantly revised:

- 13.2.1 RISK MANAGEMENT PROCESS
- 13.2.4 Durability and legibility of marking
- 13.2.5 Battery markings

- 13.2.8 POTENTIAL EQUALIZATION TERMINAL
- 13.2.14 USABILITY OF ME EQUIPMENT
- 13.3.1 Humidity preconditioning
- 13.3.2 Impedance of PE connection
- 13.3.7 CREEPAGE DISTANCES and AIR CLEARANCES
- 13.3.12 Instability (in transport position; excluding transport; from horizontal and vertical forces and from unwanted lateral movement)
- 13.3.13 Castors and wheels (Force for propulsion, movement over a threshold)
- 13.3.14 Safety catch evaluation
- 13.3.17 Overflow
- 13.3.18 Spillage
- 13.3.23 Impact
- 13.3.14 Drop impact
- 13.3.25 Rough handling
- 13.3.27 Actuating parts of controls
- 13.3.28 Construction of transformers
- 13.4.1 ESSENTIAL PERFORMANCE – Functional
- 13.4.3 Voltage mismatch
- 13.4.4 Limitation of voltage, current or energy
- 13.4.5 DEFIBRILLATION-PROOF APPLIED PART protection
- 13.4.6 Energy reduction
- 13.4.7 EARTH LEAKAGE CURRENT
- 13.4.9 PATIENT LEAKAGE CURRENT
- 13.4.14 Sound pressure level measurements
- 13.4.16 X-radiation (ionizing radiation) measurement
- 13.4.20 Interruption of power supply
- 13.4.28 Rechargeable battery overcharge/discharge
- 13.4.29 Mains transformers

This technical report is intended to be read in conjunction with IEC 60601-1:1988 (including the collateral provisions of IEC 60601-1-1:2000) and IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
62A/936/DTR	62A/947/RVC

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

In this technical report, the terms defined in Clause 2 of IEC 60601-1:1988 or Clause 3 of IEC 60601-1:2005 are printed in SMALL CAPITALS.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

INTRODUCTION

IEC/TR 60513, *Fundamental aspects of safety standards for medical electrical equipment* published by IEC sub-committee 62A provided the basis for inclusion of the test methods for ME EQUIPMENT in the safety standards.

"Technical requirements and test methods are interrelated elements of product standards and should always be considered together.

Product standards should identify where medically informed judgements are required in deciding whether a particular requirement applies.

Wherever possible, the standards should contain test specifications for completely and clearly checking compliance with the technical requirements. In some cases, a compliance statement such as 'visual inspection', 'manual testing' or similar is adequate for this purpose if such a method gives an accurate assessment.

It should be easy to recognize which test methods apply to each technical requirement. Appropriate headings should designate the appropriate test and a reference should be made to the clause containing the requirement. This also applies for references which are made to other relevant test standards."

It was deemed necessary to support IEC 60601-1 with guidelines for general testing PROCEDURES for MEDICAL ELECTRICAL EQUIPMENT.

In developing the test PROCEDURES, the advice given in IEC/TR 60513 and ISO/IEC Guide 51 was considered as follows:

- a) test results should be reproducible within defined limits. When considered necessary, the test method should incorporate a statement as to its limit of uncertainty;
- b) where the sequence of tests can influence the results, the correct sequence should be specified.

There is also growing support for the idea that all the test PROCEDURES for ME EQUIPMENT should be found within one international standard.

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*, highlights the need for a single series of requirements covering test PROCEDURES.

IEC/TR 60513 includes a major new principle referring to testing:

"In specifying minimum safety requirements, provision is made for assessing the adequacy of the design PROCESS where this provides an appropriate alternative to the application of laboratory testing with specific pass/fail criteria, (e.g. in assessing the safety of new technologies such as programmable electronic systems)."

GENERAL TESTING PROCEDURES FOR MEDICAL ELECTRICAL EQUIPMENT

1 Scope and object

This technical report applies to MEDICAL ELECTRICAL EQUIPMENT (as defined in Subclauses 3.63 of IEC 60601-1:2005 and 2.2.15 of IEC 60601-1:1988), hereinafter referred to as ME EQUIPMENT.

The object of this technical report is to provide guidance on general testing PROCEDURES according to IEC 60601-1:1988 (including the collateral provisions of IEC 60601-1-1:2000) and IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60086-4, *Primary batteries – Part 4: Safety of lithium batteries*

IEC 60127-1, *Miniature fuses – Part 1: Definitions for miniature fuses and general requirements for miniature fuse-links*

IEC 60252-1, *AC motor capacitors – Part 1: General – Performance, testing and rating – Safety requirements – Guide for installation and operation*

IEC 60364-4-41, *Low voltage electrical installations – Part 4-41: Protection for safety – Protection against electric shock*

IEC 60417, *Graphical symbols for use on equipment*. Available from: <http://www.graphical-symbols.info/equipment>

IEC/TR 60513, *Fundamental aspects of safety standards for medical electrical equipment*

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*
IEC 60529:1989/AMD1:1999¹

IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety*²
IEC 60601-1:1998/AMD1:1991
IEC 60601-1:1998/AMD2:1995

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012³

¹ A consolidated version 2.1 (2001) exists that includes IEC 60529:1989 and its Amendment 1:1999.

² The second edition of IEC 60601-1, cancelled and replaced by the third edition in 2005.

³ A consolidated version 3.1 (2012) exists that includes IEC 60601-1:2005 and its Amendment 1:2012.