



**ANSI/ASA S3.55-2015/Part 3 /
IEC 60318-3:2014**

Reaffirmed June 5, 2020

AMERICAN NATIONAL STANDARD

**Electroacoustics – Simulators of Human Head
and Ear – Part 3: Acoustic Coupler for the
Calibration of Supra-aural Earphones
Used in Audiometry**

(a nationally adopted international standard)

Secretariat:

Acoustical Society of America

Approved on May 15, 2015

American National Standards Institute, Inc.

This standard specifies an acoustic coupler for objective and reproducible measurements of supra-aural audiometric earphones in the frequency range from 125 Hz to 8000 Hz. The results of these measurements can be used for specifying reference equivalent threshold sound pressure levels (RETSPL) for the calibration of audiometers.

AMERICAN NATIONAL STANDARD

Electroacoustics – Simulators of Human Head and Ear – Part 3: Acoustic Coupler for the Calibration of Supra-aural Earphones Used in Audiometry (a nationally adopted international standard)

ANSI/ASA S3.55-2015/Part 3 /
IEC 60318-3 :2014

Accredited Standards Committee S3, Bioacoustics

Standards Secretariat
Acoustical Society of America
1305 Walt Whitman Road, Suite 300
Melville, NY 11747

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



AMERICAN NATIONAL STANDARD

Electroacoustics – Simulators of Human Head and Ear – Part 3: Acoustic Coupler for the Calibration of Supra-aural Earphones Used in Audiometry (a nationally adopted international standard)

Secretariat:

Acoustical Society of America

Approved on May 15, 2015 by:

American National Standards Institute, Inc.

Abstract

This standard specifies an acoustic coupler for objective and reproducible measurements of supra-aural audiometric earphones in the frequency range from 125 Hz to 8000 Hz. The results of these measurements can be used for specifying reference equivalent threshold sound pressure levels (RETSPL) for the calibration of audiometers.

AMERICAN NATIONAL STANDARDS ON BIOACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
1305 Walt Whitman Road, Suite 300
Melville, New York 11747
Telephone: 1 (631) 390-0215
Fax: 1 631-923-2875
E-mail: asastds@acousticalsociety.org

© 2015 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

Contents

1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Construction	1
	4.1 General	1
	4.2 Cavity dimensions	2
	4.3 Static pressure equalization	4
	4.4 Calibrated pressure type microphone	4
5	Coupling of earphone to acoustic coupler	5
6	Calibration	6
	6.1 Reference environmental conditions	6
	6.2 Method of calibration	6

Figures

Figure 1 – Dimensions of acoustic coupler	3
Figure 2 – Coupling of earphone to coupler	5

Tables

Table 1 – Height of the coupler as a function of the acoustic volume of the microphone	4
Table 2 – Values of maximum permitted uncertainties U_{max} for a level of confidence of approximately 95 %	7

Foreword

[This Foreword is for information only, and is not a part of the American National Standard ANSI/ASA S3.55-2015/Part 3 / IEC 60318-3:2014 American National Standard Electroacoustics – Simulators of Human Head and Ear – Part 3: Acoustic Coupler for the Calibration of Supra-aural Earphones Used in Audiometry. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the standard.]

This standard comprises a part of a group of definitions, standards, and specifications for use in bioacoustics. It was developed and approved by Accredited Standards Committee S3 Bioacoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S3 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the fields of psychological and physiological acoustics, including aspects of general acoustics which pertain to biological safety, tolerance and comfort.

This standard is a nationally adopted international standard (NAIS). It was undertaken as part of the revision of ANSI/ASA S3.7-1995 (R2008), Method for Coupler Calibration of Earphones, which, in addition to coupler calibration methods, contains detailed information about the audiometric ear, the 6cc coupler, and the 2cc coupler. Several years ago, IEC 60318 was reorganized into several parts, which now include the same information about the aforementioned couplers (IEC 60318, Parts 1, 3, and 5, respectively). Given that the manufacture of these couplers has changed little if at all in more than 25 years, the fact that the IEC and ANSI/ASA specifications for these couplers are essentially identical, and the fact that manufacturers of these couplers do not produce different versions of the devices to meet alternative versions of the standards, harmonization was deemed appropriate. The first step in this process has been the move of the detailed coupler specifications from the previous version of ANSI/ASA S3.7 to NAIS ANSI/ASA S3.55, Parts 1, 3, and 5, which correspond directly to their IEC 60318 counterparts. References in other standards that previously pointed to ANSI/ASA S3.7 for a particular coupler will now point to the appropriate NAIS ANSI/ASA S3.55 part instead as these documents are revised and updated. The next revision of ANSI/ASA S3.7 will focus solely on the calibration methods for earphones, and will also point to the appropriate NAIS ANSI/ASA S3.55 part, where the detailed coupler information now resides.

This standard is an identical national adoption of IEC 60318-3 Ed. 2.0 b:2014 *Electroacoustics – Simulators of human head and ear – Part 3: Acoustic coupler for the calibration of supra-aural earphones used in audiometry*, which was prepared by IEC/TC 29. However, in conformance with ANSI and IEC rules, the words “this part of ANSI/ASA S3.55 / IEC 60318” replace the words “this part of IEC 60318” where they appear in the IEC document, decimal points were substituted in place of the decimal commas used in IEC documents, and American English spelling is used in place of British English spelling. At the time this Standard was submitted to Accredited Standards Committee S3, Bioacoustics for approval, the membership was as follows:

C.J. Struck, *Chair*
P.B. Nelson, *Vice-Chair*

S.B. Blaeser, *Secretary*

Acoustical Society of AmericaC.J. Struck
..... P.B. Nelson (Alt.)

American Academy of Audiology C. Schweitzer
..... T. Ricketts (Alt.)

American Academy of Otolaryngology, Head and Neck Surgery, Inc.	R.A. Dobie
.....	A. Kim (Alt.)
American Industrial Hygiene Association	T.K. Madison
.....	D. Driscoll (Alt.)
American Speech-Language-Hearing Association (ASHA)	L.A. Wilber
.....	N. DiSarno (Alt.)
Beltone/GN Resound	S. Petrovic
Council for Accreditation in Occupational Hearing Conservation (CAOHC)	L.D. Hager
Diagnostic Group	T. McColley
.....	P. Dobrowski (Alt.)
ETS – Lindgren Acoustic Systems	S. Dunlap
.....	D. Winker (Alt.)
Etymotic Research, Inc.	M.C. Killion
.....	J.K. Stewart (Alt.)
Food and Drug Administration	S-C Peng
G.R.A.S. Sound & Vibration	B. Schustrich
.....	J. Soendergaard (Alt.)
Hearing Industries Association	A. Bopp
.....	M. Jones (Alt.)
National Electrical Manufacturers Association, Signaling Protection & Communication Section (NEMA – 3SB)	J. McNamara
.....	R. Reiswig (Alt.)
National Hearing Conservation Association	G.L. Poling
.....	R. Danielson (Alt.)
National Institute for Occupational Safety and Health (NIOSH)	M.R. Stephenson
.....	W.J. Murphy (Alt.)
National Institute of Standards and Technology	R.P. Wagner
.....	S. Fick (Alt.)
National Park Service	M. McKenna
.....	K. Frstrup (Alt.)
Natus Medical, Inc.	Y. Hekimoglu
Ocean Conservation Research	M. Stocker
Starkey Hearing Technologies	D.A. Preves

U.S. Air Force..... R.L. McKinley
..... B.D. Simpson (Alt.)

U.S. Army Aeromedical Research Laboratory W.A. Ahroon

U.S. Army CERL..... D.K. Delaney
..... M.J. White (Alt.)

U.S. Army Research Laboratory, Human Research and Engineering Directorate A.A. Scharine
..... A. Foots (Alt.)

University of Cincinnati Animal Audiology Clinic/Bioacoustics Lab..... P.M. Scheifele
..... D.K. Brown (Alt.)

Individual Experts of the Accredited Standards Committee S3, Bioacoustics, were:

A.J. Brammer	A.J. Campanella	P.D. Schomer
R.F. Burkard	R.L. McKinley	C.J. Struck
M. Burkhard	V. Nedzelnitsky	L.A. Wilber

Working Group S3/WG 37, Couplers, Ear Simulators, and Earphones, which assisted Accredited Standards Committee S3, Bioacoustics, in the development of this standard, had the following membership.

	C.J. Struck, Chair	
M. Alexander	C.B. King	D.A. Preves
T.H. Burns	B. Kruger	J. Soendergaard
G.J. Frye	B. Mathews	J.K. Stewart

Suggestions for improvements of this standard will be welcomed. They should be sent to Accredited Standards Committee S3, Bioacoustics, in care of the Standards Secretariat of the Acoustical Society of America, 1305 Walt Whitman Road, Suite 300, Melville, New York 11747. Telephone: 631-390-0215; FAX: 631-923-2875; E-mail: asastds@acousticalsociety.org.

American National Standard

Electroacoustics – Simulators of Human Head and Ear – Part 3: Acoustic Coupler for the Calibration of Supra-aural Earphones Used in Audiometry

1 Scope

This part of ANSI/ASA S3.55 / IEC 60318 specifies an acoustic coupler for the measurement of supra-aural audiometric earphones in the frequency range from 125 Hz to 8,000 Hz.

The sound pressure developed by an earphone is not, in general, the same in the coupler as in a person's ear. However, the acoustic coupler can be used as an objective and reproducible means of measuring the output of supra-aural earphones. It can be used for specifying reference equivalent threshold sound pressure levels (RETSPL) for the calibration of audiometers.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61094-1, *Measurement microphones – Part 1: Specifications for laboratory standard microphones*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purpose of this document, the following definition applies:

3.1

acoustic coupler

device for measuring the acoustic output of sound sources where the sound pressure is measured by a calibrated microphone coupled to the source by a cavity of predetermined shape and volume which does not necessarily approximate the acoustical impedance of the normal human ear

4 Construction

4.1 General

The coupler consists essentially of a cylindrical cavity whose acoustic transfer impedance is determined by the volume of air in the cavity and its dimensions (see 4.2). A microphone with a diaphragm having high acoustic impedance is located in the base of the cylindrical cavity.

The coupler shall be made of a material that has no negative influences on its performance. For example it should be acoustically hard and dimensionally stable. The general construction of the coupler and mounting of the microphone shall aim to reduce the response to vibration of any earphone or to sound outside the cavity.