

MM21

Genomic Copy Number Microarrays for Constitutional Genetic and Oncology Applications

This guideline provides recommendations for validation, verification, performance, and interpretation of nucleic acid microarrays used for cytogenetic applications to measure copy number imbalances and loss of heterozygosity. Both constitutional and oncology applications are addressed.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeals Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeals, documented in the CLSI Standards Development Policies and Processes, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute
950 West Valley Road, Suite 2500
Wayne, PA 19087 USA
P: +1.610.688.0100
F: +1.610.688.0700
www.clsi.org
standard@clsi.org

Genomic Copy Number Microarrays for Constitutional Genetic and Oncology Applications

Hutton M. Kearney, PhD, FACMG
Shashikant Kulkarni, MS, PhD, FACMG
Karen E. Bijwaard, MS, RAC, MB(ASCP)
Madhuri R. Hegde, PhD
Lisa Kalman, PhD
Leming Shi, PhD
Richard Shippy, MS
Sarah T. South, PhD, FACMG
Erik C. Thorland, PhD
Patrik Vitazka, MD, PhD, FACMG
Emily S. Winn-Deen, PhD
Joanne M. Yeakley, PhD

Abstract

Clinical and Laboratory Standards Institute document MM21—*Genomic Copy Number Microarrays for Constitutional Genetic and Oncology Applications* discusses nucleic acid microarray technologies for diagnostic testing that a growing number of medical laboratories have adopted. The different types of microarrays and their uses in various types of laboratories have grown tremendously. MM21 specifically addresses validation, verification, performance, and interpretation of nucleic acid microarrays used for cytogenetic applications to measure copy number imbalances and loss of heterozygosity. Both constitutional and oncology applications are discussed.

Clinical and Laboratory Standards Institute (CLSI). *Genomic Copy Number Microarrays for Constitutional Genetic and Oncology Applications*. 1st ed. CLSI guideline MM21 (ISBN 1-56238-915-7 [Print]; ISBN 1-56238-916-5 [Electronic]). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2015.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org. If you or your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: customerservice@clsi.org; Website: www.clsi.org.



MM21, 1st ed.

Copyright ©2015 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedure manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. *Genomic Copy Number Microarrays for Constitutional Genetic and Oncology Applications*. 1st ed. CLSI guideline MM21. Wayne, PA: Clinical and Laboratory Standards Institute; 2015.

Reaffirmed:

January 2020

ISBN 1-56238-915-7 (Print)
ISBN 1-56238-916-5 (Electronic)
ISSN 1558-6502 (Print)
ISSN 2162-2914 (Electronic)

Volume 35, Number 15

Committee Membership

Consensus Committee on Molecular Methods

Frederick S. Nolte, PhD,
D(ABMM), F(AAM)
Chairholder
Medical University Hospital
Authority
USA

Barbara Zehnbaue, PhD,
FACMG, FACB
Vice-Chairholder
Centers for Disease Control and
Prevention
USA

Sherry A. Dunbar, PhD
Luminex Corporation
USA

Helen Fernandes, PhD
New York Presbyterian Hospital
USA

Charles E. Hill, MD, PhD
Emory University Hospital
USA

Thomas J. Lenk, PhD
Celera Diagnostics
USA

Ronald M. Przygodzki, MD
Department of Veterans Affairs
USA

Yi-Wei Tang, MD, PhD,
D(ABMM)
Memorial Sloan Kettering Cancer
Center
USA

Zivana Tezak-Fragale, PhD
FDA Center for Devices and
Radiological Health
USA

Document Development Committee on Microarrays for Cytogenetics and Oncology

Hutton M. Kearney, PhD,
FACMG
Co-Chairholder
Mayo Clinic
USA

Shashikant Kulkarni, MS, PhD,
FACMG
Co-Chairholder
Washington University School of
Medicine
USA

Karen E. Bijwaard, MS, RAC,
MB(ASCP)
FDA Center for Devices and
Radiological Health
USA

Madhuri R. Hegde, PhD
Emory University School of
Medicine
USA

Lisa Kalman, PhD
Centers for Disease Control and
Prevention
USA

Leming Shi, PhD
Fudan University
China

Richard Shippy, MS
Illumina, Inc.
USA

Sarah T. South, PhD, FACMG
University of Utah and ARUP
Laboratories
USA

Patrik Vitazka, MD, PhD, FACMG
GeneDx
USA

Joanne M. Yeakley, PhD
BioSpyder Technologies
USA

Staff

Clinical and Laboratory Standards
Institute
USA

Luann Ochs, MS
Senior Vice President – Operations

Marcy Hackenbrack, MCM,
M(ASCP)
Project Manager

Megan L. Tertel, MA, ELS
Editorial Manager

Joanne P. Christopher, MA
Editor

Alexander B. Phucas
Editor

Acknowledgment

CLSI, the Consensus Committee on Molecular Methods, and the Document Development Committee on Microarrays for Cytogenetics and Oncology gratefully acknowledge the following volunteers for their important contributions to the development of this document:

Erik C. Thorland, PhD
Mayo Clinic
USA

Emily S. Winn-Deen, PhD
Rx Dx Advisors, Inc.
USA

Contents

Abstract.....	i
Committee Membership.....	iii
Foreword.....	vii
Chapter 1: Introduction.....	1
1.1 Scope.....	1
1.2 Background.....	1
1.3 Standard Precautions.....	2
1.4 Terminology.....	2
Chapter 2: Genomic Microarray Validation, Verification, and Testing Process.....	7
2.1 Selecting a Microarray Methodology.....	8
2.2 Selecting Microarray Samples.....	13
2.3 Validating and Verifying Microarrays.....	15
2.4 Interpreting Results.....	18
2.5 Patient Reports and Disclaimers.....	24
Chapter 3: Quality Management Considerations for Microarrays.....	29
3.1 Reference DNA Sets and Quality.....	29
3.2 Quality Control Evaluation for Microarrays.....	29
3.3 Postexamination Considerations.....	30
3.4 Qualification of New Reagent Lots.....	31
Chapter 4: Conclusion.....	31
Chapter 5: Supplemental Information.....	31
References.....	32
The Quality Management System Approach.....	36
Related CLSI Reference Materials.....	37

Foreword

MM21 was developed to provide more targeted guidance on molecular microarrays. It focuses on the appropriate performance, validation, verification, and interpretation of nucleic acid microarrays primarily used for cytogenetic applications to measure copy number (CN) imbalances.

When the first guideline on molecular microarrays was published (see CLSI document MM12¹) nucleic acid microarrays were not a major part of medical laboratory test options. Today, many medical laboratories have adopted a variety of testing platforms and, as a result, the types of arrays and laboratories using arrays have increased tremendously. After reviewing CLSI document MM12,¹ the subject matter experts determined that the document, in its current form, is quite broad in scope. To allow for the most clinically relevant applications to be addressed in greater depth, CLSI document MM12¹ was split into three documents, with each document focusing on a particular field of clinical genetics:

- CN and absence of heterozygosity detection arrays (MM21)
- Expression arrays and methylation profiling (see CLSI document MM12¹)
- Nucleic acid microarrays for use in microbiology/immunology laboratories (eg, pathogen profiling) (see CLSI document MM22²)

NOTE: Mandates are occasionally allowed in CLSI guidelines, in cases in which the document development committee feels strongly that a particular action is either required or prohibited, or when a guideline addresses provisions based on regulations. Throughout MM21, the use of the term “must” was evaluated by the document development committee and deemed appropriate because the uses are either 1) based on a requirement or 2) indicative of a necessary step to ensure patient safety or proper fulfillment of a procedure.

Key Words

Array-based comparative genomic hybridization, chromosomal microarrays, genomic copy number microarrays, whole genome testing

Genomic Copy Number Microarrays for Constitutional Genetic and Oncology Applications

Chapter 1: Introduction

This chapter includes:

- Document scope and applicable exclusions
- Background information pertinent to the document content
- Standard Precautions information
- Terms and definitions used in the document
- “Note on Terminology” that highlights particular use and/or variation in use of terms and/or definitions
- Abbreviations and acronyms used in the document

1.1 Scope

This guideline provides recommendations for the appropriate performance, validation or verification, and interpretation of nucleic acid microarrays used primarily for cytogenetic applications to measure copy number (CN) imbalances, traditional array-based comparative genomic hybridization (aCGH), and single nucleotide polymorphism (SNP) arrays for CN imbalances and absence of heterozygosity (AOH). Both constitutional (prenatal/postnatal) and oncology applications are addressed.

The intended users of this guideline are clinical genetics laboratorians who perform cytogenetics and molecular genetics testing.

This guideline:

- Is not intended for research laboratorians
- Is not intended to provide guidance to manufacturers
- Does not address methylation arrays, RNA expression microarrays, resequencing and genotyping arrays not intended for CN detection, microarrays for the diagnosis and monitoring of infectious diseases, or non-nucleic acid microarrays (eg, protein arrays)

1.2 Background

Chromosome abnormalities play a significant role in human genetic diseases. Abnormalities may involve chromosome number (eg, aneuploidy characterized by extra or missing chromosomes) and/or chromosome structure (eg, structural aberrations such as deletion, duplication, inversion, translocation, ring chromosome, isochromosome, and marker chromosome). Chromosomal structural defects are balanced (eg, reciprocal translocations, Robertsonian translocation) or, more often, they are unbalanced, associated with CN gains or losses in the DNA. Historically, conventional cytogenetic techniques (eg, G-