

ASME PVHO-1–2016
(Revision of ASME PVHO-1–2012)

Safety Standard for Pressure Vessels for Human Occupancy

AN AMERICAN NATIONAL STANDARD



**The American Society of
Mechanical Engineers**

ASME PVHO-1-2016
(Revision of ASME PVHO-1-2012)

Safety Standard for Pressure Vessels for Human Occupancy

AN AMERICAN NATIONAL STANDARD



**The American Society of
Mechanical Engineers**

Two Park Avenue • New York, NY • 10016 USA

Date of Issuance: August 29, 2016

The next edition of this Standard is scheduled for publication in 2019.

ASME issues written replies to inquiries concerning interpretations of technical aspects of this Standard. Periodically certain actions of the ASME PVHO Committee may be published as Cases. Cases and interpretations are published on the ASME Web site under the Committee Pages at <http://cstools.asme.org/> as they are issued.

Errata to codes and standards may be posted on the ASME Web site under the Committee Pages to provide corrections to incorrectly published items, or to correct typographical or grammatical errors in codes and standards. Such errata shall be used on the date posted.

The Committee Pages can be found at <http://cstools.asme.org/>. There is an option available to automatically receive an e-mail notification when errata are posted to a particular code or standard. This option can be found on the appropriate Committee Page after selecting “Errata” in the “Publication Information” section.

ASME is the registered trademark of The American Society of Mechanical Engineers.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The Standards Committee that approved the code or standard was balanced to assure that individuals from competent and concerned interests have had an opportunity to participate. The proposed code or standard was made available for public review and comment that provides an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not “approve,” “rate,” or “endorse” any item, construction, proprietary device, or activity.

ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable letters patent, nor assumes any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representative(s) or person(s) affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations of this document issued in accordance with the established ASME procedures and policies, which precludes the issuance of interpretations by individuals.

No part of this document may be reproduced in any form,
in an electronic retrieval system or otherwise,
without the prior written permission of the publisher.

The American Society of Mechanical Engineers
Two Park Avenue, New York, NY 10016-5990

Copyright © 2016 by
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS
All rights reserved
Printed in U.S.A.

CONTENTS

Foreword	vii
Committee Roster	viii
Correspondence With the PVHO Committee	x
Summary of Changes	xii
Section 1 General Requirements	1
1-1 Introduction	1
1-2 Scope	1
1-3 Exclusions	1
1-4 User Requirements	1
1-5 Manufacturer’s Data Report	2
1-6 Materials	2
1-7 Design and Fabrication Requirements	2
1-8 Pressure Relief Devices	13
1-9 Marking	13
1-10 Nonmetallic Materials and Toxicity Off-Gas Testing	14
1-11 Risk Analysis	15
Section 2 Viewports	16
2-1 General	16
2-2 Design	16
2-3 Material	68
2-4 Fabrication	73
2-5 Inspection	74
2-6 Marking	77
2-7 Pressure Testing	77
2-8 Installation of Windows in Chambers	79
2-9 Repair of Damaged Windows Prior to Being Placed in Service	79
2-10 Guidelines for Application of the Requirements of Section 2	81
Section 3 Quality Assurance for PVHO Manufacturers	85
3-1 General	85
3-2 Responsibilities	85
Section 4 Piping Systems	86
4-1 General	86
4-2 Material Requirements	87
4-3 Design of Components	89
4-4 Selection and Limitations of Piping Components	89
4-5 Selection and Limitations of Piping Joints	90
4-6 Supports	91
4-7 Inspection	91
4-8 Testing	92
4-9 Systems	93
Section 5 Medical Hyperbaric Systems	99
5-1 General	99
5-2 PVHO System Design	99
5-3 Gas Systems	99
5-4 Control Systems and Instrumentation	100
5-5 Environmental Systems	100

Section 6	Diving Systems	101
6-1	General	101
6-2	Design	102
6-3	Pressure Boundary	104
6-4	Systems	105
6-5	Testing	110
6-6	Quick-Acting Closures for Diving Bells and Emergency Evacuation Systems	110
Section 7	Submersibles	112
7-1	General	112
7-2	Pressure Boundary	113
7-3	Piping	114
7-4	Electrical Systems	114
7-5	Life Support	115
7-6	Fire Protection	116
7-7	Navigation	116
7-8	Communications	117
7-9	Instrumentation	117
7-10	Buoyancy, Stability, Emergency Ascent, and Entanglement	117
7-11	Emergency Equipment	118
Figures		
1-7.13.1-1	Geometry of Cylinders	7
1-7.13.1-2	Stiffener Geometry	8
1-7.13.1-3	Sections Through Rings	8
1-7.13.5-1	Values of t/R_0 and L_c/R_0	12
1-9-1	Form of Nameplate, U.S. Customary	14
1-9-2	Form of Nameplate, Metric	14
2-2.2.1-1	Standard Window Geometries — Part 1	18
2-2.2.1-2	Standard Window Geometries — Part 2	19
2-2.2.1-3	Standard Window Geometries — Part 3	20
2-2.2.1-4	Standard Window Geometries — Part 4	21
2-2.5.1-1	Short-Term Critical Pressure of Flat Disk Acrylic Windows — Part 1	26
2-2.5.1-2	Short-Term Critical Pressure of Flat Disk Acrylic Windows — Part 2	27
2-2.5.1-3	Short-Term Critical Pressure of Flat Disk Acrylic Windows — Part 3	28
2-2.5.1-4	Short-Term Critical Pressure of Conical Frustum Acrylic Windows — Part 1	29
2-2.5.1-5	Short-Term Critical Pressure of Conical Frustum Acrylic Windows — Part 2	30
2-2.5.1-6	Short-Term Critical Pressure of Spherical Sector Acrylic Windows — Part 1	31
2-2.5.1-7	Short-Term Critical Pressure of Spherical Sector Acrylic Windows — Part 2	32
2-2.5.1-8	Short-Term Critical Pressure of Cylindrical Acrylic Windows Pressurized Internally — Part 1	34
2-2.5.1-9	Short-Term Critical Pressure of Cylindrical Acrylic Windows Pressurized Internally — Part 2	35
2-2.5.1-10	Short-Term Critical Pressure of Cylindrical Acrylic Windows Pressurized Externally	36
2-2.5.1-11	Short-Term Elastic Buckling of Cylindrical Acrylic Windows Between Supports Under External Hydrostatic Pressure — Part 1	37
2-2.5.1-12	Short-Term Elastic Buckling of Cylindrical Acrylic Windows Between Supports Under External Hydrostatic Pressure — Part 2	38
2-2.5.1-13	Short-Term Elastic Buckling of Cylindrical Acrylic Windows Between Supports Under External Hydrostatic Pressure — Part 3	39
2-2.5.1-14	Short-Term Critical Pressure of Hyperhemispherical and NEMO-Type Acrylic Windows — Part 1	40

2-2.5.1-15	Short-Term Critical Pressure of Hyperhemispherical and NEMO-Type Acrylic Windows — Part 2	41
2-2.10.1-1	Seat Cavity Requirements — Conical Frustum Window, Spherical Sector Window With Conical Edge, and Flat Disk Window	45
2-2.10.1-2	Seat Cavity Requirements — Double Beveled Disk Window	46
2-2.10.1-3	Seat Cavity Requirements — Spherical Sector Window With Square Edge	47
2-2.10.1-4	Seat Cavity Requirements — Hemispherical Window With Equatorial Flange	48
2-2.10.1-5	Seat Cavity Requirements — Cylindrical Window	49
2-2.10.1-6	Seat Cavity Requirements — Hyperhemispherical Window	50
2-2.10.1-7	Seat Cavity Requirements — NEMO Window (Standard Seat)	51
2-2.10.1-8	Seat Cavity Requirements — NEMO Window (Seat With Extended Cyclic Fatigue Life)	52
2-2.11.10-1	Bevels on Window Edges — Flat Disk Windows, Conical Frustum Windows, Spherical Sector Windows, Hyperhemispheres	54
2-2.11.10-2	Bevels on Window Edges — Flanged Hemispherical Window, Spherical Sector Window With Square Edge, External Pressure and Internal Pressure of Cylindrical Windows	55
2-2.11.11-1	Acceptable Configurations for Clear Viewport Retaining Covers	57
2-2.14.11-1	Dimensional Tolerances for Penetrations in Acrylic Windows	61
2-2.14.15-1	Dimensional Tolerances for Inserts in Acrylic Windows	63
2-2.14.16-1	Typical Shapes of Inserts	64
2-2.14.22-1	Seal Configurations for Inserts in Acrylic Windows	66
2-2.14.24-1	Restraints for Inserts in Acrylic Windows	67
4-9.14.2-1	Flow Diagram of Apparatus for Measuring the Concentration of Hydrocarbons in a Stream of Air or Other Gas After It Has Passed Through a Test Hose	98
6-4.5.2.2-1	Placement and Design of Markings of Hyperbaric Rescue Systems Designed to Float in Water	109
6-4.5.2.2-2	Markings of Hyperbaric Rescue Systems Designed to Float in Water	109
Tables		
1-10-1	Conversion Factor, F_p (for PVHO Occupation Exceeding 8 hr)	15
2-2.3.1-1	Conversion Factors for Acrylic Flat Disk Windows	23
2-2.3.1-2	Conversion Factors for Acrylic Conical Frustum Windows and Double Beveled Disk Windows	23
2-2.3.1-3	Conversion Factors for Acrylic Spherical Sector Windows With Conical Edge, Hyperhemispherical Windows With Conical Edge, and NEMO-Type Windows With Conical Edge	24
2-2.3.1-4	Conversion Factors for Acrylic Spherical Sector Windows With Square Edge and Hemispherical Windows With Equatorial Flange	24
2-2.3.1-5	Conversion Factors for Acrylic Cylindrical Windows	25
2-2.3.2-1	Conical Frustum Windows for Design Pressures in Excess of 10,000 psi (69 MPa)	25
2-2.14.13-1	Specified Values of Physical Properties for Polycarbonate Plastic	62
2-2.14.13-2	Specified Values of Physical Properties for Cast Nylon Plastic	62
2-3.4-1	Specified Values of Physical Properties for Each Lot	69
2-3.4-2	Specified Values of Physical Properties for Each Casting	71
2-4.5-1	Annealing Schedule for Acrylic Windows	75
4-2.1.1-1	Maximum Allowable Stress Values for Seamless Pipe and Tube Materials Not Listed in Nonmandatory Appendix A of ASME B31.1	87
4-7.1-1	Mandatory Minimum Nondestructive Examinations for Pressure Welds in Piping Systems for Pressure Vessels for Human Occupancy	92
4-9.14.2-1	Maximum Allowable Concentration of Hydrocarbons in Air Passing Through Hose	98

Forms		
GR-1	Manufacturer’s Data Report for Pressure Vessels for Human Occupancy	5
GR-1S	Manufacturer’s Data Report Supplementary Sheet	6
VP-1	Fabrication Certification for Acrylic Windows	17
VP-2	Acrylic Window Design Certification	59
VP-3	Material Manufacturer’s Certification for Acrylic	70
VP-4	Material Testing Certification for Acrylic	72
VP-5	Pressure Testing Certification	78
VP-6	Acrylic Window Repair Certification	82
Mandatory Appendices		
I	Reference Codes, Standards, and Specifications	119
II	Definitions	121
Nonmandatory Appendices		
A	Design of Supports and Lifting Attachments	127
B	Recommendations for the Design of Through-Pressure Boundary Penetrations	128
C	Recommended Practices for Color Coding and Labeling	131
D	Guidelines for the Submission of a PVHO Case for the Use of Nonstandard Designs, Materials, and Construction	132
E	Guidelines for Preparing a PVHO Performance-Based Case for Flexible Chambers	138
F	Useful References	153

FOREWORD

Early in 1971, an ad hoc committee was formed by action of the ASME Codes and Standards Policy Board to develop design rules for pressure vessels for human occupancy. The importance of this task was soon recognized, and the ASME Safety Code Committee on Pressure Vessels for Human Occupancy (PVHO) was established in 1974 to continue the work of the ad hoc committee. Initially, this committee was to confine its activity to the pressure boundary of such systems. It was to reference existing ASME Boiler and Pressure Vessel Code (BPVC) sections, insofar as practicable, adapting them for application to pressure vessels for human occupancy. The common practice hitherto had been to design such chambers in accordance with Section VIII, Division 1 of the ASME BPVC; however, a number of important considerations were not covered in those rules. Among these were requirements for viewports and the in-service use of pressure relief valves, and special material toughness requirements. This Standard provides the necessary rules to supplement that section, and also Section VIII, Division 2 of the BPVC. The user is expected to be familiar with the principles and application of the Code sections.

BPVC criteria furnish the baseline for design. In PVHO-1, design temperature is limited to 0°F to 150°F (−18°C to 66°C). Supporting structure and lifting loads are given special attention. Certain design details permitted by Section VIII are excluded. A major addition is the inclusion of design rules for acrylic viewports (Section 2). The formulation of rules for these vital and critical appurtenances was one of the reasons for establishing the PVHO Committee. Finally, all chambers designed for external pressure are required to be subjected to an external pressure hydrostatic test or pneumatic test.

The 2007 edition was completely rewritten and reformatted from the 2002 edition. Section 1, General Requirements, is intended to be used for all PVHOs, regardless of use. The rules for external pressure design were expanded to include unstiffened and ring-stiffened cylinders, in addition to spheres. Other additions included sections pertaining to application-specific PVHOs. Sections were included for medical hyperbaric systems, diving systems, submersibles, and quality assurance. The piping section was expanded. Where possible, Mandatory Appendices were incorporated into the body of the document. All Forms were revised to reflect the document (PVHO-1), an abbreviation denoting the corresponding section (e.g., General Requirements is GR), and the form number within that Section. An example is PVHO-1 Form GR-1.

The 2012 edition included expansions made to the General Requirements, Viewports, and Diving Systems Sections.

The 2016 edition includes additional expansions made to the General Requirements, Viewports, Medical Hyperbaric Systems, and Diving Systems Sections. It includes a new Nonmandatory Appendix for preparing PVHO performance-based Cases for flexible chambers. There is continuing work being accomplished by the Subcommittees in the areas of PVHOs using nonstandard materials, including nonmetallic PVHOs. A companion document (PVHO-2) that covers in-service guidelines for PVHOs has been published.

The 2016 edition of PVHO-1 was approved and adopted by the American National Standards Institute as meeting the criteria as an American National Standard on January 20, 2016. Previous editions were published in 1977, 1981, 1984, 1987, 1993, 1997, 2002, 2007, and 2012.

ASME PRESSURE VESSELS FOR HUMAN OCCUPANCY COMMITTEE

(The following is the roster of the Committee as of November 10, 2015.)

STANDARDS COMMITTEE OFFICERS

G. K. Wolfe, *Chair*
J. Witney, *Vice Chair*
G. E. Moino, *Secretary*

STANDARDS COMMITTEE PERSONNEL

M. W. Allen , Microbaric Oxygen Systems, LLC	G. E. Moino , The American Society of Mechanical Engineers
J. E. Crouch , Southwest Research Institute	H. Pauli , DNV GL
W. F. Crowley, Jr. , Aerospace & Undersea Support Services, LLC	J. P. Hierholzer , <i>Alternate</i> , DNV GL
W. Davison , Oxyheal Health Group	S. D. Reimers , Reimers Systems, Inc.
B. Faircloth , FMS Engineering, LLC	G. Richards , Blanson Ltd.
E. G. Fink , Fink Engineering Ltd.	T. C. Schmidt , Lockheed Martin
M. A. Frey , Naval Sea Systems Command	J. S. Selby , S.O.S. Medical Group Ltd.
T. R. Galloway , Naval Sea Systems Command	P. Selby , <i>Alternate</i> , S.O.S. Medical Group Ltd.
G. P. Jacob , Navy Experimental Diving Unit	J. C. Sheffield , International ATMO, Inc.
C. B. Kemper III , Kemper Engineering Services, LLC	K. A. Smith , U.S. Coast Guard
W. Kohnen , Hydrospace Group, LLC	R. C. Smith , Naval Facilities Engineering Service Center
J. D. Lawrence , U.S. Coast Guard	D. Talati , Sechrist Industries, Inc.
P. A. Lewis , Hyperbaric Technologies, Inc.	R. Thomas , American Bureau of Shipping (ABS)
J. R. Maison , Adaptive Computer Technology, Inc.	M. R. Walters , Oceaneering International, Inc.
T. T. Marohl , Consultant	J. Witney , Atlantis Submarines International
	G. K. Wolfe , Southwest Research Institute

SUBCOMMITTEE ON DESIGN AND PIPING

T. C. Schmidt , <i>Chair</i> , Lockheed Martin	B. Humberstone , Global Diving and Salvage
G. Richards , <i>Vice Chair</i> , Blanson Ltd.	G. P. Jacob , Navy Experimental Diving Unit
M. W. Allen , Microbaric Oxygen Systems, LLC	C. B. Kemper III , Kemper Engineering Services, LLC
F. Burman , DAN Southern Africa	K. A. Wohlfeil , <i>Alternate</i> , Kemper Engineering Services, LLC
W. F. Crowley, Jr. , Aerospace & Undersea Support Services, LLC	R. Kumar , Reimers Systems, Inc.
W. Davison , Oxyheal Health Group	P. A. Lewis , Hyperbaric Technologies, Inc.
R. K. Dixit , Reimers Systems, Inc.	S. D. Reimers , Reimers Systems, Inc.
B. Faircloth , FMS Engineering, LLC	D. A. Renear , Aqua-Air Industries, Inc.
P. Forte , Woods Hole Oceanographic Institution	C. D. Johnstone , <i>Alternate</i> , Blanson Ltd.
M. A. Frey , Naval Sea Systems Command	J. S. Selby , S.O.S. Medical Group Ltd.
T. R. Galloway , Naval Sea Systems Command	P. Selby , <i>Alternate</i> , S.O.S. Medical Group Ltd.
R. M. Webb , <i>Alternate</i> , Naval Sea Systems Command	R. Thomas , American Bureau of Shipping (ABS)
C. Gaumont , Gaumont Medical Gaumont	M. R. Walters , Oceaneering International, Inc.
	J. Witney , Atlantis Submarines International

SUBCOMMITTEE ON DIVING SYSTEMS

T. R. Galloway , <i>Chair</i> , Naval Sea Systems Command	J. D. Lawrence , U.S. Coast Guard
M. R. Walters , <i>Vice Chair</i> , Oceaneering International, Inc.	T. Gilman , <i>Alternate</i> , U.S. Coast Guard
M. W. Allen , Microbaric Oxygen Systems, LLC	P. A. Lewis , Hyperbaric Technologies, Inc.
W. F. Crowley, Jr. , Aerospace & Undersea Support Services, LLC	T. T. Marohl , Consultant
W. Davison , Oxyheal Health Group	H. Pauli , DNV GL
B. Faircloth , FMS Engineering, LLC	D. A. Renear , Aqua-Air Industries, Inc.
E. G. Fink , Fink Engineering Ltd.	J. S. Selby , S.O.S. Medical Group Ltd.
B. Humberstone , Global Diving and Salvage	P. Selby , <i>Alternate</i> , S.O.S. Medical Group Ltd.
C. B. Kemper III , Kemper Engineering Services, LLC	K. A. Smith , U.S. Coast Guard
K. A. Wohlfeil , <i>Alternate</i> , Kemper Engineering Services, LLC	R. Thomas , American Bureau of Shipping (ABS)

SUBCOMMITTEE ON GENERAL REQUIREMENTS

M. A. Frey, *Chair*, Naval Sea Systems Command
M. W. Allen, *Vice Chair*, Microbaric Oxygen Systems, LLC
J. E. Crouch, Southwest Research Institute
T. R. Galloway, Naval Sea Systems Command
R. M. Webb, *Alternate*, Naval Sea Systems Command

G. P. Jacob, Navy Experimental Diving Unit
S. D. Reimers, Reimers Systems, Inc.
J. S. Selby, S.O.S. Medical Group Ltd.
P. Selby, *Alternate*, S.O.S. Medical Group Ltd.
G. K. Wolfe, Southwest Research Institute

SUBCOMMITTEE ON MEDICAL HYPERBARIC SYSTEMS

M. W. Allen, *Chair*, Microbaric Oxygen Systems, LLC
W. Davison, *Vice Chair*, Oxyheal Health Group
T. Dingman, Healogics
K. W. Evans, Perry Baromedical
E. G. Fink, Fink Engineering Ltd.
C. Foreman, U.S. Food and Drug Administration
W. T. Gurnee, Oxyheal Health Group

C. B. Kemper III, Kemper Engineering Services, LLC
K. A. Wohlfeil, *Alternate*, Kemper Engineering Services, LLC
P. A. Lewis, Hyperbaric Technologies, Inc.
H. Pauli, DNV GL
S. D. Reimers, Reimers Systems, Inc.
J. C. Sheffield, International ATMO, Inc.
D. Talati, Sechrist Industries, Inc.

SUBCOMMITTEE ON POST CONSTRUCTION

R. C. Smith, *Chair*, Naval Facilities Engineering Service Center
W. Davison, *Vice Chair*, Oxyheal Health Group
M. W. Allen, Microbaric Oxygen Systems, LLC
J. E. Crouch, Southwest Research Institute
W. F. Crowley, Jr., Aerospace & Undersea Support Services, LLC
T. Dingman, Healogics
M. A. Frey, Naval Sea Systems Command
T. R. Galloway, Naval Sea Systems Command
R. M. Webb, *Alternate*, Naval Sea Systems Command
B. Humberstone, Global Diving and Salvage

G. P. Jacob, Navy Experimental Diving Unit
C. B. Kemper III, Kemper Engineering Services, LLC
K. A. Wohlfeil, *Alternate*, Kemper Engineering Services, LLC
J. D. Lawrence, U.S. Coast Guard
T. Gilman, *Alternate*, U.S. Coast Guard
G. Richards, Blanson Ltd.
T. C. Schmidt, Lockheed Martin
J. C. Sheffield, International ATMO, Inc.
D. Talati, Sechrist Industries, Inc.
J. Witney, Atlantis Submarines International

SUBCOMMITTEE ON VIEWPORTS

J. Witney, *Chair*, Atlantis Submarines International
G. Richards, *Vice Chair*, Blanson Ltd.
M. W. Allen, Microbaric Oxygen Systems, LLC
B. Faircloth, FMS Engineering, LLC
C. B. Kemper III, Kemper Engineering Services, LLC
K. A. Wohlfeil, *Alternate*, Kemper Engineering Services, LLC
W. Kohnen, Hydrospace Group, LLC
J. D. Lawrence, U.S. Coast Guard

T. Gilman, *Alternate*, U.S. Coast Guard
D. A. Renear, Aqua-Air Industries, Inc.
T. C. Schmidt, Lockheed Martin
R. C. Smith, Naval Facilities Engineering Service Center
J. Stromer, Triton Submarines
D. Talati, Sechrist Industries, Inc.
R. Thomas, American Bureau of Shipping (ABS)
P. Everley, *Delegate*, Stanley Plastics Ltd.

SUBCOMMITTEE ON SUBMERSIBLES

W. Kohnen, *Chair*, Hydrospace Group, LLC
W. F. Crowley, Jr., Aerospace & Undersea Support Services, LLC
T. R. Galloway, Naval Sea Systems Command
R. M. Webb, *Alternate*, Naval Sea Systems Command
H. Pauli, DNV GL

T. C. Schmidt, Lockheed Martin
K. A. Smith, U.S. Coast Guard
J. D. Lawrence, *Alternate*, U.S. Coast Guard
R. Thomas, American Bureau of Shipping (ABS)
J. Witney, Atlantis Submarines International

HONORARY MEMBERS

R. J. Dzikowski
F. T. Gorman
L. G. Malone
R. P. Swanson

CORRESPONDENCE WITH THE PVHO COMMITTEE

General. ASME Standards are developed and maintained with the intent to represent the consensus of concerned interests. As such, users of this Standard may interact with the Committee by requesting interpretations, proposing revisions or a Case, and attending Committee meetings. Correspondence should be addressed to:

Secretary, PVHO Standards Committee
The American Society of Mechanical Engineers
Two Park Avenue
New York, NY 10016-5990
<http://go.asme.org/Inquiry>

Proposing Revisions. Revisions are made periodically to the Standard to incorporate changes that appear necessary or desirable, as demonstrated by the experience gained from the application of the Standard. Approved revisions will be published periodically.

The Committee welcomes proposals for revisions to this Standard. Such proposals should be as specific as possible, citing the paragraph number(s), the proposed wording, and a detailed description of the reasons for the proposal, including any pertinent documentation.

Interpretations. Upon request, the PVHO Standards Committee will render an interpretation of any requirement of the Standard. Interpretations can only be rendered in response to a written request sent to the Secretary of the PVHO Standards Committee.

The request for an interpretation should be clear and unambiguous. It is further recommended that the inquirer submit his/her request in the following format:

Subject: Cite the applicable paragraph number(s) and the topic of the inquiry.
Edition: Cite the applicable edition of the Standard for which the interpretation is being requested.
Question: Phrase the question as a request for an interpretation of a specific requirement suitable for general understanding and use, not as a request for an approval of a proprietary design or situation. The inquirer may also include any plans or drawings that are necessary to explain the question; however, they should not contain proprietary names or information.

Requests that are not in this format may be rewritten in the appropriate format by the Committee prior to being answered, which may inadvertently change the intent of the original request.

ASME procedures provide for reconsideration of any interpretation when or if additional information that might affect an interpretation is available. Further, persons aggrieved by an interpretation may appeal to the cognizant ASME Committee or Subcommittee. ASME does not "approve," "certify," "rate," or "endorse" any item, construction, proprietary device, or activity.

PVHO Cases. The ASME PVHO Committee issues Cases. Although PVHO Cases do not constitute formal revisions of the Standard, they may be used as representing considered opinions of the PVHO Committee. Once approved, Cases are posted on the ASME PVHO Committee Web page at <http://cstools.asme.org> and are subsequently published with the next edition of this Standard. PVHO Cases are not part of the Standard and are included for information only following the Nonmandatory Appendices.

Requests for PVHO Cases shall provide the following:

- (a) *Statement of Need.* Provide a brief explanation of the need for the revision(s) or addition(s).
- (b) *Background Information.* Provide background information to support the revision(s) or addition(s) including any data or changes in technology that form the basis for the request that will allow the Committee to adequately evaluate the proposed revision(s) or addition(s). Sketches, tables, figures, and graphs should be submitted as appropriate. When applicable, identify any pertinent paragraphs in the Standard that would be affected by the revision(s) or addition(s) and paragraphs in the Standard that reference the paragraphs that are to be revised or added.

Furthermore, the proposed Case should be written as a question and a reply in the same format as existing Cases. Requests for PVHO Cases should also indicate the applicable edition to which the proposed Case applies.

Errata. ASME issues errata (corrections of errors introduced in the Standard during the publishing process; e.g., typographical errors, misspellings, grammatical errors, incorrect publication of revisions). Once approved, errata are posted on the ASME PVHO Committee Web page at <http://cstools.asme.org> and are subsequently published with the next edition of the Standard.

Attending Committee Meetings. The PVHO Standards Committee regularly holds meetings and/or telephone conferences that are open to the public. Persons wishing to attend any meeting and/or telephone conference should contact the Secretary of the PVHO Standards Committee.

ASME PVHO-1–2016 SUMMARY OF CHANGES

Following approval by the PVHO Committee and ASME, and after public review, ASME PVHO-1–2016 was approved by the American National Standards Institute on January 20, 2016.

ASME PVHO-1–2016 contains editorial changes, revisions, and corrections identified by a margin note, **(16)**, placed next to the affected areas.

<i>Page</i>	<i>Location</i>	<i>Change</i>
1	1-1	Revised
2	1-6	Subparagraphs (a), (b), (f) revised
3	1-7.6	Subparagraph (g) revised
4	1-7.13	Title revised
	1-7.13.1	Revised
10–11	1-7.13.4	Subparagraph (d)(3) figure number corrected
13	1-7.13.6	(1) Title revised (2) First paragraph added (3) Subparagraph (a) revised
	1-7.15	Revised
22	2-2.4.1	Subparagraph (a) revised
26	Fig. 2-2.5.1-1	Revised
27	Fig. 2-2.5.1-2	Revised
28	Fig. 2-2.5.1-3	Revised
39	Fig. 2-2.5.1-13	Revised
40	Fig. 2-2.5.1-14	Revised
41	Fig. 2-2.5.1-15	Revised
49	Fig. 2-2.10.1-5	R_m changed to R_n
51	Fig. 2-2.10.1-7	β changed to ϵ
52	Fig. 2-2.10.1-8	β changed to ϵ
53, 56	2-2.11.11	Added, and following paragraphs redesignated
54	Fig. 2-2.11.10-1	Revised
57	Fig. 2-2.11.11-1	Added
62	Table 2-2.14.13-2	Signs added under Specified Values columns
70	2-3.7	Subparagraph (c)(1) revised
83–84	2-10.5	Revised

<i>Page</i>	<i>Location</i>	<i>Change</i>
86–87	4-1.2	(1) Subparagraph 4-1.2.2(c) deleted (2) Subparagraph 4.1.2.2(d) redesignated as 4-1.2.3, and original 4-1.2.3 redesignated as 4-1.2.4
	4-1.2.4	Revised
	4-1.2.5	Added
99	5-1.1	Revised
100	5-5.5.1	Revised
	5-5.6	Added, and following paragraphs redesignated
102	6-1.5	Reference to Nonmandatory Appendix F updated
106–111	6-4.4	Added
	6-4.5	Added
	Fig. 6-4.5.2.2-1	Added
	Fig. 6-4.5.2.2-2	Added
	6-4.6	Added
	6-6	Added
114	7-3.1.3	Reference to para. 4-2.4.5 corrected
121–125	Mandatory Appendix II	(1) Definitions of <i>clump weight</i> , <i>design pressure</i> , <i>drop weight</i> , <i>examination</i> , <i>guide wire(s)</i> , <i>inspection</i> , <i>material specification</i> , <i>operating pressure</i> , <i>testing</i> added (2) Definition of <i>hyperbaric stretcher</i> revised
136	Table D-7.1-1	Redesignated
138, 153	Nonmandatory Appendix E	Added, and former Nonmandatory Appendix E redesignated as Nonmandatory Appendix F

SPECIAL NOTE:

The Cases to PVHO-1 follow the last page of this Edition. However, they are not part of the Standard itself.

INTENTIONALLY LEFT BLANK

SAFETY STANDARD FOR PRESSURE VESSELS FOR HUMAN OCCUPANCY

Section 1 General Requirements

(16) 1-1 INTRODUCTION

This Standard defines the requirements that are applicable to all Pressure Vessels for Human Occupancy (PVHOs) fabricated to this Standard (Sections 1 through 4) and shall be used in conjunction with specific requirements in other Sections (Sections 5 through 7, as applicable) and Mandatory Appendices of this Standard. In the event of conflict between Sections 1 through 4 and other Sections (5 through 7), the application-specific requirements from Sections 5 through 7 shall govern.

PVHOs shall be designed, fabricated, inspected, tested, marked, and stamped in accordance with the requirements of this Standard and of the ASME Boiler and Pressure Vessel Code (the Code), Section VIII, Division 1 or Division 2, unless otherwise permitted within this Standard.

In-service requirements for PVHOs are found in ASME PVHO-2.

1-2 SCOPE

1-2.1 Application

This Standard applies to all pressure vessels that enclose a human within their pressure boundary while under internal or external pressure exceeding a differential pressure of 2 psi (15 kPa). PVHOs include, but are not limited to, submersibles, diving bells, and personnel transfer capsules, as well as decompression, recompression, hypobaric, and hyperbaric PVHOs.

1-2.2 Geometry

The scope of this Standard in relation to the geometry is the pressure boundary as defined in the User's Design Specification and shall include, but not be limited to, the following:

- (a) shells of revolution
- (b) openings and their reinforcements
- (c) nozzles and other connections
- (d) flat heads
- (e) quick-actuating closures

- (f) vessel penetrations
- (g) attachments and supports
- (h) access openings
- (i) viewports
- (j) pressure relief devices
- (k) pressure-retaining covers for vessel openings

1-2.3 Limitations

The pressure boundary of the PVHO shall be as follows:

- (a) welding end connection for the first circumferential joint for welded connections
- (b) the first threaded joint for screwed connections
- (c) the face of the first flange for bolted, flanged connections
- (d) the first sealing surface for proprietary connections or fittings

1-3 EXCLUSIONS

The following types of vessels are excluded from this Standard:

- (a) nuclear reactor containments
- (b) pressurized airplane cabins
- (c) aerospace vehicle cabins
- (d) caissons

1-4 USER REQUIREMENTS

It is the responsibility of the user, or an agent acting for the user who intends that a PVHO be designed, fabricated, inspected, tested, marked, stamped, and certified to be in compliance with this Standard, to provide or cause to be provided for such PVHO, a User's Design Specification. The User's Design Specification shall set forth the intended operating conditions of the PVHO to provide the basis for design. It shall identify the external environment to which the PVHO will be exposed, the intended function of the PVHO, mechanical loads imposed on the PVHO, specific installation requirements, and applicable codes and standards.