

American National Standard

*for Ophthalmics –
Corneal Topography and
Tomography Systems –
Standard Terminology, Requirements*



ANSI
Z80.23-2018
Revision of
ANSI Z80.23-2008
(R2013)

American National Standard
for Ophthalmics –

**Corneal Topography and
Tomography Systems –
Standard Terminology, Requirements**

Sponsor

The Vision Council

Approved November 9, 2018

Published January 31, 2019

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards -

The Vision Council
Z80 Secretariat
225 Reinekers Lane
Suite 700
Alexandria, VA 22314

Published by

The Vision Council
225 Reinekers Lane
Suite 700
Alexandria, VA 22314

Copyright © 2019 by The Vision Council
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	ii
1 Scope and purpose.....	1
1.1 Scope.....	1
1.2 Purpose.....	1
2 Normative references.....	1
3 Terminology	2
4 Requirements.....	8
4.1 Area measured	8
4.2 Measurement sample density	8
4.3 Measurement and report of performance	8
5 Test methods and test devices	8
5.1 Types of test	8
5.1.1 Accuracy	8
5.1.2 Responsibility.....	8
5.2 Test surfaces	9
5.2.1 Reflection-based systems.....	9
5.2.2 Luminous surface systems	9
5.2.3 Optical sectioning systems	9
5.2.4 Interference-based systems.....	9
5.2.5 Specification of test surfaces	9
5.2.6 Verification of test surfaces.....	10
5.2.7 Type testing using test surfaces	10
5.3 Data collection	12
5.3.1 Test surfaces	12
5.3.2 Human corneas.....	12
5.4 Analysis of the data.....	12
5.4.1 Structure of the accuracy data set	12
5.4.2 Structure of the repeatability data set	12
5.4.3 Analysis of the paired data sets	13
5.4.3.1 Special considerations for the comparison of corneal elevation values	13
5.4.4 Report of performance	14
5.4.4.1 Accuracy	14
5.4.4.2 Repeatability	14
6 Accompanying documents.....	14
7 Marking	15

	Page
Tables	
1	Conic section descriptors..... 7
2	Test surfaces for type testing..... 11
3	Tolerance levels for test surface measurements 11
4	Analysis zones for accuracy and repeatability testing 13
Figures	
1	Illustration of axial curvature, K_a , axial radius of curvature, r_a , meridional curvature, K_m , and meridional radius of curvature, r_m 7
2	Illustration of the corneal vertex and the apex 7
Annexes	
A	Test surfaces for corneal topographers and corneal tomographers .. 16
B	Standardized displays for corneal topographers and corneal tomographers 18
C	Calculation of area weighting values 21
D	Pre-processing of axial elevation data prior to analysis of paired data sets 23

Foreword (This foreword is not part of American National Standard ANSI Z80.23-2018.)

This American National Standard continues to address the expressed needs of those members of the ophthalmic community who use corneal topography in clinical settings, those who manufacture corneal topographers and those who teach others in the use of the information collected by corneal topographers. In particular, there continues to be a need for standardization of the terms and definitions used in the field, for standardization of the methods used for characterizing the performance of these instruments and for standardization of displays of corneal topographical information. The experts who worked together to initially create this standard felt that, at the time, there was not sufficient consensus within the ophthalmic community to set performance requirements for these instruments beyond those for minimum area measured and measurement sample density. Later, in working with international experts in the field, it was decided that instruments should be tested on a selected set of test surfaces to ensure that the calibration to set their scale factors had been done correctly. The standard continues to address standardization of the methods for testing these instruments, for assessing their performance, and for reporting the results thus obtained.

The number and type of test surfaces to be used has been changed to include only test surfaces for which the results can be verified. When these surfaces are tilted or rotated the expected surface measurements are easy to predict. These surfaces were considered to be adequate as minimum verification surfaces for corneal topographers; if a corneal topography system can measure these surfaces well, it will be a clinically useful instrument. The method for standardization of color maps has been changed in an effort to improve the user's ability to discern just-noticeable differences in corneal topography. The user always has the option of using a scale with less resolution but with greater range, as long as the scale recommended in this document is available to be used.

This standard contains four annexes. Annex A is informative and is not considered part of this standard. Annexes B through D are normative and are considered part of this standard.

This standard was created by a special working group created by the Z80 Subcommittee on Ophthalmic Instruments and included experts in the field of corneal topography from the clinical, manufacturing and academic areas of the ophthalmic community. Suggestions for improvement of this standard will be welcome. They should be sent to the Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Carl Tubbs, M.D., Chairman
Neal Roche, Vice-Chairman
William Benjamin, O.D., Secretary
Michael Vitale, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Advanced Medical Technology Association	Michael Pflieger
American Academy of Ophthalmology	Carl Tubbs
American Academy of Optometry	David Loshin
American Ceramic Society	Lyle Rubin
American Glaucoma Society	Steven Gedde
American Optometric Association.....	William Benjamin
American Society of Cataract and Refractive Surgery	Stephen Klyce
Contact Lens Institute	Stan Rogaski
Contact Lens Manufacturers Association	Martin Dalsing
Cornea Society	Michael Belin
Department of Veterans Affairs	John Townsend
Food & Drug Administration CDRH/Division.....	Don Calogero
Johnson & Johnson Vision	Priya Janakiraman
National Association of Optometrists & Opticians	Nick Mileti
Optical Laboratory Association	Steve Sutherland
Opticians Association of America	Tom Hicks
Sunglass Association of America	Tibor Gross
The Vision Council.....	Michael Vitale
ISO TC 172/SC7	Michael Vitale

Individual Member

Ralph Stone

The members of the subcommittee on Ophthalmic Instruments who contributed to this standard are:

Charles E. Campbell, Chair
William L. Brown, O.D., Ph.D,
Karl Citek, O.D. Ph.D.
Bruce Drum, Ph.D.
Stephen Farrer
Stephen Klyce, Ph.D.
Dexiu Shi, Ph.D.
David Sliney, Ph.D.
Robert Rosenberg, O.D.

American National Standard
for Ophthalmics –

Corneal Topography and Tomography Systems – Standard Terminology, Requirements

1 Scope and purpose

1.1 Scope

This American National Standard applies to instruments, systems and methods that are intended to measure the shape of the cornea of the human eye over a majority of its area. The measurements of the anterior and/or posterior surface in local areas may be of curvature and/or three dimensional topographical measurements of the surface. The measurements may be used to derive more global parameters used to characterize the surfaces. Instruments classified as ophthalmometers or keratometers are not covered by this standard.

1.2 Purpose

This standard defines certain terms that are peculiar to the characterization of the corneal shape so that they may be standardized throughout the field of vision care and have common meaning for all those who have occasion to participate in this area.

This standard sets forth minimum requirements for instruments and systems that fall into the class of corneal topographers.

This standard sets forth tests and verification procedures that will verify that a system or instrument complies with the standard and so qualifies as a corneal topographer or corneal tomographer in the meaning of this standard.

This standard sets forth certain tests and verification procedures that will allow the verification of capabilities of systems that are beyond the minimum required for corneal topographers or corneal tomographers.

2 Normative references

The following standard contains provisions that, through reference in this text, constitute provisions of this American National Standard. At the time of publication, the edition indicated was valid. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below.

ANSI Z80.20-2004, *Ophthalmics – Contact Lenses – Standard terminology, tolerances, measurements and physicochemical properties*¹⁾

ANSI Z80.28-2017, *Ophthalmic Instruments – Methods for reporting optical aberrations of eyes*¹⁾

ISO 8429:1986, *Optics and optical instruments – Ophthalmology – Gradual dial scale*¹⁾

¹⁾ For electronic copies of some standards, visit ANSI's Electronic Standards Store (ESS) at www.ansi.org. For printed versions of all these standards, contact Global Engineering Documents, 15 Inverness Way East, Englewood, CO 80112-5704, (800) 854-7179.