



Medical electrical equipment

Part 2.23: Particular requirements for the basic safety and essential performance of transcutaneous partial pressure monitoring equipment (IEC 60601-2-23:2011 (ED. 3.0), MOD)



AS 60601.2.23:2018

This Australian Standard® was prepared by HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 5 June 2018.

This Standard was published on 29 June 2018.

The following are represented on Committee HE-003:

- Australasian College of Physical Scientists and Engineers in Medicine
- Australian and New Zealand College of Anaesthetists
- Australian Chamber of Commerce and Industry
- Australian Radiation Protection and Nuclear Safety Agency
- Australian Society of Anaesthetists
- Certification Body Australia (Certification Interests Australia)
- College of Biomedical Engineering Engineers Australia
- Engineers Australia
- Medical Technology Association of Australia
- Therapeutic Goods Administration

This Standard was issued in draft form for comment as DR AS 60601.2.23:2018.

Keeping Standards up-to-date

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

www.standards.org.au

www.saiglobal.com (sales and distribution)

ISBN 978 1 76072 107 7



Medical electrical equipment

Part 2.23: Particular requirements for the basic safety and essential performance of transcutaneous partial pressure monitoring equipment (IEC 60601-2-23:2011 (ED. 3.0), MOD)

Originated as AS/NZS 3200.2.23:1994.
Second edition 2001.
Revised and redesignated as AS 60601.2.23:2018.

COPYRIGHT

© IEC 2018 — All rights reserved
© Standards Australia Limited 2018

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth).

Published by SAI Global Pty Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia.

Preface

This Standard was prepared by the Australian members of Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment, to supersede AS/NZS 3200.2.23—2001, *Medical electrical equipment, Part 2.23: Particular requirements for safety—Transcutaneous partial pressure monitoring equipment*.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

The objective of this Standard is to specify the requirements for basic safety and essential performance of transcutaneous partial pressure monitoring equipment. This Standard applies to transcutaneous monitors used with adults, children and neonates, and it includes the use of these devices in foetal monitoring during birth.

This Standard is an adoption with national modifications, and has been reproduced from, IEC 60601-2-23:2011 (ED. 3.0), *Medical electrical equipment — Part 2-23: Particular requirements for the basic safety and essential performance of transcutaneous partial pressure monitoring equipment*. Appendix ZZ lists the variations to IEC 60601-2-23:2011 for the application of this Standard in Australia.

As this document has been reproduced from an International Standard, the following applies:

- (a) In the source text 'this International Standard' should read 'Australian Standard'.
- (b) A full point substitutes for a comma when referring to a decimal marker.

Australian or Australian/New Zealand Standards that are identical adoptions of international normative references may be used interchangeably. Refer to the online catalogue for information on specific Standards.

The terms 'normative' and 'informative' are used in Standards to define the application of the appendices or annexes to which they apply. A 'normative' appendix or annex is an integral part of a Standard, whereas an 'informative' appendix or annex is only for information and guidance.

NOTES

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	10
201.4 General requirements	11
201.5 General requirements for testing of ME EQUIPMENT	11
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	12
201.7 ME EQUIPMENT identification, marking and documents	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	13
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	14
201.10 Protection against unwanted and excessive radiation HAZARDS	14
201.11 Protection against excessive temperatures and other HAZARDS	14
201.12 Accuracy of controls and instruments and protection against hazardous outputs	18
201.13 HAZARDOUS SITUATIONS and fault conditions	23
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	23
201.15 Construction of ME EQUIPMENT.....	23
201.16 ME SYSTEMS	23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	24
202 Electromagnetic compatibility – Requirements and tests	24
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	28
Annexes	34
Annex AA (informative) Particular guidance and rationale.....	35
Annex BB (informative) Alarm diagrams of Clause 208/IEC 60601-1-8:2006	43
Index of defined terms used in this particular standard.....	46
Figure 201.101 – TRANSDUCER cable strain relief test.....	14
Figure 201.102 – Foam block test (see 201.11.1.2.2.104 and 201.11.1.2.2.105).....	15
Figure 201.103 – Linearity and hysteresis test set-up – Gas mix chamber, assembled.....	20
Figure 201.104 – Linearity and hysteresis test set-up – Gas mix chamber, manufacturing dimensions	21
Figure 201.105 – Linearity and hysteresis test set-up – Gas mix chamber, dimensions of hose connector	22
Figure 202.101 – Set-up for radiated and conducted EMISSIONS testing according to 202.6.1.1.2 a)	25
Figure 202.102 – Set-up for radiated immunity test according to 202.6.2.3.2	27
Figure BB.1 – NON-LATCHING ALARM SIGNALS without ALARM RESET.....	43
Figure BB.2 – NON-LATCHING ALARM SIGNALS with ALARM RESET	43
Figure BB.3 – LATCHING ALARM SIGNALS with ALARM RESET	44
Figure BB.4 – Two ALARM CONDITIONS with ALARM RESET.....	44

Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements	11
Table 201.102 – Required readings and tolerances	19
Table 201.103 – Calibration test gases	19
Table 208.101 – ALARM CONDITION priorities	29
Table 208.102 – Characteristics of the burst of auditory ALARM SIGNALS	30

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-23: Particular requirements for the basic safety and essential performance of transcutaneous partial pressure monitoring equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-23 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 1999 and constitutes a technical revision. This edition of IEC 60601-2-23 was revised to align structurally with the 2005 edition of IEC 60601-1.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/885/FDIS	62D/907/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of TRANSCUTANEOUS PARTIAL PRESSURE MONITORING EQUIPMENT. It amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard.

The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the more important requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this Standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-23: Particular requirements for the basic safety and essential performance of transcutaneous partial pressure monitoring equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of TRANSCUTANEOUS PARTIAL PRESSURE MONITORING EQUIPMENT as defined in 201.3.63 and hereinafter referred to as ME EQUIPMENT, whether this ME EQUIPMENT is stand alone or part of a system.

This standard applies to transcutaneous monitors used with adults, children and neonates, and it includes the use of these devices in foetal monitoring during birth.

This standard does not apply to haemoglobin saturation oximeters or to devices applied to surfaces of the body other than the skin (for example conjunctiva, mucosa).

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the General Standard.

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for TRANSCUTANEOUS PARTIAL PRESSURE MONITORING EQUIPMENT as defined in 201.3.63.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

¹ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*