

Australian Standard[®]

Medical suction equipment

**Part 1: Electrically-powered
suction equipment—
Safety requirements**

This Australian Standard was prepared by Committee HT/4, Medical Gases and Pipeline Services. It was approved on behalf of the Council of Standards Australia on 22 July 1992 and published on 16 November 1992.

The following interests are represented on Committee HT/4:

Association of Consulting Engineers, Australia
Australian Society of Anaesthetists
Confederation of Australian Industry
Department of Health, Housing and Community Services (Commonwealth)
Department of Health, N.S.W.
Health Department, W.A.
Institute of Hospital Engineering, Australia
Medical Industry Association of Australia
Metal Trades Industry Association of Australia
Public Works Department, N.S.W.
Royal Australasian College of Surgeons
South Australian Health Commission

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS 2120.1—1992

Australian Standard[®]

Medical suction equipment

**Part 1: Electrically-powered
suction equipment—
Safety requirements**

First published as part of AS 2120—1977.
Revised and redesignated in part as AS 2120.1—1992.

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 7738 X

PREFACE

This Standard was prepared by the Standards Australia Committee on Medical Gases and Pipeline Services, under the direction of the Multitechnics Standards Policy Board, to supersede, in part, AS 2120—1977, *Suction systems for medical use in hospitals*.

This Standard is the first in a series of Standards for medical suction equipment, and deals only with safety requirements for electrically-powered suction equipment. Part 2 deals with manually-powered suction equipment. Part 3 deals with suction equipment powered from a vacuum or pressure source.

The clauses of this Standard supplement or modify the corresponding clauses in AS 3200.1/NZS 6150, *Approval and test specifications—Electromedical equipment—General requirements*, Part 1: *General requirements for safety*. Where the reference in the text of this Standard includes an 'addition', 'replacement', 'substitution' or 'modification' of the relevant requirements, tests or explanatory notes of AS 3200.1/NZS 6150, these changes are made to the relevant text which then become part of that Standard. It will be observed that when no such change is necessary, the words 'the requirements given in (the relevant) clause of AS 3200.1/NZS 6150 apply' are used.

Where the requirements of AS 3200.1/NZS 6150 do not apply, the words 'not used' are used. Although this Standard closely follows AS 3200.1/NZS 6150 in format and technical content, some of the requirements of that publication have been modified to take account of local conditions.

Where this Standard deviates technically from ISO 10079-1:1991, *Medical suction equipment*, Part 1: *Electrically powered suction equipment—Safety requirements*, by way of additional or different requirements, such deviations are indicated by a rule in the margin against the Clause, or part thereof, affected. Minor changes are not indicated. An Annex to the Standard lists the variations from ISO 10079-1.

This Standard has been prepared in response to a need for a safety and performance Standard for suction systems. Suction is used to clear the airway and remove unwanted material from body cavities. Suction is also used to assist drainage and decompress body cavities. Suction and vacuum systems are used widely both in health care facilities such as hospitals, for domiciliary care of patients who are nursed at home, and in emergency situations both outside hospitals in field conditions, and during transport in ambulances.

In this Standard, vacuum readings are specified as gauge (relative) pressures to assist clinical personnel. However, this is not intended to prevent engineering groups from using absolute vacuum in their design process. The specifications in this Standard assume an ambient pressure of 1 atmosphere at sea level.

Test methods other than those specified in this Standard, but of equal or greater accuracy, may be used to verify compliance with the given requirements. However, in the event of a dispute, the methods specified in this Standard are to be used as the reference methods.

This Standard requires reference to Australian and ISO Standards. These are listed in Appendix L.

A rationale for the most important requirements is given in Appendix O. It is considered that a knowledge of the reasons for the requirements will not only facilitate the proper application of the Standard, but will expedite any subsequent revision.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
SECTION 1 GENERAL	
1 SCOPE	5
2 TERMINOLOGY AND DEFINITIONS	5
3 GENERAL REQUIREMENTS	7
4 GENERAL REQUIREMENTS FOR TESTS	7
5 CLASSIFICATION	7
6 IDENTIFICATION, MARKING AND DOCUMENTS	7
7 POWER INPUT	8
SECTION 2 ENVIRONMENTAL CONDITIONS	
8 BASIC SAFETY CATEGORIES	9
9 REMOVABLE PROTECTIVE MEANS	9
10 ENVIRONMENTAL CONDITIONS	9
11 (NOT USED)	9
12 (NOT USED)	9
SECTION 3 PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
13 GENERAL	9
14 REQUIREMENTS RELATED TO CLASSIFICATION	9
15 LIMITATION OF VOLTAGE AND ENERGY	9
16 ENCLOSURES AND PROTECTIVE COVERS	9
17 SEPARATION	9
18 PROTECTIVE EARTHING, FUNCTIONAL EARTHING AND POTENTIAL EQUALIZATION	9
19 CONTINUOUS LEAKAGE CURRENTS AND PATIENT AUXILIARY CURRENTS	9
20 DIELECTRIC STRENGTH	9
SECTION 4 PROTECTION AGAINST MECHANICAL HAZARDS	
21 MECHANICAL STRENGTH	10
22 MOVING PARTS	10
23 SURFACES, CORNERS AND EDGES	10
24 STABILITY IN NORMAL USE	10
25 EXPELLED PARTS	10
26 VIBRATION AND NOISE	10
27 PNEUMATIC AND HYDRAULIC POWER	10
28 SUSPENDED MASSES	10
SECTION 5 PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
29 X-RADIATION	11
30 ALPHA, BETA, GAMMA AND NEUTRON RADIATION AND OTHER PARTICLE RADIATION	11
31 MICROWAVE RADIATION	11
32 LIGHT RADIATION (INCLUDING LASERS)	11
33 INFRA-RED RADIATION	11
34 ULTRAVIOLET RADIATION	11
35 ACOUSTICAL ENERGY (INCLUDING ULTRASONIC)	11
36 ELECTROMAGNETIC COMPATIBILITY	11
SECTION 6 PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
37 LOCATIONS AND BASIC REQUIREMENTS	11
38 MARKING AND ACCOMPANYING DOCUMENTS	11

	<i>Page</i>
39 COMMON REQUIREMENTS FOR CATEGORY AP AND CATEGORY APG EQUIPMENT	11
40 REQUIREMENTS AND TESTS FOR CATEGORY AP EQUIPMENT—PARTS AND COMPONENTS THEREOF	11
41 REQUIREMENTS AND TESTS FOR CATEGORY APG EQUIPMENT—PARTS AND COMPONENTS THEREOF	11
 SECTION 7 PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
42 EXCESSIVE TEMPERATURES	12
43 FIRE PREVENTION	12
44 OVERFLOW, SPILLAGE, LEAKAGE, HUMIDITY, INGRESS OF LIQUIDS, CLEANING, STERILIZATION AND DISINFECTION	12
45 PRESSURE VESSELS AND PARTS SUBJECT TO PRESSURE	14
46 HUMAN ERRORS	14
47 ELECTROSTATIC CHARGES	14
48 MATERIALS IN APPLIED PARTS IN CONTACT WITH THE BODY OF THE PATIENT	14
49 INTERRUPTION OF THE POWER SUPPLY	14
 SECTION 8 ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
50 ACCURACY OF OPERATING DATA	15
51 PROTECTION AGAINST HAZARDOUS OUTPUT	15
 SECTION 9 ABNORMAL OPERATION AND FAULT CONDITIONS— ENVIRONMENTAL TESTS	
52 ABNORMAL OPERATION AND FAULT CONDITIONS	16
53 ENVIRONMENTAL TESTS	16
 SECTION 10 CONSTRUCTIONAL REQUIREMENTS	
54 GENERAL	17
55 ENCLOSURES AND COVERS	17
56 COMPONENTS AND GENERAL ASSEMBLY	17
57 MAINS PARTS, COMPONENTS AND LAYOUT	19
58 PROTECTIVE EARTHING—TERMINALS AND CONNECTIONS	19
59 CONSTRUCTION AND LAYOUT	19
 APPENDICES	
L REFERENCED DOCUMENTS	23
M LUMEN (PASSAGEWAY) SIZE AND ITS EFFECTS ON FLOW	24
N TABLE OF TYPICAL RANGE OF VOLUMES FOR COLLECTION CONTAINERS FOR SPECIFIC USES	24
O RATIONALE STATEMENT	25
ANNEX	26

STANDARDS AUSTRALIA

Australian Standard
Medical suction equipment

Part 1: Electrically-powered suction equipment—Safety requirements

SECTION 1 GENERAL

1 SCOPE* This Standard is based on AS 3200.1/NZS 6150. As stated in Clause 1.3 of AS 3200.1/NZS 6150, the requirements of this Standard take precedence over those of AS 3200.1/NZS 6150.

The Scope and Object given in Clause 1 of AS 3200.1/NZS 6150 applies except that Clause 1.1 shall be replaced by the following:

This Standard specifies minimum safety and performance requirements for medical and surgical suction equipment (see Figure 1) in health care facilities such as hospitals, for domiciliary care of patients and for field and transport use. Although equipment may be driven by centrally-powered, piped vacuum systems, compressed gases, electricity or be manually powered for a variety of applications, this Standard addresses only transportable equipment powered electrically.

Excluded from this Standard are the following:

- (a) Central power source (vacuum/compressed air generation), piping systems of vehicles and buildings, and wall connectors.
- (b) Catheter tubes, drains, curettes and suction tips.
- (c) Syringes.
- (d) Dental suction equipment, complying with AS 2686.2.
- (e) Waste gas scavenging systems.
- (f) Laboratory suction.
- (g) Autotransfusion systems.
- (h) Passive urinary drainage.
- (i) Closed systems for wound drainage.
- (j) Gravity gastric drainage.
- (k) Orally-operated mucous extractors.
- (l) Suction equipment where the collection container is downstream of the vacuum pump.
- (m) Equipment marked as suction unit for permanent tracheostomy.
- (n) Ventouse (obstetric) equipment.
- (o) Neonatal mucous extractors.
- (p) Suction equipment marked for endoscopic use only.

1.4 Environmental conditions The requirements given in Clause 1.4 of AS 3200.1/NZS 6150 apply except that the following modification shall be made to Clause 1.4(b)(i):

Substitute '+5°C' for '+10°C' and '+35°C' for '40°C'.

For field and transport use, environmental conditions shall be as specified in Clause 53.

2 TERMINOLOGY AND DEFINITIONS For the purposes of this Standard, the definitions given in Clause 2 of AS 3200.1/NZS 6150 apply except for the following variations:

- (a) The definition given in Clause 2.1.5 shall be replaced by the following:
Applied part—all parts in the liquid pathway.
- (b) The following shall be added to the definition given in Clause 2.4.3:
Safety extra-low voltage (SELV)—includes the electrical sources which are isolated (e.g. car battery) and do not require a separate transformer or converter with separate windings.

For the purpose of this Standard, the definitions below also apply.

2.1 Breast pump—vacuum pump for the collection of breast milk.

2.2 Collection container—container in which fluids or solid particles are collected.

2.3 Collection container assembly—collection container and its closure.

* See also Appendix O (of this Standard).