



BSI Standards Publication

# Industrial process control devices — Thermographic cameras

---

Part 1: Metrological characterization

## National foreword

This Published Document is the UK implementation of IEC TS 63144-1:2020.

The UK participation in its preparation was entrusted to Technical Committee GEL/65/2, Elements of systems.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2020  
Published by BSI Standards Limited 2020

ISBN 978 0 539 00010 8

ICS 25.040.40; 17.200

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 May 2020.

### Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

---



# IEC TS 63144-1

Edition 1.0 2020-04

# TECHNICAL SPECIFICATION

---

**Industrial process control devices – Thermographic cameras –  
Part 1: Metrological characterization**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

---

ICS 17.200, 25.040.40

ISBN 978-2-8322-7969-4

**Warning! Make sure that you obtained this publication from an authorized distributor.**

## CONTENTS

FOREWORD.....	6
INTRODUCTION.....	8
1 Scope.....	9
2 Normative references .....	9
3 Terms and definitions .....	9
4 Symbols .....	14
5 Abbreviated terms .....	14
6 Determination of technical data .....	15
6.1 General.....	15
6.2 Measuring temperature range .....	15
6.2.1 General .....	15
6.2.2 Required parameters .....	16
6.2.3 Examples of indications .....	16
6.2.4 Test condition, method and procedure for measuring temperature range .....	16
6.3 Noise equivalent temperature difference ( <i>NETD</i> ).....	16
6.3.1 General .....	16
6.3.2 Required parameters .....	16
6.3.3 Examples of indications .....	16
6.3.4 Test condition, method and procedure for noise equivalent temperature difference .....	17
6.4 Measuring distance ( <i>d</i> ).....	18
6.4.1 General .....	18
6.4.2 Required parameters .....	18
6.4.3 Examples of indications .....	18
6.4.4 Test condition, method and procedure for measuring distance.....	18
6.5 Field of view ( <i>FOV</i> ) .....	18
6.5.1 General .....	18
6.5.2 Required parameters .....	19
6.5.3 Examples of indications .....	19
6.5.4 Test condition, method and procedure for field of view .....	19
6.6 Number of image elements .....	19
6.7 Detector format used (number of detector elements used) .....	19
6.8 Instantaneous field of view ( <i>IFOV</i> ) .....	20
6.8.1 General .....	20
6.8.2 Required parameters .....	20
6.8.3 Example of indications.....	20
6.8.4 Test condition, method and procedure for instantaneous field of view.....	20
6.9 Slit response function (SRF) .....	20
6.9.1 General .....	20
6.9.2 Required parameters .....	21
6.9.3 Examples of indications .....	21
6.9.4 Test condition, method and procedure for slit response function .....	21
6.10 Minimum field of view for temperature measurement ( <i>MFOV<sub>T</sub></i> ).....	22
6.10.1 General .....	22
6.10.2 Required parameters .....	23
6.10.3 Example of indications.....	23

6.10.4	Test condition, method and procedure for minimum field of view for temperature measurement .....	23
6.11	Spectral range .....	24
6.11.1	General .....	24
6.11.2	Examples of indications .....	24
6.11.3	Test condition, method and procedure for spectral range .....	24
6.12	Emissivity setting .....	24
6.12.1	General .....	24
6.12.2	Examples of indications .....	24
6.12.3	Test condition, method and procedure for emissivity setting .....	24
6.13	Influence of the internal instrument temperature .....	24
6.13.1	General .....	24
6.13.2	Required parameters .....	25
6.13.3	Examples of indications .....	25
6.13.4	Test condition, method and procedure for influence of the internal instrument temperature .....	25
6.14	Influence of the humidity .....	26
6.14.1	General .....	26
6.14.2	Required parameters .....	26
6.14.3	Example of indications .....	26
6.14.4	Test condition, method and procedure for influence of the humidity .....	26
6.15	Long-term stability .....	26
6.15.1	General .....	26
6.15.2	Required parameters .....	26
6.15.3	Example of indication .....	26
6.15.4	Test condition, method and procedure for long-term stability .....	27
6.16	Short-term stability .....	27
6.16.1	General .....	27
6.16.2	Required parameters .....	27
6.16.3	Example of indication .....	28
6.16.4	Test condition, method and procedure for short-term stability .....	28
6.17	Repeatability .....	28
6.17.1	General .....	28
6.17.2	Required parameters .....	28
6.17.3	Example of indication .....	29
6.17.4	Test condition, method and procedure for repeatability .....	29
6.18	Interchangeability (spread of production) .....	29
6.18.1	General .....	29
6.18.2	Required parameters .....	29
6.18.3	Example of indication .....	30
6.18.4	Test condition, method and procedure for interchangeability (spread of production) .....	30
6.19	Response time .....	30
6.19.1	General .....	30
6.19.2	Required parameters .....	34
6.19.3	Example of indication .....	34
6.19.4	Test condition, method and procedure for response time .....	34
6.20	Exposure time .....	35
6.20.1	General .....	35
6.20.2	Required parameters .....	36

6.20.3	Example of indication .....	36
6.20.4	Test condition, method and procedure for exposure time .....	36
6.21	Warm-up time .....	37
6.21.1	General .....	37
6.21.2	Required parameters .....	37
6.21.3	Examples of indication .....	37
6.21.4	Test condition, method and procedure for warm-up time .....	37
6.22	Integration time setting range .....	38
6.22.1	General .....	38
6.22.2	Required parameters .....	38
6.22.3	Example of indication .....	38
6.23	Refresh rate .....	38
6.23.1	General .....	38
6.23.2	Example of indication .....	39
6.23.3	Test condition, method and procedure for refresh rate .....	39
6.24	Non-uniformity (inhomogeneity of detector responsivity) .....	39
6.24.1	General .....	39
6.24.2	Required parameters .....	39
6.24.3	Example of indication .....	39
6.24.4	Test condition, method and procedure for non-uniformity .....	39
6.25	Inhomogeneity equivalent temperature difference (IETD) .....	40
6.25.1	General .....	40
6.25.2	Required parameters .....	40
6.25.3	Examples of indications .....	40
6.25.4	Test condition, method and procedure for inhomogeneity equivalent temperature difference .....	40
6.26	Operating temperature range and air humidity range .....	41
6.26.1	General .....	41
6.26.2	Example of indication .....	41
6.26.3	Test condition, method and procedure for operating temperature range and air humidity range .....	41
6.27	Size-of-source effect (SSE) .....	41
6.27.1	General .....	41
6.27.2	Required parameters .....	41
6.27.3	Examples of indications .....	42
6.27.4	Test condition, method and procedure for size-of-source effect .....	42
Annex A (informative) Change in the indicated temperature caused by a 1 % change in the radiative interchange .....		43
Bibliography .....		44
Figure 1 – Schematic measuring setup .....		15
Figure 2 – Slit response function .....		21
Figure 3 – Minimum size of a measuring spot for temperature measurement .....		22
Figure 4 – Synchronous signal acquisition for a quantum detector .....		31
Figure 5 – Asynchronous signal acquisition for a quantum detector .....		32
Figure 6 – Asynchronous signal acquisition for a thermal detector (best case) .....		33
Figure 7 – Asynchronous signal acquisition for a thermal detector (worst case) .....		34
Figure 8 – Example of the measurement of the warm-up time .....		38

Table 1 – Symbols .....	14
Table A.1 – Change in the indicated temperature.....	43

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**INDUSTRIAL PROCESS CONTROL DEVICES –  
THERMOGRAPHIC CAMERAS –****Part 1: Metrological characterization**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. In exceptional circumstances, a technical committee may propose the publication of a Technical Specification when

- the required support cannot be obtained for the publication of an International Standard, despite repeated efforts, or
- the subject is still under technical development or where, for any other reason, there is the future but no immediate possibility of an agreement on an International Standard.

Technical Specifications are subject to review within three years of publication to decide whether they can be transformed into International Standards.

IEC TS 63144-1, which is a Technical Specification, has been prepared by subcommittee 65B: Measurement and control devices, of IEC technical committee 65: Industrial-process measurement, control and automation.

The text of this Technical Specification is based on the following documents:

Draft TS	Report on voting
65B/1129/DTS	65B/1159/RVDTS

Full information on the voting for the approval of this Technical Specification can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 63144 series, published under the general title *Industrial process control devices – Thermographic cameras*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

## INTRODUCTION

Thermographic cameras (also called "thermographic imagers" or "infrared cameras") are being increasingly used for spatially and temporally resolved, non-contact radiation temperature measurement. Tracing the temperature values indicated by these instruments to the International Temperature Scale (ITS-90) is gaining in importance for the comparability of measurements. The precondition for their calibration and metrological application with low uncertainties is to accurately describe and determine the essential metrological data of thermographic cameras. Whereas there are international regulations to determine the technical specifications for radiation thermometers – namely IEC TS 62492-1 and IEC TS 62492-2 – there is a lack of such regulations for thermographic cameras in such a detailed form.

This document is Part 1 of a series of technical specifications for thermographic cameras. It is intended to improve comparability and testability of the essential metrological technical data of thermographic cameras. To this end, unambiguous procedures are laid down for the indication and the determination of this technical data. Future IEC TS 63144-2 is intended to specifically address the absolute calibration procedures and the corresponding uncertainties for thermographic cameras in more depth and detail.

# INDUSTRIAL PROCESS CONTROL DEVICES – THERMOGRAPHIC CAMERAS –

## Part 1: Metrological characterization

### 1 Scope

This part of IEC 63144 applies, in the field of metrology, to the statement and testing of technical data in datasheets and instruction manuals for thermographic cameras that are used to measure the temperature of surfaces. This includes, unless otherwise stated, both two-dimensional and one-dimensional (line cameras or line scanners) temperature measuring instruments, independently of the scanning principle (fixed multi-element detector or scanning camera system).

This document describes standard test methods to determine relevant metrological data of thermographic cameras. Manufacturers and sellers can choose relevant data and can state that the data shall be compliant with this Technical Specification.

### 2 Normative references

There are no normative references in this document.

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE The term “uncertainty” used in this document is precisely derived from the above databases and more specifically from ISO/IEC Guide 98-3:2008 [3].

#### 3.1

##### **blackbody radiator**

radiator that emits radiation in a very good approximation of Planck's radiation law

Note 1 to entry: A blackbody radiator is a source of thermal radiation with an effective emissivity close to 1.

#### 3.2

##### **coverage factor**

numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

Note 1 to entry: Coverage factor,  $k$ , is a number larger than one and, typically, in the range from 2 to 3.

#### 3.3

##### **detector format used**

number of detector elements (detector pixels) that have actually been used to record the image

Note 1 to entry: The term "pixel", which is frequently used, will not be used in this document since it can refer both to the detector (detector element) and to the image (image element).