



TECHNICAL MEMORANDUM:
MEASUREMENT UNCERTAINTY FOR
LIGHTING EQUIPMENT CALIBRATION
USING INTEGRATING SPHERES
AN AMERICAN NATIONAL STANDARD



ANSI/IES TM-31-20

**TECHNICAL MEMORANDUM:
MEASUREMENT UNCERTAINTY FOR
LIGHTING EQUIPMENT CALIBRATION
USING INTEGRATING SPHERES
AN AMERICAN NATIONAL STANDARD**

Publication of this Technical Memorandum
has been approved by IES.
Suggestions for revisions
should be directed to IES.

**Prepared by
The IES Testing Procedures Committee**



Copyright 2020 by the Illuminating Engineering Society.

Approved by the IES Standards Committee, February 23, 2018, as a Transaction of the Illuminating Engineering Society.

Approved February 25, 2020 as an American National Standard.

All rights reserved. No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IES.

Published by the Illuminating Engineering Society, 120 Wall Street, New York, New York 10005.

IES Standards are developed through committee consensus and produced by the IES Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, please forward them to Brian Liebel, IES Director of Standards, at standards@ies.org or the above address for verification and correction. The IES welcomes and urges feedback and comments.

Printed in the United States of America.

ISBN# 978-0-87995-397-3

DISCLAIMER

IES publications are developed through the consensus standards development process approved by the American National Standards Institute. This process brings together volunteers representing varied viewpoints and interests to achieve consensus on lighting recommendations. While the IES administers the process and establishes policies and procedures to promote fairness in the development of consensus, it makes no guaranty or warranty as to the accuracy or completeness of any information published herein.

The IES disclaims liability for any injury to persons or property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of, or reliance on this document.

In issuing and making this document available, the IES is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is the IES undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

The IES has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does the IES list, certify, test or inspect products, designs, or installations for compliance with this document. Any certification or statement of compliance with the requirements of this document shall not be attributable to the IES and is solely the responsibility of the certifier or maker of the statement.

AMERICAN NATIONAL STANDARD

Approval of an American National Standard requires verification by ANSI that the requirements for due process, consensus, and other criteria have been met by the standards developer.

Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made toward their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether that person has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation to any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised at any time. The procedures of the American National Standards Institute require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of approval. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Prepared by the IES Testing Procedures Committee.

C. Cameron Miller, *Chair*
Becky Kuebler, *Vice-Chair*
David Ellis, *Secretary*
Jianzhong Jiao, *Treasurer*

Members

C. Andersen	M. Grather	S. Longo	M. Sapcoe
A. Baker	J. Hospodardky	J. Marella	G. Steinberg
R. Berger	J. Hulett	P. McCarthy	J. Yon
R. Bergman	P.C. Hung	G. McKee	
E. Bretschneider	A. Jackson	E. Radkov	
P. Elizondo	M. Kotrebai	D. Randolph	

Advisory Members

L. Ayers	R. Heinisch	A. Mendoza	L. Stafford
R. Bergin	K. Hemmi	D. Miletich	W. Sun
C. Bloomfield	Y. Hiebert	M. Minarczyk	L. Swainston
K. Broughton	R. Higley	Z. Mooney	J. Swiernik
E. Carter	J. Hospodarsky	M. Nadal	R. Taylor
D. Chan	S. Hua	D. Nava	K. Tracy
P.-T. Chou	D. Jenkins	D. O'Hare	R. Tuttle
G. Connelly	T. Kawabata	Y. Ohno	T. Uchida
J. Creveling	R. Kelley	J. Pan	J. Vollers
J. Dakin	J. Lawton	D. Park	K. Wagner
L. Davis	L. Leetzoe	N. Peimanovic	J. Walker
J. Demirjian	J. Leland	E. Perkins	Y. Wang
M. Dejardins	K. Lerbs	M. Piscitelli	H. Waugh
M. Duffy	R. Levin	F. Plank	D. Weiss
K. C. Fletcher	R. Li	C. Richards	J. Welch
J. Frazer	M. Liang	E. Richman	K. Wilcox
J. Fye	K. Liepmann	D. Rogers	V. Wu
A. Gelder	R. Low	M. Royer	R. Young
Y. Guan	M.-H. Lu	T. Schneider	J. Zhang
K. Haraguchi	J. Marsh	A. Serres	Y. Zong

CONTENTS

Introduction	1
1.0 Scope	1
2.0 Normative References	1
3.0 Definitions	2
3.1 array spectrometer	2
3.2 combined standard uncertainty	2
3.3 coverage factor	2
3.4 expanded uncertainty	2
3.5 photometer	2
3.6 primary standard (lamp)	2
3.7 relative uncertainty	3
3.8 secondary standard (lamp)	3
3.9 standard deviation (experimental)	3
3.10 standard uncertainty	3
3.11 uncertainty (of measurement)	3
3.12 working standard (lamp)	4
4.0 Guide to the Expression of Uncertainty in Measurement (GUM): A Summary	4
4.1 Introduction	4
4.2 Description of the Model Equation and Input Parameter Estimation	4
4.3 Standard Uncertainty Evaluation	4
4.3.1 Type A Uncertainty Evaluation	5
4.3.2 Type B Uncertainty Evaluation	5
4.4 Sensitivity Coefficients and Combined Standard Uncertainty	5
4.5 Expanded Uncertainty and Degrees of Freedom	7
5.0 Calibration of Integrating Spheres	8
5.1 Calibration Using a Total Luminous Flux Standard Lamp	8
5.1.1 Model for an Integrating Sphere with Photometer	8
5.1.2 Additional Uncertainty Components	11
5.1.3 Integrating Sphere with Photometer Sensitivity Coefficients	12
5.1.4 Example Uncertainty Budget	12
5.1.5 Measurand and Combined Standard Uncertainty	12

5.1.6	Degrees of Freedom and Expanded Uncertainty	13
5.1.7	Model Modifications	13
5.1.8	Multiple Standard Lamps.....	14
5.1.9	Describing the Model Equation in Relative Terms.....	17
5.1.10	Integrating Sphere with Spectrometer Detector.....	19
5.2	Calibration Using a Total Spectral Radiant Flux Standard Lamp.....	21
5.2.1	Model for an Integrating Sphere with Spectrometer	21
5.2.2	Wavelength to Channel Data Conversion	24
5.2.3	Uncertainty in Spectral Radiance	25
5.2.4	Additional Uncertainty Components.....	27
5.2.5	Integrating Sphere with Spectroradiometer Sensitivity Coefficients.....	30
5.2.6	Example Uncertainty Budget	30
5.2.7	Measurand and Combined Standard Uncertainty	30
5.2.8	Degrees of Freedom and Expanded Uncertainty	30
5.2.9	Model Modifications	30
6.0	Calibration of Working Standard Lamps Using an Integrating Sphere.....	33
6.1	Creating Total Luminous Flux Working Standard Lamps	33
6.1.1	Model for an Integrating Sphere with Photometer	33
6.1.2	Additional Uncertainty Components.....	36
6.1.3	Integrating Sphere with Photometer Sensitivity Coefficients for Working Standard Lamps.....	36
6.1.5	Example Uncertainty Budget	37
6.1.6	Measurand and Combined Uncertainty	37
6.1.7	Degrees of Freedom and Expanded Uncertainty	37
6.1.8	Model Modifications	38
6.1.9	Integrating Sphere with an Array Spectrometer	40
6.2	Creating Total Spectral Radiant Flux Working Standard Lamps	41
6.2.1	Model for an Integrating Sphere with Array Spectrometer.....	41
6.2.2	Additional Uncertainty Components.....	45
6.2.3	Integrating Sphere with Array Spectrometer Sensitivity Coefficients for Working Standard Lamps ..	45
6.2.4	Example Uncertainty Budget	46
6.2.5	Measurand and Combined Uncertainty	46
6.2.6	Degrees of Freedom and Expanded Uncertainty	47
6.3	Creating a Luminous Flux Working Standard Lamp from a Total Spectral Radiant Flux Standard Lamp ...	47
6.3.1	SI Unit Traceability: Total Spectral Radiant and Luminous Flux Scales.....	47
6.3.2	Integrating-Sphere Transfer Model: Total Spectral Radiant Flux to Luminous Flux.....	48
6.3.3	Additional Uncertainty Components.....	53
6.3.4	Array Spectrometer Data Summation Uncertainty.....	53
6.3.5	Lumen Sensitivity Coefficients for Integrating Sphere with Array Spectrometer	55
6.3.6	Example Uncertainty Budget	55
6.3.7	Measurand and Combined Standard Uncertainty	56
6.3.8	Degrees of Freedom and Expanded Uncertainty	56

Annex A – Tolerance, Uncertainty, and Confidence Intervals	56
Annex B – Partial Derivatives	61
Annex C – Sensitivity Coefficients Calculated as Slopes	63
Annex D – Total Spectral Radiant Flux Data Measured with an Array Spectrometer and Integrating Sphere	64
Annex E – Total Spectral Radiant Flux Data and Signal for Working Standard Lamp Transfer	66
Acknowledgements	69
Informative References	69

1.0 Introduction and Scope

1.1 Introduction

Uncertainty is inherent in every measurement. To truly understand the value of a measurement, the uncertainty of such should be analyzed. A critical step of this analysis is to know the uncertainty contributed by the instruments used to make the measurement. This Technical Memorandum provides a detailed and step-by-step method for determining uncertainty of certain calibration procedures common to instruments used in lumen measurement laboratories. This document uses the structure proposed in the *Guide to the Expression of Uncertainty in Measurement* (GUM) for its recommendations. For each calibration procedure presented in this document, a mathematical model is developed, estimates of each input quantity are discussed, and guidance is provided for the evaluation of the standard uncertainty of each input estimate. With this information, the combined standard uncertainty is calculated, effective degrees of freedom are calculated, and finally, the expanded uncertainty is reported.

The intent of this document is to provide a common approach to the uncertainty analysis for calibration of lumen measuring equipment (integrating spheres) with standard incandescent lamps, including halogen, that have been assigned values of total luminous flux and/or total spectral radiant flux. In addition, uncertainty budget templates are provided for calibration of incandescent (including halogen) working standard lamps that are virtually identical to the standard lamps they are compared against. These templates are starting points for uncertainty analysis. A laboratory applying these templates is expected to estimate each input quantity and the standard uncertainty of each input quantity, except where industry standard values are specifically noted in this document. Additional input quantities may be required based on specific situations that arise in the uniqueness of each laboratory measurement system. Future documents will cover topics such as uncertainty determination for colorimetric calibrations, and photometric and colorimetric measurements for testing general lighting products.

1.2 Scope

This document provides templates for the analysis of measurement uncertainty for the photometric calibration of integrating sphere systems using standard incandescent lamps, as well as the creation of working standard lamps using integrating sphere systems that are effectively identical to the primary standard lamps. It does not provide templates for the analysis of measurement uncertainty for colorimetric calibration of working standard lamps or for the testing of general lighting products.

2.0 Normative References

JCGM 100-2008, *Evaluation of Measurement Data – Guide to the Expression of Uncertainty in Measurement* (GUM). Bureau International des Poids et Mesures (BIPM); Sep 2008. Free download: www.bipm.org/en/publications/guides/gum.html.

ANSI/IES LS-1-20, *Nomenclature and Definitions for Illuminating Engineering*. New York: Illuminating Engineering Society; 2017. Free access from www.ies.org/standards/definitions/