



ANSI/ASA S1.15-2021/Part 4/ IEC  
61094-4:1995

AMERICAN NATIONAL STANDARD

**Electroacoustics – Measurement microphones – Part  
4: Specifications for working standard microphones  
(a nationally adopted international standard)**

**Secretariat:**

**Acoustical Society of America**

**Approved on July 12, 2021:**

**American National Standards Institute, Inc.**

**Abstract**

This part of IEC 1094 is applicable to working standard microphones. It specifies mechanical dimensions and certain electroacoustical characteristics for working standard microphones used in measuring systems for the determination of sound pressure to enable these microphones to be used as transfer standards in the calibration of acoustic measurement instruments.



## AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).

These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.

Acoustical Society of America  
Standards Secretariat  
1305 Walt Whitman Road, Suite 300  
Melville, New York 11747  
Telephone: + 1 (516) 576-2341  
E-mail: [standards@acousticalsociety.org](mailto:standards@acousticalsociety.org)



© 2021 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

These materials are subject to copyright claims of ISO or IEC and ASA. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of ASA.

## Contents

1	Scope and object .....	1
2	Normative references .....	1
3	Definitions .....	2
3.1	Working standard microphone.....	2
3.2	Principal axis of a microphone.....	2
3.3	Reference plane of a microphone.....	2
3.4	Effective front volume of a microphone.....	2
3.5	Correction grid.....	2
4	Reference environmental conditions .....	2
5	Classification of working standard microphones.....	2
5.1	General .....	2
5.2	Type designation .....	3
6	Characteristics of working standard microphones.....	3
6.1	Sensitivity.....	3
6.2	Effective front volume .....	3
6.3	Upper limit of the dynamic range of a microphone.....	3
6.4	Linearity range of the microphone sensitivity level.....	4
6.5	Static pressure dependence of microphone sensitivity .....	4
6.6	Temperature dependence of microphone sensitivity.....	4
6.7	Humidity dependence of microphone sensitivity .....	4
6.8	Stability of microphone sensitivity.....	4
6.9	Pressure equalizing leakage .....	5
7	Specifications.....	5
7.1	Mechanical dimensions .....	5
7.2	Electroacoustical specifications .....	6
7.3	Identification markings .....	7
	Annex A (informative) Provisional ground-shield reference configuration for Type WS3 microphones.....	10

## Tables

Table 1 – Nominal mechanical dimensions and tolerance limits for working standard microphones in Figure 1 .....	6
Table 2 – Electroacoustical specifications for working standard microphones.....	7
Table 3 – Tolerance values for the frequency response (in decibels).....	9

## Figures

Figure 1 – Mechanical configuration of Microphones.....	5
Figure 2 – Tolerance curves for frequency response on a logarithmic frequency scale, where $f_0$ , $f_1$ and $f_2$ are specified in Table 2.....	7
Figure A.1 – Mechanical attachment to microphone, showing the provisional ground-shield reference configuration for Type WS3 microphone.....	10

## Foreword

*This Foreword is for information only and is not a part of the American National Standard ANSI/ASA S1.15-2021/Part 4/ IEC 61094-4:1995 American National Standard Electroacoustics - Measurement microphones - Part 4: Specifications for working standard microphones. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the standard.*

This standard is a nationally adopted international standard (NAIS). This standard comprises a part of a group of definitions, standards, and specifications for use in acoustics. It was developed and approved by Accredited Standards Committee S1 Acoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S1 is as follows:

*Standards, specifications, methods of measurement and test, and terminology in the field of physical acoustics, including architectural acoustics, electroacoustics, sonics and ultrasonics, and underwater sound, but excluding those aspects which pertain to biological safety, tolerances, and comfort.*

The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.

The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.

Documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.

In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.

The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards. Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 1094-4 has been prepared by IEC Technical Committee 29: Electroacoustics. The text of this standard is based on the following documents:

DIS	Report on voting
29/295/DIS	29/312/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

Note that in this national adoption, the decimal sign is a comma on the line and international English spelling is used throughout.

From 1 January 1997 all IEC publications were reissued with a designation in the 60000 series. Therefore, the numbers of the IEC standards that appear in this adopted version of the IEC standard have been editorially updated accordingly. Annex A is for information only.

At the time this Standard was submitted to Accredited Standards Committee S1, Acoustics for approval, the membership was as follows:

Angelique Scharine, *Chair*  
Richard J. Peppin, *Vice-Chair*  
  
Nancy Blair-DeLeon, *Secretary*

<b>Acoustical Society of America</b> .....	Richard J. Peppin Robert D. Hellweg (Alt.)
<b>Air-Conditioning, Heating and Refrigeration Institute</b> .....	Stephen J. Lind
<b>ETS-Lindgren Acoustic Systems</b> .....	Douglas Winker Michael Black (Alt.)
<b>Information Technology Industry Council</b> .....	Willem M. Beltman Stephanie Barrett (Alt.)
<b>Josephson Engineering, Inc.</b> .....	David L. Josephson
<b>National Council of Acoustical Consultants</b> .....	Anthony Nash
<b>National Institute of Standards &amp; Technology (NIST)</b> .....	Randall P. Wagner Richard Allen (Alt.)
<b>OTOJOY</b> .....	Thomas Kaufmann
<b>PCB Group</b> .....	Ken Cox Chad M. Walber (Alt.)
<b>Scantek, Inc.</b> .....	Steve Marshall William Gallagher (Alt.)
<b>TSI, Incorporated</b> .....	Steve Michalski Rich Bechtel (Alt.)
<b>U.S. Air Force (USAF)</b> .....	Richard L. McKinley Frank Mobley (Alt.)
<b>U.S. Army Construction Engineering Research Laboratory</b> .....	Michelle Swearingen Michael J. White (Alt.)
<b>U.S. Naval Surface Warfare Center - Carderock</b> .....	Kuangcheng Wu Jason Smoker (Alt.)

Individual Experts of Accredited Standards Committee S1, Acoustics, were:

William Gallagher  
Peter Hanes  
Tomasz R. Letowski

Richard McKinley  
Karl Peterman  
Christopher J. Struck

Chad Walber  
Lixue Wu

Working Group S1/WG 1, Standard Microphones and their Calibration, which assisted Accredited Standards Committee S1, Acoustics, in the development of this standard, had the following membership.

Chad.M. Walber, *Chair*  
Christopher J. Struck, *Vice-Chair*

David L. Josephson

Randall P. Wagner

Lixue Wu

Suggestions for improvements to this standard will be welcomed. They should be sent to Accredited Standards Committee S1, Acoustics, in care of the Standards Secretariat of the Acoustical Society of America, 1305 Walt Whitman Road, Suite 300, Melville, New York 11747. Telephone: + 1 (516) 576-2341; E-mail: [standards@acousticalsociety.org](mailto:standards@acousticalsociety.org).

# American National Standard

## Electroacoustics – Measurement microphones – Part 4: Specifications for working standard microphones (a nationally adopted international standard)

### 1 Scope and object

This part of ASA S1.15/IEC 61094 is applicable to working standard microphones.

It specifies mechanical dimensions and certain electroacoustical characteristics for working standard microphones used in measuring systems for the determination of sound pressure to enable these microphones to be used as transfer standards in the calibration of acoustic measurement instruments.

It establishes a system for classifying working standard microphones into a number of types according to their dimensions and properties in order to facilitate the specification of measurement systems, the calibration of measuring systems and microphones by sound calibrators, and the interchangeability of microphones in given measuring and calibration systems.

It does not specify the transduction principle by which working standard microphones operate.

### 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ASA S1.15/IEC 61094. At the time of publication, the editions indicated were valid. All normative documents are subject to revision, and parties to agreements based on this part of ASA S1.15/IEC 61094 are encouraged to investigate the possibility of applying the most recent editions of the normative documents listed below. Members of IEC and ISO maintain registers of currently valid International Standards.

IEC 60942 *Electroacoustics – Sound calibrators*

IEC 61094-1 *Electroacoustics – Measurement microphones - Part 1: Specifications for laboratory standard microphones*

IEC 61094-2 *Electroacoustics – Measurement microphones - Part 2: Primary method for pressure calibration of laboratory standard microphones by the reciprocity technique*

IEC 61094-3 *Electroacoustics – Measurement microphones - Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique*

ANSI/ASME B1.1-1982 *Unified Inch Screw Threads (UN & UNR Thread Form)* \*

ISO/IEC Guide 98-3 *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement*

---

\* Reference is given to ANSI B1.1 in the absence of an equivalent International Standard.