



BSI Standards Publication

**Organic and organo-mineral fertilizers -  
Determination of the content of specific  
elements by ICP-AES after extraction by water**

---

## National foreword

This Published Document is the UK implementation of CEN/TS 17774:2022.

The UK participation in its preparation was entrusted to Technical Committee CII/37, Fertilisers and related chemicals.

A list of organizations represented on this committee can be obtained on request to its committee manager.

### Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication is not to be regarded as a British Standard.

© The British Standards Institution 2022  
Published by BSI Standards Limited 2022

ISBN 978 0 539 17958 3

ICS 65.080

### **Compliance with a Published Document cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 June 2022.

### Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

---

TECHNICAL SPECIFICATION  
SPÉCIFICATION TECHNIQUE  
TECHNISCHE SPEZIFIKATION

**CEN/TS 17774**

April 2022

ICS 65.080

English Version

**Organic and organo-mineral fertilizers - Determination of  
the content of specific elements by ICP-AES after  
extraction by water**

Engrais organiques et organo-minéraux -  
Détermination de la teneur en éléments spécifiques par  
ICP-AES après extraction à l'eau

Organische und organisch-mineralische Düngemittel -  
Bestimmung des Gehalts spezifischer Elemente mittels  
ICP-AES nach Extraktion mit Wasser

This Technical Specification (CEN/TS) was approved by CEN on 13 March 2022 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>		Page
<b>European foreword</b> .....		<b>3</b>
<b>Introduction</b> .....		<b>4</b>
<b>1</b>	<b>Scope</b> .....	<b>5</b>
<b>2</b>	<b>Normative references</b> .....	<b>5</b>
<b>3</b>	<b>Terms and definitions</b> .....	<b>5</b>
<b>4</b>	<b>Principle</b> .....	<b>5</b>
<b>5</b>	<b>Interferences</b> .....	<b>6</b>
<b>5.1</b>	<b>General</b> .....	<b>6</b>
<b>5.2</b>	<b>Spectral interferences</b> .....	<b>6</b>
<b>5.3</b>	<b>Transport interferences</b> .....	<b>6</b>
<b>5.4</b>	<b>Excitation interferences</b> .....	<b>6</b>
<b>5.5</b>	<b>Chemical interferences</b> .....	<b>6</b>
<b>5.6</b>	<b>Memory interferences</b> .....	<b>6</b>
<b>6</b>	<b>Reagents</b> .....	<b>7</b>
<b>7</b>	<b>Apparatus</b> .....	<b>7</b>
<b>8</b>	<b>Procedure</b> .....	<b>8</b>
<b>8.1</b>	<b>Preparation of test and blank solution</b> .....	<b>8</b>
<b>8.2</b>	<b>Preparation of the calibration solutions</b> .....	<b>8</b>
<b>8.3</b>	<b>Measurement</b> .....	<b>8</b>
<b>9</b>	<b>Calculation and expression of the results</b> .....	<b>11</b>
<b>10</b>	<b>Test report</b> .....	<b>11</b>
<b>Bibliography</b> .....		<b>12</b>

## **European foreword**

This document (CEN/TS 17774:2022) has been prepared by Technical Committee CEN/TC 260 “Fertilizers and liming materials”, the secretariat of which is held by DIN

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a standardization request given to CEN by the European Commission and the European Free Trade Association.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Introduction**

This document is part of a modular approach and concerns the analytical measurement step. “Modular” means that a test standard concerns a specific step in assessing a property and not the whole chain of measurements. Inductively coupled plasma atomic emission spectrometry (ICP-AES) is nowadays widely used and a well-established method in many laboratories.

## 1 Scope

This document specifies a method for the determination of boron (B), cobalt (Co), copper (Cu), iron (Fe), manganese (Mn), molybdenum (Mo) and zinc (Zn) in organic fertilizers and organo-mineral fertilizers extracts using inductively coupled plasma-atomic emission spectrometry (ICP-AES).

NOTE Alternatively, inductively coupled plasma mass spectrometry (ICP-MS) can be used for the measurement if the user proves that the method gives the same results.

This method is applicable to water extracts prepared according to CEN/TS 17766. The method can be used for the determination of other elements, provided the user has verified the applicability.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CEN/TS 17766, *Organic and organo-mineral fertilizers — Extraction by water for subsequent determination of elements*

EN 12944-1, *Fertilizers and liming materials — Vocabulary — Part 1: General terms*

EN 12944-2, *Fertilizers and liming materials — Vocabulary — Part 2: Terms relating to fertilizers*

EN ISO 3696:1995, *Water for analytical laboratory use — Specification and test methods (ISO 3696:1987)*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1 and EN 12944-2 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

## 4 Principle

The method is based on the ICP-AES measurement of the concentration of boron (B), cobalt (Co), copper (Cu), iron (Fe), manganese (Mn), molybdenum (Mo) and zinc (Zn) in fertilizer extracts prepared according to CEN/TS 17766. The elements are determined after appropriate dilution of the extract. The solution is dispersed by a nebulizer of the ICP-AES instrument and the resulting aerosol is transported into the plasma. Element specific emission spectra are produced by a radio-frequency inductively coupled argon plasma where atoms or ions are excited at high temperature. The emission spectra are dispersed by a spectrometer, and the intensities of the emission lines are monitored by photosensitive devices. Multi-element determinations using sequential or simultaneous optical systems and axial or radial viewing of the plasma may be used.

The method may be used for the determination of other elements, provided the user has verified the applicability.