



BSI Standards Publication

## Organic and organo-mineral fertilizers — Determination of the mercury content

---

## National foreword

This Published Document is the UK implementation of CEN/TS 17769:2022.

The UK participation in its preparation was entrusted to Technical Committee CII/37, Fertilisers and related chemicals.

A list of organizations represented on this committee can be obtained on request to its committee manager.

### Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication is not to be regarded as a British Standard.

© The British Standards Institution 2022  
Published by BSI Standards Limited 2022

ISBN 978 0 539 17899 9

ICS 65.080

### **Compliance with a Published Document cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 July 2022.

### Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

---

TECHNICAL SPECIFICATION  
SPÉCIFICATION TECHNIQUE  
TECHNISCHE SPEZIFIKATION

**CEN/TS 17769**

April 2022

ICS 65.080

English Version

**Organic and organo-mineral fertilizers - Determination of  
the mercury content**

Engrais organiques et organo-minéraux -  
Détermination de la teneur en mercure

Organische und organisch-mineralische Düngemittel -  
Bestimmung des Quecksilbergehaltes

This Technical Specification (CEN/TS) was approved by CEN on 13 March 2022 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>	Page
<b>European foreword</b> .....	<b>3</b>
<b>Introduction</b> .....	<b>4</b>
<b>1 Scope</b> .....	<b>5</b>
<b>2 Normative references</b> .....	<b>5</b>
<b>3 Terms and definitions</b> .....	<b>5</b>
<b>4 Principle</b> .....	<b>5</b>
<b>5 Interferences</b> .....	<b>5</b>
<b>6 Reagents</b> .....	<b>6</b>
<b>7 Apparatus</b> .....	<b>7</b>
<b>8 Procedure</b> .....	<b>7</b>
<b>9 Calculation and expression of the results</b> .....	<b>9</b>
<b>10 Test report</b> .....	<b>10</b>
<b>Annex A (informative) Methods for mercury determination</b> .....	<b>11</b>
<b>A.1 General</b> .....	<b>11</b>
<b>A.2 Cold vapour generation (hydride generation)</b> .....	<b>11</b>
<b>A.3 Direct amalgamation technique</b> .....	<b>12</b>
<b>A.4 Inductively coupled plasma mass spectrometry (ICP-MS)</b> .....	<b>12</b>
<b>Bibliography</b> .....	<b>13</b>

## **European foreword**

This document (CEN/TS 17769:2022) has been prepared by Technical Committee CEN/TC 260 “Fertilizers and liming materials”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Introduction**

This document concerns the analytical measurement step for the determination of mercury in organic fertilizers and organo-mineral fertilizers after digestion by aqua regia according to CEN/TS 17768. The document covers cold vapour generation followed by mercury determination using atomic absorption spectrophotometry (AAS). Different cold vapour generation techniques can be used (flow injection, segmented flow, batch). The document also includes a method based on a direct amalgamation technique which is widely used in many analytical laboratories. It is also possible to use other suitable methods of mercury determination described in Annex A, if users prove that the methods give the same results as the methods described in this document.

## 1 Scope

This document specifies a method for determination of the content of mercury (Hg) in organic fertilizers and organo-mineral fertilizers using (cold) vapour generation apparatus coupled to an atomic absorption spectrophotometer and a method using a direct amalgamation technique. It is applicable to aqua regia digests prepared according to CEN/TS 17768.

NOTE It is also possible to use other suitable methods for the determination of mercury described in Annex A if users prove that the method gives the same results as the methods described in this standard.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CEN/TS 17773, *Organic and organo-mineral fertilizers — Determination of the dry matter content*

## 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

## 4 Principle

### 4.1 Vapour generation atomic absorption spectrophotometry (VG-AAS)

Mono- and divalent mercury is reduced to the elemental form by tin(II) chloride or sodium borohydride in an acid medium. Elemental mercury is stripped off from the solution and determined in the form of an atomic gas by an atomic absorption spectrophotometer.

### 4.2 Direct amalgamation (DA)

The sample is thermally decomposed in an oxygen rich environment. The decomposition products are carried to an amalgamator that selectively traps mercury. After the system is flushed with oxygen to remove any remaining gases or decomposition products, the amalgamator is rapidly heated, releasing mercury vapour. Flowing oxygen carries the mercury vapour through absorbance cells positioned in the light path of a single wavelength atomic absorption spectrophotometer. Absorbance is measured at 253,7 nm as a function of mercury concentration.

## 5 Interferences

### 5.1 Vapour generation atomic absorption spectrophotometry (VG-AAS)

The matrix of the solution analysed is dominated by the acids used in the digestion step. Tin(II) chloride as a reduction substance is recommended, because sodium borohydride reduces many elements commonly found in organic fertilizers and organo-mineral fertilizers digests to the elemental state, which can cause matrix problems under particular circumstances. However, it is still possible to use sodium borohydride as a reduction agent. The interferences due to the presence of other elements in