

ASA/ANSI S2.2-1959 (R2020)



AMERICAN NATIONAL STANDARD

**Methods for the Calibration of Shock and
Vibration Pickups**

(a nationally adopted international standard)

Secretariat:

Acoustical Society of America

American National Standards Institute, Inc.

ANSI[®]
S2.2-1959

Reaffirmed by ANSI
July 10, 2001

Reaffirmed by ANSI
May 19, 2006

Reaffirmed by ANSI
June 28, 2011

Reaffirmed by ANSI
June 16, 2016

Reaffirmed by ANSI June 19, 2020

American National Standard

**Methods for the
Calibration of Shock and
Vibration Pickups**

Secretariat

**Acoustical Society of America
American Society of Mechanical Engineers**

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires verification by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made toward their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Published by

**American National Standards Institute
11 West 42nd Street, New York, New York 10036**

Copyright ©1960 by American National Standards Institute
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Foreword

(This Foreword is not a part of American Standard¹ Methods for the Calibration of Shock and Vibration Pickups, S2.2-1959.)

This American Standard¹ comprises a part of a group of definitions, standards, and specifications for use in work concerning acoustics, vibration, and mechanical shock. It has been developed under the Sectional Committee Method of ASA procedure, under the co-sponsorship of the Acoustical Society of America and the American Society of Mechanical Engineers.

The S2 Committee under whose jurisdiction this standard was developed has the following scope:

Standards, specifications, methods of measurement and test, and terminology in the fields of mechanical shock and vibration, but excluding those aspects that pertain to biological safety, tolerance, and comfort.

Various subcommittees have been organized to take care of the committee's program and this standard was developed by Subcommittee S2-W-27, whose personnel is shown below.

Suggestions for improvement gained in the use of this standard will be welcomed. They should be sent to the United States of America Standards Institute, 10 East 40th Street, New York, N. Y. York 11

The organizations which participated in this work and the names of their representatives, as listed at the time this standard was submitted to the S2 Sectional Committee for approval, were as follows:

H. M. TRENT, *Chairman*

C. E. CREDE, *Vice-Chairman*

S. DAVID HOFFMAN, *Secretary*

<i>Organization Represented</i>	<i>Name of Representative</i>
Acoustical Society of America.....	H. M. TRENT
American-British-Canadian Navy Tripartite Standardization.....	I. VIGNESS
American Institute of Electrical Engineers.....	H. E. STIELING
American Society of Mechanical Engineers.....	W. W. BENNETT (<i>Alt</i>)
	C. A. ARENTS
	C. E. CREDE
	D. C. KENNAED, JR
	HOWARD S. BEAN (<i>Alt</i>)
American Society for Testing Materials.....	CYRIL M. HARRIS
Canadian Standards Association (<i>Liaison</i>).....	T. D. NORTHWOOD
Electronic Industries Association.....	E. W. BISSON
National Bureau of Standards, U. S. Department of Commerce.....	WALTER RAMBERG
	R. R. BOUCHE (<i>Alt</i>)
	S. EDELMAN (<i>Alt</i>)
National Electrical Manufacturers Association.....	ROBERT O. FEHR
	A. C. HAGG (<i>Alt</i>)
Society of Automotive Engineers Inc.....	C. T. MOLLOY
Society for Experimental Stress Analysis.....	EDWARD G. FISCHER
Telephone Group.....	R. L. HANSON
U. S. Department of the Air Force (<i>Liaison</i>).....	D. C. KENNAED, JR
	ROBERT F. WILKUS (<i>Alt</i>)
U. S. Department of the Army, Signal Corps (<i>Liaison</i>).....	JOSEPH J. OLIVERI
U. S. Department of Defense, Office of Assistant Secretary for Research and Development.....	ELIAS KLEIN
U. S. Department of the Navy, Bureau of Ships, Code 621.....	J. R. SULLIVAN
U. S. Department of the Navy, Bureau of Ordnance, Special Products Office.....	L. SCHLESINGER
	JOHN C. NEW (<i>Alt</i>)
Members-at-Large.....	LAURENCE BATCHELDER
	LEO L. BERANEK
	FLOYD A. FIRESTONE
	F. V. HUNT
	W. D. NEFF
	A. P. G. PETERSON
	R. W. YOUNG

Writing Group S2-W-27 on Calibration of Shock and Vibration Measuring Instruments, which prepared this standard, had the following personnel:

WALTER RAMBERG, *Chairman until December 1958*

R. R. BOUCHE, *Chairman after December 1958*

ABRAHAM I. DRANETZ

J. LLOYD JONES

H. L. RICH

S. P. THOMPSON

K. UNHOLTZ

IRWIN VIGNESS

¹ All American Standards are now designated USA Standards.

Contents

SECTION	PAGE
Introduction.....	6
1. Purpose and Scope.....	6
1.1 Purpose.....	6
1.2 Scope.....	6
1.3 Present and Future Choices of Methods.....	6
1.4 Related Standards.....	6
2. Definitions.....	7
2.1 Letter Symbols.....	7
2.2 Pickup.....	7
2.3 Input.....	8
2.4 Output.....	8
2.5 Sensitivity, Amplitude Sensitivity, Phase Lag.....	8
2.6 Amplitude Distortion, Frequency Distortion, Proportional Phase Lag, Phase Distortion.....	8
2.7 Calibration Factor.....	9
2.8 Transverse Sensitivity Ratio.....	9
2.9 Exciter.....	9
2.10 Damping Ratio.....	9
2.11 Undamped Natural Frequency.....	9
2.12 Damped Natural Frequency.....	9
2.13 Resonant Frequency.....	9
3. Characteristics to be Measured.....	9
3.1 General.....	9
3.2 Direct Response.....	9
3.3 Response to Motion in Other Directions and About Other Axes.....	10
3.4 Damping Ratio and Undamped Natural Frequency.....	10
3.5 Nonlinearity.....	12
4. Standard Calibration Methods.....	12
4.1 Tabulation of Standard Methods.....	13
4.2 Tilting Support Calibrator.....	13
4.3 Centrifuge Calibrator.....	14
4.4 Rectilinear Electrodynamic Vibration Pickup Calibrator.....	15
4.4.1 Sensitivity of Electrodynamic Calibrator Determined by Reciprocity Method.....	16
4.4.2 Calibration of Vibration Pickups on Calibrator, General Method.....	17
4.5 Physical Pendulum Calibrator.....	18
4.6 Ballistic Pendulum Calibrator.....	20
5. Calibration by Comparison with Calibrated Pickup.....	22
Appendix	
A1. Survey of Other Calibration Methods.....	23
A1.1 Static Calibration Methods.....	23
A1.1.1 Static Load Method.....	23
A1.2 Dynamic Calibration by Sinusoidal Excitation.....	23
A1.2.1 Rectilinear Vibration Exciter.....	24
A1.2.2 Angular Vibration Exciter.....	25
A1.2.3 Tilted Centrifuge.....	26
A1.2.4 Dual Centrifuge.....	26
A1.2.5 Resonant Beam Vibrator.....	26
A1.2.6 Acoustical Cavity Method.....	27
A1.2.7 Condenser Microphone Method.....	27

A1.3 Dynamic Calibration by Transient Excitation.....	28
A1.3.1 Free-Fall Test	28
A1.3.2 Pulse Calibrator.....	29
A1.3.3 Air Gun	29
A1.3.4 Shock Test Applying Measured Impact Acceleration	29
A1.3.5 Hydraulic Shock Test.....	30
A1.3.6 Impact Drop Test.....	30
A1.3.7 Elastic Pulses in Long Bars.....	30
A2. Theory of Reciprocity Method for Linear Electromechanical Pickups.....	30
A3. Reciprocity Method for Electrodynamic Calibrators	32
A3.1 Theory.....	32
A3.2 Measurements.....	34
A3.2.1 Transfer Admittance Measurements.....	35
A3.2.2 Voltage Ratio Measurements.....	35
 Table 1 Estimated Ranges and Errors of Standard Methods.....	 13
 Figures	
Fig. 1 Response of Idealized Linear Acceleration Pickup.....	10
Fig. 2 Response of Idealized Linear Displacement or Velocity Pickup	11
Fig. 3 Phase Lag of Idealized Linear Pickup	11
Fig. 4 Response of Idealized Linear Pickup to a Sudden Impulse.....	12
Fig. 5 Tilting Support for Static Calibration of Acceleration Pickups.....	14
Fig. 6 Schematic Cross-sectional View of Rectilinear Electrodynamic Vibration Pickup Calibrator.....	15
Fig. 7 Circuit for Calibration by Method A.....	18
Fig. 8 Pendulum Calibrator.....	19
Fig. 9 Recorded Output of Pickup on Physical Pendulum Calibrator.....	20
Fig. 10 Calibration of Pickup A on Ballistic Pendulum Calibrator.....	20
Fig. 11 Simple RC Network for Integrating Pickup Output.....	21
Fig. 12 Comparison Method for Calibrating Pickup 2 Against Calibrated Pickup 1.....	21
Fig. A1 Vibrating Wedge.....	23
Fig. A2 Rectilinear Vibration Pickup Calibrator with Piezoelectric Drive.....	25
Fig. A3 Dual Centrifuge.....	26
Fig. A4 Resonant Beam Vibrator.....	27
Fig. A5 Impact Load Acting on Mass and Displacement of Mass of Linear Mass-Spring System, with Natural Period T	28
Fig. A6 Typical Oscillograph Record from Free-Fall Test of Velocity Pickup.....	29
Fig. A7 Shock Test Applying Measured Impact Acceleration.....	29
Fig. A8 Circuits and Polygons for Making Transfer Admittance Measurements on Electro- dynamic Calibrator and Pickups.....	34
Fig. A9 Circuits and Polygons for Making Voltage Ratio Measurements on Electrodynamic Calibrator.....	36
 References	 37

USA Standard Methods for the Calibration of Shock and Vibration Pickups

Introduction

The calibration of shock and vibration pickups has become increasingly important as the need has grown for accurate measurements of the shocks and vibrations to which aircraft, missiles, and other high performance equipment are subjected in service. Numerous methods have been used or proposed for these calibrations. These are briefly described in Section 4 of this standard and in the Appendix. Section 4 contains a tabulation and detailed description of five "standard methods" which have proven themselves to be reliable means for the calibration of shock and vibration pickups. A tabulation (see 4.1) lists approximate ranges of application of each standard method and estimates of attainable accuracy. The description, which follows the tabulation, gives details of performing calibrations by each of the standard methods. Section A1 contains a brief survey of other methods which have been used to calibrate vibration pickups, but which have not found sufficient acceptance as yet to be included among the standard methods. The methods described range from precision techniques suitable for laboratory use only to procedures appropriate to field use.

It is anticipated that the standard will undergo frequent revision to keep pace with the advancement of the art in this new and growing field. Section 4 will grow as methods in Section A1 prove their adequacy as standards. Section A1 will expand as new methods make their appearance.

Calibration methods for both shock and vibration pickups are included in the same standard because it is impracticable to draw a line of demarcation between pickups used in measurements of shocks and vibrations. Using Fourier analysis [1]¹ any shock can be described by the superposition of a sufficient number of sinusoidal vibrations of proper frequency, amplitude, and phase angle.

The standard is limited to the calibration of acceleration, velocity, and displacement pickups. It is not concerned with pickups used for measurements of force, pressure, or strain even though some of these may be calibrated by similar methods.

¹ Numbers in brackets denote references at the end of the standard.

1. Purpose and Scope

1.1 Purpose. This standard is designed to acquaint the user with the general principles of calibration of shock and vibration pickups and to describe concisely several standard methods which have proven to give reliable and reproducible results. Further details concerning these methods are given in the Appendix. Also, other methods that have not as yet reached the stage of development of the standard methods are described briefly in the Appendix.

1.2 Scope. Five methods have been selected as standard methods for the calibration of acceleration, velocity, and displacement pickups. They are described in Section 4 of this standard.

It is impracticable to calibrate all pickups by one standard method over the entire frequency and amplitude range of vibrations and shocks to be measured by the pickups. Several methods are accordingly described. Each method is limited to a range of frequency and amplitude, and to the weight of pickup that can be calibrated. The limitations may include, in addition, other variables such as volume of the pickup and temperature of operation.

1.3 Present and Future Choices of Methods. A pickup should be calibrated by the standard method or methods corresponding to its range of frequency, amplitude, and mass. The choice of method is facilitated by the tabulation of standard methods with estimated ranges and accuracies given in 4.1.

Other methods of calibration, of the types referred to in the Appendix, may be used in the future as these methods attain the degree of reliability characteristic of a standard method.

1.4 Related Standards. This standard is coordinated with the following standards which are closely related to the calibration of pickups:

American Standard¹ Method for Specifying the Characteristics of Pickups for Shock and Vibration Measurement, Z24.21-1957;

American Standard¹ Method for Specifying the Characteristics of Auxiliary Equipment for Shock and Vibration Measurements, S2.4-1960;² and

American Standard¹ Shock and Vibration Terminology, S1.1-1960.²

² These standards will become available in the latter part of 1960.

¹All American Standards are now designated USA Standards.